

MCP Hahnemann University

Operated by



Radiation Physics & Safety Department
Hahnemann Division
Mail Stop 106 • 245 N. 15th Street • Philadelphia, PA 19102-1192
TEL 215.762.4050 • FAX 215.762.7683
www.mcphu.edu

March 12, 2001

Tom Thompson
U.S. Nuclear Regulatory Commission
Region 1
Telefacsimile No. 610-337-5269

Dear Mr. Thompson:

We confirm that surveys will be performed around patients implanted with eye plaques as follows:

1. The patient will be surveyed for dislodged or misplaced sources immediately upon implant the eye plaque as previously described.
2. The room in which the patient is housed will be surveyed along each wall where the highest dose rates would be expected in lieu of performing a survey in contiguous restricted and unrestricted areas. If the dose rate exceeds 2 millirem per hour along any wall, then a measurement will be performed in the room or area adjacent to the patient room sharing the wall along which the dose rate exceeded 2 millirem per hour.

If you have any questions, please feel free to contact me.

Sincerely,

Kent Lambert, M.S., CHP
Director, Radiation Safety

129172 TKT
NMSS/RGN MATERIALS-002