

## Office Memorandum • UNITED STATES GOVERNMENT

TO : R. W. Kirkman, Director, Compliance Division  
New York Operations Office

DATE: October 20, 1960

FROM : Leo Dubinski, Director, Inspection Division  
Oak Ridge Operations Office

SUBJECT: ALPHATRON PRESSURE TRANSDUCER CONTAINING 375 MC H 3

SYMBOL: MI:JBH

During a recent inspection made by this office of the U. S. National Aeronautics and Space Administration it was noted that the licensee had received a 375-mc tritium foil incorporated in a National Research Corporation pressure transducer model No. 0716. The National Research Corporation is located at 70 Memorial Drive, Cambridge 42, Massachusetts. The licensee stated that when he received the pressure transducer, there was no labeling of any kind to denote that it contained byproduct material. Since the National Research Corporation is in your area and most likely is an AEC licensee, we are forwarding this information to you. Whether or not it is of any significance, we have no way of knowing.

Leo Dubinski  
Leo Dubinski

TO	ACTION	SIGNATURE	DATE
CMP			

ITEM # H

A/4

(1)