



# NRC EQUIPMENT CORPORATION

A SUBSIDIARY OF NATIONAL RESEARCH CORPORATION

CABLE ADDRESS: NARESCO

160 CHARLEMONT STREET, NEWTON HIGHLANDS 61, MASSACHUSETTS

TELEPHONE  
DECATUR 2-5800

December 19, 1962

Mr. Richard Gilbert  
Division of Compliance  
U.S. Atomic Energy Commission  
376 Hudson Street  
New York 14, New York

Dear Mr. Gilbert:

On December 13, 1962 you called our company and spoke to Mr. F. H. Greene, Jr. regarding the labeling of our commercial shipments of Alphatron R vacuum gauges containing both Radium 226 and Tritium foils. Mr. Greene has asked me to reply to your questions since I am concerned with this problem here at NRC Equipment Corporation.

We no longer sell any device containing tritium foil either commercially as a catalogue item or on a special contract basis. In May 1961, it was discovered that the tritium foils used in our Model 0716 Alphatron Pressure Transducer were leaking tritium gas at a rate in excess of the allowable AEC limits for sealed sources. We immediately reported this incident to the AEC and to the manufacturer of these sources. We further sent out a letter to all customers who had purchased this instrument, informing them of the problem and of their obligations as AEC licensed recipients of this device. All sales of these gauges were stopped and no further shipments of tritium were made in any form. A survey of our premises was also made by Industrial Hygiene Associates, to evaluate any potential hazard to our personnel. No hazard was found. Since this time we have not considered these foils safe enough for our purposes and have disposed of all inventory of sources in accordance with AEC regulations.

At the present time we are complying with all existing ICC regulations regarding the labeling and shipment of radioactive material (Radium 226) for our commercial Alphatrons. The outer box is labelled with the standard ICC label for Group I or II "D" (Radioactive) poison. Copies of these labels are enclosed.

ITEM # 6

R/G

2

This label notes the type, amount and intensity of gamma radiation in units (mrhm) emanating from the package. These limits are well within levels set by the ICC. For instance, the intensity of a gamma radiation for our most "active" Alphatron (in ICC units) is less than 0.8.

Our instruction literature also includes a section on the operation of the gauge for maximum safety and references the possible registration requirements of the various states. The outside labelling of the radioactive section of the gauge is clearly labelled in accordance with AEC requirements.

We hope this information will answer your questions. If you have any further questions please do not hesitate to contact me.

Very truly yours,

NRC EQUIPMENT CORPORATION

*F. L. Torney, Jr.*

F. L. Torney, Jr.  
Radiation Safety Officer - NRCEC

*Cambridge*  
*Eff 4-5400*

FLT/ld  
cc/FHG  
HAS

Enc.

