

Q BioChem, L.L.C.

A USEPA Contract Laboratory

Analysis, Research & Technology Development

FAX Cover Sheet

To: To Whom It May Concern Date: 4/17/01Affiliation: NRCFAX Number: 404-562-4955From: Kelly CrueySubject: Most Recent Leak Test Certifications.Comments: Please call myself or Sue Deegan
with any further questions.

Thank you

Kelly / Sue

Confidentiality Notice

This message is intended only for the use of the individual or entity to which it is addressed and contains information that is privileged and confidential. If you have received this communication in error, please notify the sender immediately by telephone and return the original message to us at the address below at our cost.

Total No. of Pages (including this sheet): 6 Please call us with any transmission problems.

1401 Municipal Road, Roanoke, VA 24012-1309 • 800-897-0297
540-265-7211 (voice) • 540-563-4866 (fax)

ML011140



DETECTOR SERVICE CENTER, INC.
P.O. Box 1044
Cary, NC 27512

Steve Scarabin
Randy Watts
(919) 469-0259

CERTIFICATE OF INSPECTION

CUSTOMER _____

ADDRESS _____

ATTENTION: _____

DETECTOR TYPE: NI63

SERIAL NO. F5338

DATE OF TEST: 4-5-01

AMOUNT OF
CONTAMINATION

To Be Completed By DSC

HOUSING: 0.91×10^{-4} pCi

ENTRANCE: 2.0×10^{-4} pCi

EXIT: 1.4×10^{-4} pCi

WITHIN NORMAL LIMIT:

☒ YES

NO

NEXT WIPE TEST REQUIRED:

Oct. 01

BY: Steve Scarabin INSPECTOR



DETECTOR SERVICE CENTER, INC.
P.O. Box 1044
Cary, NC 27512

Steve Scurabin
Randy Watts
(919) 469-0259

CERTIFICATE OF INSPECTION

CUSTOMER _____

ADDRESS _____

ATTENTION: _____

DETECTOR TYPE: NI 63

SERIAL NO. ^{P:} 0813

DATE OF TEST: 4-5-01

To Be Completed By DSC

AMOUNT OF CONTAMINATION: 1.2×10^{-4} pci

WITHIN NORMAL LIMIT: ☒ YES ☐ NO

NEXT WIPE TEST REQUIRED: Oct. 01

BY: Steve Scurabin INSPECTOR



DETECTOR SERVICE CENTER, INC.
P.O. Box 1044
Cary, NC 27512

Steve Scarabin
Randy Watts
(919) 469-0259

CERTIFICATE OF INSPECTION

CUSTOMER _____
ADDRESS _____

ATTENTION: _____

DETECTOR TYPE: NI63

SERIAL NO. F4693

DATE OF TEST: 4-5-01

AMOUNT OF
CONTAMINATION

To Be Completed By DSC

HOUSING: 1.0×10^{-4} pCi

ENTRANCE: 0.90×10^{-4} pCi

EXIT: 0.95×10^{-4} pCi

WITHIN NORMAL LIMIT: (YES) NO

NEXT WIPE TEST REQUIRED: Oct. 01

BY: Steve Scarabin INSPECTOR



DETECTOR SERVICE CENTER, INC.
P.O. Box 1044
Cary, NC 27512

Steve Scarabin
Randy Watts
(919) 469-0259

CERTIFICATE OF INSPECTION

CUSTOMER _____

ADDRESS _____

ATTENTION: _____

DETECTOR TYPE: NI 63

SERIAL NO. F 4694

DATE OF TEST: 4-5-01

AMOUNT OF
CONTAMINATION

To Be Completed By DSC

HOUSING: $0.70 \times 10^{-4} \mu\text{Ci}$

ENTRANCE: $1.0 \times 10^{-4} \mu\text{Ci}$

EXIT: $1.1 \times 10^{-4} \mu\text{Ci}$

WITHIN NORMAL LIMIT: YES NO

NEXT WIPE TEST REQUIRED: Oct. 01

BY: Steve Scarabin INSPECTOR



DETECTOR SERVICE CENTER, INC.
P.O. Box 1044
Cary, NC 27512

Steve Scarabin
Randy Watts
(919) 469-0259

CERTIFICATE OF INSPECTION

CUSTOMER _____

ADDRESS _____

ATTENTION: _____

DETECTOR TYPE: N163

SERIAL NO. F1127

DATE OF TEST: 4-5-01

AMOUNT OF
CONTAMINATION

To Be Completed By DSC

HOUSING: $1.16 \times 10^{-4} \mu\text{Ci}$

ENTRANCE: $2.6 \times 10^{-4} \mu\text{Ci}$

EXIT: $1.3 \times 10^{-4} \mu\text{Ci}$

WITHIN NORMAL LIMIT:

YES

NO

NEXT WIPE TEST REQUIRED: Oct. 01

BY: Steve Scarabin INSPECTOR