

From:
To:
Date:
Subject:

Exp
Doddccv@
Jan Barnes
Tue, Jul 25, 2000 10:51 AM
Phase shifts

Ian:

The values for the different notches with the 100% set to 40 are:

	300 kHz	400 kHz
100	40	40
60	27	23
40	23	19
20	15	11

These can vary a degree or so. A number of calibrations on different standards should be done to average out the results. Talk to you later.

Caius

CC:

<WLS@nrc.gov>

Information in this record was deleted
in accordance with the Freedom of Information
Act, exemptions
FOIA- 2001-0256

6/47