

Constellation Nuclear

Calvert Cliffs Nuclear Power Plant

*A Member of the
Constellation Energy Group*

March 13, 2002

U. S. Nuclear Regulatory Commission
Washington, DC 20555

ATTENTION: Document Control Desk

SUBJECT: Calvert Cliffs Nuclear Power Plant
Unit Nos. 1 & 2; Docket Nos. 50-317 & 50-318
10 CFR Part 26 Report; Fitness for Duty

This letter is to report an unsatisfactory performance testing result that was conducted by Quest Diagnostics Incorporated, Norristown, PA, as required by 10 CFR Part 26, Appendix A, Subpart 2.8(e)(4). One blind performance test urine sample containing amphetamine and methamphetamine was reported as negative on February 16, 2002.

Quest Diagnostics Incorporated was contacted on February 19, 2002, and reported that the specimen tested negative on initial screening. The specimen was re-assayed by the initial test (immunoassay) and was assayed by GC/MS confirmed method, and confirmed positive for amphetamine and methamphetamine. The Corrective Action Report from Quest Diagnostics Incorporated is attached.

The following attachments were gathered in our investigation:


- (1) Certificate of Analysis from ElSohly Laboratories, Incorporated dated February 19, 2002, for the amphetamine and methamphetamine Batch A-1955.
- (2) Corrective Action Report from Quest Diagnostics Incorporated, dated February 28, 2002

Questions about this incident can be addressed to Ms. Patricia A. Harrison, FFD Program Administrator, at 410-495-3751.

1021

Should you have questions regarding this matter, we will be pleased to discuss them with you.

Very truly yours,

A handwritten signature in black ink, appearing to read "Gary Dieblough". The signature is fluid and cursive, with the first name "Gary" being more legible than the last name "Dieblough".

LLW/TER/bjd

Attachments: As stated

cc: **(Without Attachments)**
R. S. Fleishman, Esquire
J. E. Silberg, Esquire
Director, Project Directorate I-1, NRC
D. M. Skay, NRC

H. J. Miller, NRC
Resident Inspector, NRC
R. I. McLean, DNR