

SEP 4 1969

Jack R. Roeder, Chief, Materials Inspection
and Enforcement Branch, Division of Compliance, HQ

COMPLIANCE INQUIRY MEMORANDUM
WESTINGHOUSE ELECTRIC CORPORATION
BOX 2287
PITTSBURGH, PENNSYLVANIA 15230
LICENSE NO. 37-09442-01
FILM BADGE EXPOSURE OF 17.4 REM

On August 18, 1969 a CIM was sent to you advising you that this office would take no further action in connection with the reported exposure of 17.4 rem which, on the basis of the licensee's reported information at that time indicated that the film badge had been inadvertently left in close proximity to a 30 Ci Co-60 source during calibration of survey instruments. On August 28, 1969 CO:I received a copy of a letter dated August 26, 1969 from the licensee to the Director, Division of Compliance, transmitting a copy of the report of exposure indicating that a valid exposure had occurred. CO:I will investigate the circumstances of the exposure, the investigation scheduled for September 9, 1969.

Alvin F. Ryan
Investigation Specialist

OFFICE ▶	CO	CO				
SURNAME ▶	Ryan:en	Nelson				
DATE ▶	9/3/69					