

U. S. ATOMIC ENERGY COMMISSION
BYPRODUCT MATERIAL LICENSE
Supplementary Sheet

Page 1 of 2 Pages

License Number 37-09442-01
Amendment No. 10

Westinghouse Electric Corporation
Box 2278
Pittsburgh, Pennsylvania 15230

In accordance with letter dated February 18, 1971, License Number 37-09442-01 is amended as follows:

Condition 14. is amended to read:

14. A. Each sealed source acquired from another person and containing byproduct material, other than Hydrogen 3, with a half-life greater than thirty days and in any form other than gas shall be tested for contamination and/or leakage prior to use. In the absence of a certificate from a transferor indicating that a test has been made within six months prior to the transfer, the sealed source shall not be put into use until tested.
- B. Each sealed source fabricated by the licensee shall be inspected and tested for construction defects, leakage, and contamination prior to use or transfer as a sealed source. If the inspection or test reveals any construction defects or 0.005 microcurie or greater of contamination, the source shall not be used or transferred as a sealed source until it has been repaired and decontaminated.
- C. (1) Each sealed source containing byproduct material, other than Hydrogen 3, with a half-life greater than thirty days and in any form other than gas shall be tested for leakage and/or contamination at intervals not to exceed six months except that each source designed for the purpose of emitting alpha particles shall be tested at intervals not to exceed three months.
- (2) Notwithstanding the periodic leak test required by this condition, any licensed sealed source is exempt from such leak tests when the source contains 100 microcuries or less of beta and/or gamma emitting material or 10 microcuries or less of alpha emitting material.
- D/20

U. S. ATOMIC ENERGY COMMISSION
BYPRODUCT MATERIAL LICENSE

Page 2 of 2 Pages

Supplementary Sheet

License Number 37-09442-01
Amendment No. 10

14. continued

- C. (3) Except for alpha sources, the periodic leak test required by this condition does not apply to sealed sources that are stored and not being used. The sources excepted from this test shall be tested for leakage prior to any use or transfer to another person unless they have been leak tested within six months prior to the date of use or transfer.
- D. The test shall be capable of detecting the presence of 0.005 microcurie of radioactive material on the test sample. The test sample shall be taken from the sealed source or from the surfaces of the device in which the sealed source is permanently or semipermanently mounted or stored on which one might expect contamination to accumulate. Records of leak test results shall be kept in units of microcuries and maintained for inspection by the Commission.
- E. If the test required by Subsection A. or C. of this condition reveals the presence of 0.005 microcurie or more of removable contamination, the licensee shall immediately withdraw the sealed source from use and shall cause it to be decontaminated and repaired or to be disposed of in accordance with Commission regulations. A report shall be filed within 5 days of the test with the Director, Division of Materials Licensing, U. S. Atomic Energy Commission, Washington, D. C., 20545, describing the equipment involved, the test results, and the corrective action taken. A copy of such report shall also be sent to the Director, Region I, Division of Compliance, USAEC, 970 Broad Street, Newark, New Jersey, 07102.

For the U. S. Atomic Energy Commission

Special Agent in Charge
Robert E. Brinkman
Materials Branch

by

Division of Materials Licensing
Washington, D. C. 20545

Date MAR 5 1971