

CHAPTER 219. STANDARDS FOR PROTECTION AGAINST RADIATION

Subch.	Sec.
A. GENERAL PROVISIONS	219.1
B. [Reserved]	219.21
C. [Reserved]	219.31
D. RADIATION DOSE LIMITS FOR INDIVIDUAL MEMBERS OF THE PUBLIC	219.51
E. TESTING FOR LEAKAGE OR CONTAMINATION OF SEALED SOURCES	219.61
F. [Reserved]	219.71
G. [Reserved]	219.91
H. [Reserved]	219.111
I. STORAGE AND CONTROL OF LICENSED OR REGISTERED SOURCES OF RADIATION	219.131
J. PRECAUTIONARY PROCEDURES	219.151
K. [Reserved]	219.181
L. [Reserved]	219.201
M. REPORTS	219.221
N. [Reserved]	219.241

Authority

The provisions of this Chapter 219 issued and amended under sections 301 and 302 of the Radiation Protection Act (35 P. S. §§ 7110.301 and 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20), unless otherwise noted.

Source

The provisions of this Chapter 219 adopted December 18, 1987, effective December 19, 1987, 17 Pa.B. 5235, unless otherwise noted.

Cross References

This chapter cited in 25 Pa. Code § 215.32 (relating to exemption qualifications); 25 Pa. Code § 217.1 (relating to purpose and scope); 25 Pa. Code § 217.144 (relating to incidental radioactive material produced by a particle accelerator); 25 Pa. Code § 220.2 (relating to posting of notices to workers); 25 Pa. Code § 224.1 (relating to purpose and scope); 25 Pa. Code § 225.1 (relating to purpose and scope); 25 Pa. Code § 225.74 (relating to training and testing); 25 Pa. Code § 225.84 (relating to operating and emergency procedures); 25 Pa. Code § 225.91 (relating to radiation survey meter requirements); 25 Pa. Code § 226.1 (relating to purpose and scope); 25 Pa. Code § 228.31a (relating to limitations); 25 Pa. Code § 228.33a (relating to facility and shielding requirements); 25 Pa. Code § 228.38 (relating to radiation safety surveys); 25 Pa. Code § 232.1 (relating to purpose and scope); 25 Pa. Code § 236.15 (relating to protection of individuals during operations); 25 Pa. Code § 236.208 (relating to specific technical information); 25 Pa. Code § 236.209 (relating to technical analyses); 25 Pa. Code § 236.225 (relating to requirements for issuance of license); and 25 Pa. Code § 236.403 (relating to facility operation plan).

219-1

(282367) No. 324 Nov. 01

Subchapter A. GENERAL PROVISIONS

Sec.	
219.1.	Purpose.
219.2.	Scope.
219.3.	Definitions.
219.4.	[Reserved].
219.5.	Incorporation by reference.
219.6.	Effect of incorporation of 10 CFR Part 20.
219.7.	Effect of incorporation of 10 CFR 20.1403 "Criteria for license termination under restricted conditions."
219.11—219.15.	[Reserved].

GENERAL PROVISIONS**§ 219.1. Purpose.**

(a) This chapter establishes standards for protection against ionizing radiation resulting from activities conducted under licenses or registrations issued by the Department. Licensees and registrants shall comply with this chapter.

(b) The requirements of this chapter are designed to control the receipt, possession, use, transfer and disposal of sources of radiation by a licensee or registrant so the total dose to an individual, including doses resulting from all sources of radiation other than background radiation, does not exceed the standards for protection against radiation prescribed in this chapter. This chapter does not limit actions that may be necessary to protect health and safety.

Source

The provisions of this § 219.1 amended November 17, 1995, effective November 18, 1995, 25 Pa.B. 5085. Immediately preceding text appears at serial page (170385).

§ 219.2. Scope.

Except as specifically provided in other chapters of this article, this chapter applies to persons licensed or registered by the Department to receive, possess, use, transfer or dispose of sources of radiation. The limits in this chapter do not apply to doses due to background radiation, to exposure of patients to radiation for the purpose of medical diagnosis or therapy or to voluntary participation in medical research programs.

Source

The provisions of this § 219.2 amended November 17, 1995, effective November 18, 1995, 25 Pa.B. 5085. Immediately preceding text appears at serial pages (170385) to (170386).

§ 219.3. Definitions.

The following term, when used this subchapter, has the following meaning, unless the context clearly indicates otherwise:

Medical reportable event for radiation—producing machine therapy—The administration to a human being, except for an administration resulting from a direct intervention of a patient that could not have been reasonably prevented by the licensee or registrant, that results in one of the following:

- (i) An administration of a therapeutic radiation dose to the wrong individual.
- (ii) An administration of a dose for therapy when the result is an increase in the total expected doses inside or outside of the intended treatment volume for organs, tissue or skin that exceeds 20% of the total prescribed dose for the intended target volume.
- (iii) A total dose delivered to the treatment site identified in a written directive for therapy that differs from the total prescribed dose by more than 20%, or for a fractionated dose, when the weekly administered dose differs from the weekly prescribed dose by more than 30%.

Source

The provisions of this § 219.3 adopted November 17, 1995, effective November 18, 1995, 25 Pa.B. 5085; amended September 14, 2001, effective September 15, 2001, 31 Pa.B. 5239. Immediately preceding text appears at serial pages (204035) to (204037).

§ 219.4. [Reserved].

Source

The provisions of this § 219.4 adopted November 17, 1995, effective November 18, 1995, 25 Pa.B. 5085; reserved September 14, 2001, effective September 15, 2001, 31 Pa.B. 5239. Immediately preceding text appears at serial page (204037).

§ 219.5. Incorporation by reference.

- (a) Except as provided in this chapter, the requirements of 10 CFR Part 20 (relating to standards for protection against radiation) are incorporated by reference.
- (b) Notwithstanding the requirements incorporated by reference, 20.1006, 20.1009, 20.2206(a)(1), (3), (4) and (5), 20.2401 and 20.2402 are not incorporated by reference.

Source

The provisions of this § 219.5 adopted September 14, 2001, effective September 15, 2001, 31 Pa.B. 5239.

§ 219.6. Effect of incorporation of 10 CFR Part 20.

To reconcile differences between this chapter and the incorporated sections of 10 CFR Part 20 (relating to standards for protection against radiation), the following words and phrases shall be substituted for the language in 10 CFR Part 20 as follows:

- (1) A reference to “NRC” or “Commission” means Department.

(2) A reference to "NRC or agreement state" means Department, NRC or agreement state.

(3) A reference to "licensee" includes registrant.

(4) A reference to "license" includes registration.

(5) A reference to "licensed" includes registered.

(6) A reference to "Department" in 10 CFR means the United States Department of Energy.

(7) Notifications, reports and correspondence referenced in the incorporated parts of 10 CFR shall be directed to the Department and, for NRC licenses, to the NRC until agreement state status is in effect.

Source

The provisions of this § 219.6 adopted September 14, 2001, effective September 15, 2001, 31 Pa.B. 5239.

§ 219.7. Effect of incorporation of 10 CFR 20.1403 "Criteria for license termination under restricted conditions."

The Department will not terminate a license under the conditions of restricted release as provided for in 10 CFR 20.1403 (relating to criteria for license termination under restricted conditions) until a license termination plan (LTP), approved by the Department, has been in effect for a period of time sufficient to demonstrate to the Department that continued implementation of the plan will be effective in maintaining compliance with the required conditions of the plan. The Department may choose to implement the license termination process in one or more of the following steps:

(1) The license is amended to authorize activities necessary to begin decommissioning under the LTP.

(2) After decommissioning activities are complete and the provisions of 10 CFR 20.1403 are in effect under the LTP, the license may be amended to end authorization of licensed activities. The license shall remain in effect for up to 5 years being limited to ownership/possession of the decommissioned material.

(3) At the end of the period prescribed in paragraph (2), the Department will make a determination of the effectiveness of the LTP as enacted. If the LTP has demonstrated the ability to maintain compliance with 10 CFR 20.1403, the license will be terminated subject to the revisitation provision of 10 CFR 20.1401(c) (relating to general provision and scope) regarding new evidence of a significant threat to health and safety. Otherwise, the licensee will be directed by the Department to take corrective actions as necessary to conform to 10 CFR 20.1403 and the process shall revert back to paragraph (2).

Source

The provisions of this § 219.7 adopted September 14, 2001, effective September 15, 2001, 31 Pa.B. 5239.

§§ 219.11—219.15. [Reserved].**Source**

The provisions of these §§ 219.11—219.15 reserved November 17, 1995, effective November 18, 1995, 25 Pa.B. 5085. Immediately preceding text appears at serial pages (170386) to (170393).

Subchapter B. [Reserved]**§ 219.21. [Reserved].****Source**

The provisions of this § 219.21 amended November 17, 1995, effective November 18, 1995, 25 Pa.B. 5085; reserved September 14, 2001, effective September 15, 2001, 31 Pa.B. 5239. Immediately preceding text appears at serial pages (204038) and (249251).

§ 219.22. [Reserved].**Source**

The provisions of this § 219.22 adopted December 18, 1987, effective December 19, 1987, 17 Pa. B. 5235; amended June 19, 1992, effective June 20, 1992, 22 Pa. B. 3135; reserved November 17, 1995, effective November 18, 1995, 25 Pa.B. 5085. Immediately preceding text appears at serial pages (170394) to (170396).

Subchapter C. [Reserved]**§§ 219.31—219.33. [Reserved].****Source**

The provisions of these §§ 219.31—219.33 amended November 17, 1995, effective November 18, 1995, 25 Pa.B. 5085; reserved September 14, 2001, effective September 15, 2001, 31 Pa.B. 5239. Immediately preceding text appears at serial pages (249251) to (249252) and (204041) to (204042).

§§ 219.34—219.38. [Reserved].**Source**

The provisions of these §§ 219.34—219.38 adopted November 17, 1995, effective November 18, 1995, 25 Pa.B. 5085; reserved September 14, 2001, effective September 15, 2001, 31 Pa.B. 5239. Immediately preceding text appears at serial pages (204042), (249253) to (249254) and (204045) to (204048).

§§ 219.41—219.49. [Reserved].**Source**

The provisions of these §§ 219.41—219.49 reserved November 17, 1995, effective November 18, 1995, 25 Pa.B. 5085. Immediately preceding text appears at serial pages (170398) to (170404).

**Subchapter D. RADIATION DOSE LIMITS FOR
INDIVIDUAL MEMBERS OF THE PUBLIC**

Sec.

219.51. Dose limits for individual members of the public.

219.52. [Reserved].

§ 219.51. Dose limits for individual members of the public.

In addition to incorporation by reference of 10 CFR Part 20 Subpart D (relating to dose limits for individual members of the public), registrants who met the previous limit (5 mSv or 0.5 REM in 1 year) for locations having existing radiation-producing machines or equipment or other registered radiation sources will not be required to retrofit installations existing before November 18, 1995. The Department does not require the retrofitting of shielding for the replacement of equipment in the facility as long as the equipment is being replaced with similar equipment.

Source

The provisions of this § 219.51 amended November 17, 1995, effective November 18, 1995, 25 Pa.B. 5085; amended September 14, 2001, effective September 15, 2001, 31 Pa.B. 5239. Immediately preceding text appears at serial pages (204048) and (249255).

Cross References

This section cited in 25 Pa. Code § 223.7 (relating to structural shielding); 25 Pa. Code § 227.12a (relating to area requirements); 25 Pa. Code § 228.32a (relating to shielding and safety design requirements); 25 Pa. Code § 228.38 (relating to radiation safety surveys); and 25 Pa. Code § 228.44 (relating to ventilation systems).

§ 219.52. [Reserved].**Source**

The provisions of this § 219.52 amended November 17, 1995, effective November 18, 1995, 25 Pa.B. 5085; reserved September 14, 2001, effective September 15, 2001, 31 Pa.B. 5239. Immediately preceding text appears at serial page (249256).

**Subchapter E. TESTING FOR LEAKAGE OR
CONTAMINATION OF SEALED SOURCES**

Sec.

219.61. Testing for leakage or contamination of sealed sources.

219.62—219.66. [Reserved].

§ 219.61. Testing for leakage or contamination of sealed sources.

(a) In addition to incorporation by reference of 10 CFR Part 20 (relating to standards for protection against radiation), a licensee possessing a sealed source shall assure that:

(1) Except as specified in subsection (b), each sealed source is tested for leakage or contamination and the test results are received before the sealed source is put into use unless the licensee has a certificate from the transferor indicating that the sealed source was tested within 6 months before transfer to the licensee.

(2) Each sealed source that is not designed to emit alpha particles is tested for leakage or contamination at intervals not to exceed 6 months or at alternative intervals approved by the Department under §§ 217.81—217.93 (Reserved), a licensing state or the NRC, except that the maximum interval between leak tests may not exceed 3 years.

(3) Each sealed source that is designed to emit alpha particles is tested for leakage or contamination at intervals not to exceed 3 months or at alternative intervals approved by the Department under §§ 217.81—217.93 (Reserved), an agreement state, a licensing state or the NRC, except that the maximum interval between leak tests may not exceed 3 years.

(4) For each sealed source that is required to be tested for leakage or contamination, the sealed source is tested for leakage or contamination before further use at any time there is reason to suspect that the sealed source might have been damaged or might be leaking.

(5) Except for brachytherapy sources manufactured to contain radium, tests for leakage for sealed sources shall be capable of detecting the presence of 185 Bq (0.005 μ Ci) of radioactive material on a test sample. Test samples shall be taken from the sealed source or from the surfaces of the container in which the sealed source is stored or mounted on which one might expect contamination to accumulate. For a sealed source contained in a device, test samples are obtained when the source is in the "off" position.

(6) The test for leakage for brachytherapy sources manufactured to contain radium shall be capable of detecting an absolute leakage rate of 37 Bq (0.001 μ Ci) of radon-222 in a 24-hour period when the collection efficiency for radon-222 and its progeny has been determined with respect to collection method, volume and time.

(7) Tests for contamination from radium progeny shall be taken on the interior surface of brachytherapy source storage containers and shall be capable of detecting the presence of 185 Bq (0.005 μ Ci) of any radium progeny which has a half-life greater than 4 days.

(b) A licensee need not perform tests for leakage or contamination on the following sealed sources:

- (1) Sealed sources containing only radioactive material with a half-life of less than 30 days.
- (2) Sealed sources containing only radioactive material as a gas.
- (3) Sealed sources containing 3.7 MBq (100 µCi) or less of beta or photon-emitting material or 370 kBq (10 µCi) or less of alpha-emitting material.
- (4) Sealed sources containing only hydrogen-3.
- (5) Seeds of iridium-192 encased in nylon ribbon.
- (6) Sealed sources, except teletherapy and brachytherapy sources, which are stored, are not being used, and are identified as in storage. The licensee shall, however, test each of these sealed sources for leakage or contamination and receive the test results before any use or transfer unless it has been tested for leakage or contamination within 6 months before the date of use or transfer. The maximum interval between tests for leakage or contamination may not exceed 3 years.
- (c) Tests for leakage or contamination from sealed sources shall be performed by persons specifically authorized by the Department, an agreement state, a licensing state or the NRC to perform these services.
- (d) Test results shall be kept in units of becquerel or microcurie and maintained for inspection by the Department.
- (e) The following shall be considered evidence that a sealed source is leaking:
 - (1) The presence of 185 Bq (0.005 µCi) or more of removable contamination on any test sample.
 - (2) Leakage of 37 Bq (0.001 µCi) of radon-222 per 24 hours for brachytherapy sources manufactured to contain radium.
 - (3) The presence of removable contamination resulting from the decay of 185 Bq (0.005 µCi) or more of radium.
- (f) The licensee shall immediately withdraw a leaking sealed source from use and take action to prevent the spread of contamination. The leaking sealed source shall be repaired or disposed of in accordance with this article.
- (g) Reports of test results for leaking or contaminated sealed sources shall be made under § 219.227 (relating to reports of leaking or contaminated sealed sources).

Source

The provisions of this § 219.61 adopted June 19, 1992, effective June 20, 1992, 22 Pa.B. 3135; amended November 17, 1995, effective November 18, 1995, 25 Pa.B. 5085; amended September 14, 2001, effective September 15, 2001, 31 Pa.B. 5239. Immediately preceding text appears at serial pages (249256) and (204051) to (204052).

Cross References

This section cited in 25 Pa. Code § 219.227 (relating to reports of leaking or contaminated sealed sources).

§§ 219.62—219.66. [Reserved].**Source**

The provisions of these §§ 219.62—219.66 adopted June 19, 1992, effective June 20, 1992, 22 Pa.B. 3135; reserved November 17, 1995, effective November 18, 1995, 25 Pa.B. 5085. Immediately preceding text appears at serial pages (170407) to (170410).

Subchapter F. [Reserved]**§ 219.71. [Reserved].****Source**

The provisions of this § 219.71 adopted December 18, 1987, effective December 19, 1987, 17 Pa.B. 3135; amended June 19, 1992, effective June 20, 1992, 22 Pa.B. 3135; amended November 17, 1995, effective November 18, 1995, 25 Pa.B. 5085; reserved September 14, 2001, effective September 15, 2001, 31 Pa.B. 5239. Immediately preceding text appears at serial page (204053).

§ 219.72. [Reserved].**Source**

The provisions of this § 219.72 amended November 17, 1995, effective November 18, 1995, 25 Pa.B. 5085; reserved September 14, 2001, effective September 15, 2001, 31 Pa.B. 5239. Immediately preceding text appears at serial page (204054).

§ 219.73. [Reserved]**Source**

The provisions of this § 219.73 amended November 17, 1995, effective November 18, 1995, 25 Pa.B. 5085; reserved September 14, 2001, effective September 15, 2001, 31 Pa.B. 5239. Immediately preceding text appears at serial pages (204054) and (249257).

§§ 219.74—219.76. [Reserved].**Source**

The provisions of these §§ 219.74—219.76 reserved November 17, 1995, effective November 18, 1995, 25 Pa.B. 5085. Immediately preceding text appears at serial pages (170414) to (170415).

§ 219.81. [Reserved].**Source**

The provisions of this § 219.81 reserved November 17, 1995, effective November 18, 1995, 25 Pa.B. 5085. Immediately preceding text appears at serial pages (170415) to (170417).

Subchapter G. [Reserved]

§§ 219.91—219.93. [Reserved].

Source

The provisions of these §§ 219.91—219.93 adopted November 17, 1995, effective November 18, 1995, 25 Pa.B. 5085; reserved September 14, 2001, effective September 15, 2001, 31 Pa.B. 5239. Immediately preceding text appears at serial pages (249257) to (249258) and (204057) to (204059).

Subchapter H. [Reserved]

§§ 219.111—219.113. [Reserved].

Source

The provisions of these §§ 219.111—219.113 adopted November 17, 1995, effective November 18, 1995, 25 Pa.B. 5085; reserved September 14, 2001, effective September 15, 2001, 31 Pa.B. 5239. Immediately preceding text appears at serial pages (204060) to (204062).

**Subchapter I. STORAGE AND CONTROL OF
LICENSED OR REGISTERED SOURCES OF
RADIATION**

Sec.

- 219.131. Security of stored sources of radiation.
- 219.132. Control of sources of radiation not in storage.

Source

The provisions of this Subchapter I adopted November 17, 1995, effective November 18, 1995, 25 Pa.B. 5085; amended September 14, 2001, effective September 15, 2001, 31 Pa.B. 5239, unless otherwise noted. Immediately preceding text appears at serial page (204062).

§ 219.131. Security of stored sources of radiation.

In addition to incorporation by reference of 10 CFR Part 20 (relating to standards for protection against radiation), the licensee or registrant shall secure from unauthorized removal or access radiation sources that are in storage.

§ 219.132. Control of sources of radiation not in storage.

In addition to incorporation by reference of 10 CFR Part 20 (relating to standards for protection against radiation), the licensee or registrant shall maintain control of radiation producing machines that are not in storage.

Subchapter J. PRECAUTIONARY PROCEDURES

Sec.	
219.151—219.158.	[Reserved].
219.159.	Posting of radiation-producing machines.
219.160.	Exceptions to posting requirements.
219.161.	Exemptions from labeling requirements.
219.162.	Procedures for receiving and opening packages.

Source

The provisions of this Subchapter J adopted November 17, 1995, effective November 18, 1995, 25 Pa.B. 5085, unless otherwise noted.

§§ 219.151—219.158. [Reserved].**Source**

The provisions of these §§ 219.151—219.158 adopted November 17, 1995, effective November 18, 1995, 25 Pa.B. 5085; reserved September 14, 2001, effective September 15, 2001, 31 Pa.B. 5239. Immediately preceding text appears at serial pages (249259) to (249260) and (204065).

§ 219.159. Posting of radiation-producing machines.

The registrant or licensee shall ensure that each radiation producing machine is labeled in a conspicuous manner which cautions individuals that radiation is produced when it is energized. For example:

**“CAUTION—RADIATION
THIS EQUIPMENT PRODUCES RADIATION
WHEN ENERGIZED.”**

Source

The provisions of this § 219.159 amended September 14, 2001, effective September 15, 2001, 31 Pa.B. 5239. Immediately preceding text appears at serial page (204065).

§ 219.160. Exceptions to posting requirements.

In addition to incorporation by reference of 10 CFR Part 20 (relating to standards for protection against radiation), a room or area is not required to be posted with a caution sign because of the presence of radiation machines used solely for diagnosis in the healing arts.

Source

The provisions of this § 219.160 amended September 14, 2001, effective September 15, 2001, 31 Pa.B. 5239. Immediately preceding text appears at serial pages (204065) to (204066).

§§ 219.161 and 219.162. [Reserved].**Source**

The provisions of these §§ 219.161 and 219.162 reserved September 14, 2001, effective September 15, 2001, 31 Pa.B. 5239. Immediately preceding text appears at serial pages (204066) to (204068).

Subchapter K. [Reserved]**§§ 219.181—219.186. [Reserved].****Source**

The provisions of these §§ 219.181—219.186 adopted November 17, 1995, effective November 18, 1995, 25 Pa.B. 5085; reserved September 14, 2001, effective September 15, 2001, 31 Pa.B. 5239. Immediately preceding text appears at serial pages (204068) to (204070) and (249261).

Subchapter L. [Reserved]**§§ 219.201—219.211. [Reserved].****Source**

The provisions of these §§ 219.201—219.211 adopted November 17, 1995, effective November 18, 1995, 25 Pa.B. 5085; reserved September 14, 2001, effective September 15, 2001, 31 Pa.B. 5239. Immediately preceding text appears at serial pages (249261) to (249266).

Subchapter M. REPORTS

Sec.	
219.221.	Reports of stolen, lost or missing licensed or registered sources of radiation.
219.222.	Notification of incidents and reportable events.
219.223—219.226.	[Reserved].
219.227.	Reports of leaking or contaminated sealed sources.
219.228.	Reports of medical reportable events for radiation-producing machine therapy.
219.229.	Other medical reports.

Source

The provisions of this Subchapter M adopted November 17, 1995, effective November 18, 1995, 25 Pa.B. 5085, unless otherwise noted.

§ 219.221. Reports of stolen, lost or missing licensed or registered sources of radiation.

In addition to incorporation by reference of the requirements in 10 CFR Part 20 (relating to standards for protection against radiation) covering the reporting

requirements associated with reports of theft or loss of licensed material, the following reporting requirements apply to radiation-producing machines:

(1) *Telephone reports.* Each licensee or registrant shall report to the Department by telephone immediately, after its occurrence becomes known, a stolen, lost or missing radiation producing machine.

(2) *Written reports.* Each licensee or registrant required to make a report under paragraph (1) shall, within 30 days after making the telephone report, make a written report to the Department setting forth the following information:

(i) A description of the licensed or registered source of radiation involved, including, for radiation producing machines, the manufacturer, model and serial number, type and maximum energy of radiation emitted.

(ii) A description of the circumstances under which the loss or theft occurred.

(iii) A statement of disposition, or probable disposition, of the licensed or registered source of radiation involved.

(iv) Exposures of individuals to radiation, circumstances under which the exposures occurred and the possible total effective dose equivalent to persons in unrestricted areas.

(v) Actions that have been taken, or will be taken, to recover the source of radiation.

(vi) Procedures or measures that have been, or will be, adopted to ensure against a recurrence of the loss or theft of licensed or registered sources of radiation.

(3) *Additional information.* Subsequent to filing the written report, the licensee or registrant shall also report additional substantive information on the loss or theft within 30 days after the licensee or registrant learns of the information.

(4) *Detachable reports.* The licensee or registrant shall prepare a report filed with the Department under this section so that the names of individuals who may have received exposure to radiation are stated in a separate and detachable portion of the report.

Source

The provisions of this § 219.221 amended September 14, 2001, effective September 15, 2001, 31 Pa.B. 5239. Immediately preceding text appears at serial pages (249266) to (249267) and (204077).

Cross References

This section cited in 25 Pa. Code § 225.76 (relating to reporting requirements).

§ 219.222. Notification of incidents and reportable events.

In addition to incorporation by reference of the requirements in 10 CFR 20.2202 and 20.2203 (relating to notification of incidents; and reports of expo-

tures, radiation levels and concentrations of radioactive material exceeding the constraints or limits), those notification requirements, as well as written 30-day reports under 10 CFR 20.2203(a), also apply to radiation-producing machines and NARM.

Source

The provisions of this 219.222 amended September 14, 2001, effective September 15, 2001, 31 Pa.B. 5239. Immediately preceding text appears at serial pages (204077) to (204078).

Cross References

This section cited in 25 Pa. Code § 225.76 (relating to reporting requirements).

§§ 219.223—219.226. [Reserved].

Source

The provisions of these §§ 219.223—219.226 reserved September 14, 2001, effective September 15, 2001, 31 Pa.B. 5239. Immediately preceding text appears at serial pages (204078) to (204080) and (249269).

§ 219.227. Reports of leaking or contaminated sealed sources.

If the test for leakage or contamination, required under § 219.61 (relating to testing for leakage or contamination of sealed sources), indicates a sealed source is leaking or contaminated, a report of the test shall be filed within 5 days with the Department describing the equipment involved, the test results and the corrective action taken.

Cross References

This section cited in 25 Pa. Code § 219.61 (relating to testing for leakage or contamination of sealed sources).

§ 219.228. Reports of medical reportable events for radiation-producing machine therapy.

(a) For a medical reportable event for radiation-producing machine therapy, the licensee or registrant shall do the following:

- (1) Notify the Department by telephone within 24 hours after discovery of the event.
- (2) Submit a written report to the Department within 15 days after discovery of the event. The written report shall include the licensee's or registrant's name; the prescribing physician's name; a brief description of the event; why the event occurred; the effect on the patient; what improvements are needed to prevent recurrence; actions taken to prevent recurrence; whether the licensee or registrant notified the patient, or the patient's responsible relative or guardian (for notification purposes under this section, this person will be included in subsequent references to "the patient"), and if not, why not; and if the patient

was notified, what information was provided to the patient. The report may not include the patient's name or other information that could lead to identification of the patient.

(3) Notify the referring physician and also notify the patient of the event within 24 hours after its discovery, unless the referring physician personally informs the licensee either that he will inform the patient or that, based on medical judgment, telling the patient would be harmful. The licensee or registrant is not required to notify the patient without first consulting the referring physician. If the referring physician or patient cannot be reached within 24 hours, the licensee or registrant shall notify the patient as soon as possible thereafter. The licensee or registrant may not delay appropriate medical care for the patient, including necessary remedial care as a result of the misadministration, because of delay in notification.

(4) If the patient was notified, the licensee or registrant shall also furnish, within 15 days after discovery of the event, a written report to the patient by sending one of the following:

(i) A copy of the report that was submitted to the Department.

(ii) A brief description of both the event and the consequences, as they may affect the patient, if a statement is included that the report submitted to the Department can be obtained from the licensee or registrant.

(b) The licensee or registrant shall retain a record of each medical reportable event for radiation-producing machine therapy for 5 years. The record shall contain the names of the individuals involved (including the prescribing physician, allied health personnel, the patient and the patient's referring physician), the patient's Social Security number or identification number if one has been assigned, a brief description of the event, why it occurred, the effect on the patient, what improvements are needed to prevent recurrence and the actions taken to prevent recurrence.

(c) Aside from the notification requirement, this section does not affect rights or duties of licensees or registrants and physicians in relation to each other, patients or the patient's responsible relatives or guardians.

Source

The provisions of this § 219.228 amended September 14, 2001, effective September 15, 2001, 31 Pa.B. 5239. Immediately preceding text appears at serial pages (249269) to (249270).

Cross References

This section cited in 25 Pa. Code § 219.229 (relating to other medical reports); and 25 Pa. Code § 228.35 (relating to operating procedures).

§ 219.229. Other medical reports.

Within 30 days of the discovery of either actual or suspected acute or long-term functional damage to tissue of a patient exposed to therapeutic or diagnostic radiation from a radiation-producing machine, the registrant or licensee shall

document the finding and provide a report to the Department and provide a clinical summary to the prescribing physician and the patient. The report shall be retained for at least 5 years. Exempt from this reporting requirement are any events already reported under § 219.228 (relating to reports of medical reportable events for radiation-producing machine therapy) and any functional damage to patient tissue that was an expected outcome when the causative procedures were prescribed.

Source

The provisions of this § 219.229 adopted September 14, 2001, effective September 15, 2001, 31 Pa.B. 5239.

Subchapter N. [Reserved]

§ 219.241. [Reserved].

Source

The provisions of this § 219.241 adopted November 17, 1995, effective November 18, 1995, 25 Pa.B. 5085; reserved September 14, 2001, effective September 15, 2001, 31 Pa.B. 5239. Immediately preceding text appears at serial page (249270).

APPENDIX A. [Reserved]**Source**

The provisions of this Appendix A amended November 17, 1995, effective November 18, 1995, 25 Pa.B. 5085; reserved September 14, 2001, effective September 15, 2001, 31 Pa.B. 5239. Immediately preceding text appears at serial pages (204084) to (204087).

APPENDIX B. [Reserved]**Source**

The provisions of this Appendix B amended November 17, 1995, effective November 18, 1995, 25 Pa.B. 5085; reserved September 14, 2001, effective September 15, 2001, 31 Pa.B. 5239. Immediately preceding text appears at serial pages (204088) to (204161).

APPENDIX C. [Reserved]**Source**

The provisions of this Appendix C adopted December 18, 1987, effective December 19, 1987, 17 Pa.B. 5235; amended June 19, 1987, effective June 20, 1992, 22 Pa.B. 3135; amended November 17, 1995, effective November 18, 1995, 25 Pa.B. 5085; reserved September 14, 2001, effective September 15, 2001, 31 Pa.B. 5239. Immediately preceding text appears at serial pages (204162) to (204169).

[Next page is 220-1.]

219-18

(282384) No. 324 Nov. 01

Copyright © 2001 Commonwealth of Pennsylvania

CHAPTER 220. NOTICES, INSTRUCTIONS AND REPORTS TO WORKERS; INSPECTIONS AND INVESTIGATIONS

Sec.	
220.1.	Purpose and scope.
220.2.	Posting of notices to workers.
220.3—220.5.	[Reserved].
220.6—220.8.	[Reserved].
220.9.	Incorporation by reference.
220.10.	Effect of incorporation of 10 CFR Part 19.

Authority

The provisions of this Chapter 220 issued and amended under sections 301 and 302 of the Radiation Protection Act (35 P. S. §§ 7110.301 and 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20), unless otherwise noted.

Source

The provisions of this Chapter 220 adopted December 18, 1987, effective December 19, 1987, 17 Pa.B. 5235, unless otherwise noted.

Cross References

This chapter cited in 25 Pa. Code § 215.32 (relating to exemption qualifications); 25 Pa. Code § 217.1 (relating to purpose and scope); 25 Pa. Code § 217.144 (relating to incidental radioactive material produced by a particle accelerator); 25 Pa. Code § 224.1 (relating to purpose and scope); 25 Pa. Code § 225.1 (relating to purpose and scope); 25 Pa. Code § 225.74 (relating to training and testing); 25 Pa. Code § 226.1 (relating to purpose and scope); 25 Pa. Code § 228.31a (relating to limitations); 25 Pa. Code § 232.1 (relating to purpose and scope); and 25 Pa. Code § 240.401 (relating to inspection).

§ 220.1. Purpose and scope.

This chapter establishes requirements for notices, instructions and reports by licensees or registrants to individuals engaged in activities under a license or registration. This chapter also establishes options available to the individuals in connection with Department inspections of licensees or registrants to ascertain compliance with the provisions of the act and regulations, orders and licenses issued thereunder regarding radiological working conditions. This chapter applies to persons who receive, possess, use, own or transfer radiation sources licensed by or registered with the Department under Chapters 216 and 217 (relating to registration of radiation-producing machines; and licensing of radioactive material).

§ 220.2. Posting of notices to workers.

(a) A licensee or registrant shall post current copies of the following documents:

- (1) This chapter and Chapter 219 (relating to standards for protection against radiation).

(2) The license, certificate of registration, conditions or documents incorporated into the license by reference and amendments thereto.

(3) The operating procedures applicable to activities under the license or registration.

(4) A notice of violation involving radiological working conditions, proposed imposition of civil penalty or order issued under Chapter 215 (relating to general provisions) and response from the licensee or registrant.

(b) If posting of a document specified in subsection (a)(1), (2) or (3) is not practicable, the licensee or registrant may post a notice which describes the document and states where it may be examined.

(c) Department Form ER-BRP-3, "Notice to Employees," shall be posted by a licensee or registrant as required by this article.

(d) Department documents posted under subsection (a)(4) shall be posted within 2 working days after receipt of the documents from the Department; the licensee's or registrant's response shall be posted within 2 working days after dispatch from the licensee or registrant. The documents shall remain posted for a minimum of 5 working days or until action correcting the violation has been completed, whichever is later.

(e) Documents, notices or forms posted under this section shall appear in a sufficient number of places to permit individuals engaged in work under the license or registration to observe them on the way to or from the particular work location to which the document applies. The documents, notices or forms shall be conspicuous and shall be replaced if defaced or altered.

Source

The provisions of this § 220.2 amended September 14, 2001, effective September 15, 2001, 31 Pa.B. 5239. Immediately preceding text appears at serial page (249272).

§§ 220.3—220.5. [Reserved].

Source

The provisions of these §§ 220.3—220.5 amended November 17, 1995, effective November 18, 1995, 25 Pa.B. 5085; reserved September 14, 2001, effective September 15, 2001, 31 Pa.B. 5239. Immediately preceding text appears at serial pages (249272) and (203895).

§§ 220.6—220.8. [Reserved].

Source

The provisions of these §§ 220.6—220.8 reserved September 14, 2001, effective September 15, 2001, 31 Pa.B. 5239. Immediately preceding text appears at serial pages (203895) to (203897).

§ 220.9. Incorporation by reference.

(a) Except as provided in this chapter, the requirements of 10 CFR Part 19 (relating to notices, instructions and reports to workers; inspections and investigations) are incorporated by reference.

(b) Notwithstanding the requirements incorporated by reference, 10 CFR 19.4, 19.5, 19.8, 19.30 and 19.40 are not incorporated by reference.

Source

The provisions of this § 220.9 adopted September 14, 2001, effective September 15, 2001, 31 Pa.B. 5239.

§ 220.10. Effect of incorporation of 10 CFR Part 19.

To reconcile differences between this chapter and the incorporated sections of 10 CFR Part 19 (relating to notices, instructions and reports to workers; inspections and investigations), the following words and phrases shall be substituted for the language in 10 CFR Part 19 as follows:

(1) A reference to "NRC" or "Commission" means Department.

(2) A reference to "NRC or agreement state" means Department, NRC or agreement state.

(3) A reference to "license," "licenses," "licensed" and "licensed radioactive material" also include "registration," "registrant" "registered," and "registered source of radiation," respectively.

(4) Notifications, reports and correspondence referenced in the incorporated parts of 10 CFR shall be directed to the Department and, for NRC licenses, to the NRC until agreement state status is in effect.

Source

The provisions of this § 220.10 adopted September 14, 2001, effective September 15, 2001, 31 Pa.B. 5239.

[Next page is 221-1.]

220-4

(282388) No. 324 Nov. 01

Copyright © 2001 Commonwealth of Pennsylvania

CHAPTER 221. X-RAYS IN THE HEALING ARTS**GENERAL PROVISIONS**

- Sec.
221.1. Purpose and scope.
221.2. Definitions.
221.3. Sale and installation.

ADMINISTRATIVE CONTROLS

- 221.11. Registrant responsibilities.
221.12. Records, maintenance and associated information.
221.13. Information to be submitted by persons proposing to conduct healing arts screening.
221.14. [Reserved].
221.15. Use of X-rays in research on humans.

DIAGNOSTIC INSTALLATIONS GENERAL REQUIREMENTS

- 221.21. Diagnostic equipment requirements.
221.22. Battery charge indicator.
221.23. Leakage radiation from the diagnostic source assembly.
221.24. Radiation from components other than the diagnostic source assembly.
221.25. Beam quality.
221.26. Multiple tubes.
221.27. Mechanical support of tube head.
221.28. Technique indicators.
221.29. Kilovoltage (kV) accuracy.
221.30. Exposure reproducibility.
221.31. [Reserved].
221.31a. Locks.
221.32. [Reserved].
221.32a. Radiographic beam limitation.
221.33. [Reserved].
221.33a. Radiation from capacitor energy storage equipment in standby status.
221.34. [Reserved].
221.34a. Radiation exposure control.
221.35. [Reserved].
221.35a. Fluoroscopic X-ray systems.
221.36. [Reserved].
221.36a. Limitation of useful beam of fluoroscopic equipment.
221.37. [Reserved].
221.37a. Activation of fluoroscopic tube.
221.38. [Reserved].
221.38a. Entrance exposure rate.

221-1

- 221.39. [Reserved].
- 221.39a. Barrier transmitted radiation rate limits.
- 221.40. [Reserved].
- 221.40a. Indication of tube voltage and current.
- 221.41. [Reserved].
- 221.41a. Fluoroscopic timer.
- 221.42. [Reserved].
- 221.42a. Control of scattered radiation.
- 221.43. [Reserved].
- 221.43a. Mobile fluoroscopes.
- 221.44—221.49. [Reserved].

INTRAORAL DENTAL RADIOGRAPHIC SYSTEMS

- 221.51—221.55. [Reserved].
- 221.56. [Reserved].

OTHER SYSTEMS

- 221.61. Radiation therapy simulation systems.
- 221.62. [Reserved].

THERAPEUTIC X-RAY SYSTEMS WITH ENERGIES LESS THAN 1 MEV

- 221.71. Equipment requirements.
- 221.72. Facility design requirements for systems capable of operating above 50 kVp.
- 221.73. Surveys.
- 221.74. Calibration.
- 221.75. Spot checks.
- 221.76. Operating procedures.
- 221.81—221.102. [Reserved].

COMPUTED TOMOGRAPHY X-RAY SYSTEMS

- 221.201. Definitions.
- 221.202. Equipment requirements.
- 221.203. Facility design requirements.
- 221.204. Radiation measurements and performance evaluations.
- 221.205. Operating procedures.

Authority

The provisions of this Chapter 221 issued under section 301 of the The Atomic Energy Development and Radiation Control Act (73 P. S. § 1301) (Repealed), unless otherwise noted.

Source

The provisions of this Chapter 221 adopted February 1, 1972, effective February 2, 1972, 2 Pa.B. 212, unless otherwise noted.

Cross References

This chapter cited in 25 Pa. Code § 215.32 (relating to exemption qualifications); 25 Pa. Code § 225.104 (relating to X-ray detection systems for explosives, weapons and illegal items); 28 Pa. Code § 501.4 (relating to regulations); and 28 Pa. Code § 565.12 (relating to radiology service policy).

GENERAL**§ 221.1. Purpose and scope.**

This chapter establishes requirements for the use of X-ray equipment by or under the supervision of a licensed practitioner of the healing arts. A registrant who uses X-rays in the healing arts shall comply with this chapter. This chapter is in addition to, and not in substitution for, other applicable provisions of this article.

Authority

The provisions of this § 221.1 amended under section 302 of the Radiation Protection Act (35 P. S. § 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20).

Source

The provisions of this § 221.1 adopted February 1, 1972, effective February 2, 1972, 2 Pa.B. 212; amended December 18, 1987, effective December 19, 1987, 17 Pa.B. 5235. Immediately preceding text appears at serial pages (117772), (42615) and (4877) to (4880).

Cross References

This section cited in 25 Pa. Code § 221.35a (relating to fluoroscopic X-ray systems).

§ 221.2. Definitions.

The following words and terms, when used in this chapter, have the following meanings, unless the context clearly indicates otherwise:

AAPM—American Association of Physicists in Medicine.

ACR—American College of Radiology.

Aluminum equivalent—The thickness of type 1100 aluminum alloy—the nominal chemical composition of type 1100 aluminum alloy is 99% minimum aluminum, .12% copper—affording the same attenuation, under specified conditions, as the material in question.

Automatic exposure control—A device which automatically controls one or more technique factors in order to obtain at preselected locations a desired quantity of radiation.

Beam axis—A line from the source through the centers of the X-ray fields.

Beam-limiting device—A device providing a means to restrict the dimensions of the X-ray field.

Cephalometric device—A device intended for the radiographic visualization and measurement of the dimensions of the human head.

Certified components—Components of X-ray systems which are subject to regulations promulgated under the Radiation Control for Health and Safety Act of 1968 (42 U.S.C.A. §§ 263b—263n).

Certified system—An X-ray system which has one or more certified components.

Changeable filter—A filter, exclusive of inherent filtration, which can be added to or removed from the useful beam through an electronic, mechanical or physical process.

Coefficient of variation (C)—The ratio of the standard deviation to the mean value of a population of observations. It is estimated using the following equation:

$$C = \frac{s}{\bar{X}} = \frac{1}{\bar{X}} \left[\sum_{i=1}^n \frac{(X_i - \bar{X})^2}{n-1} \right]^{1/2}$$

where

s = Estimated standard deviation of the population.

\bar{X} = Mean value of observations in sample.

X_i = i^{th} observation in sample.

n = Number of observations in sample; $n > 1$.

Contact therapy system—An X-ray system used for therapy with the X-ray tube port placed in contact with or within 5 centimeters of the surface being treated.

Control panel—The part of the X-ray control upon which are mounted the switches, knobs, pushbuttons and other hardware necessary for manually setting the technique factors.

Dead-man switch—A switch so constructed that a circuit closing contact can be maintained only by continuous pressure on the switch by the operator.

Dental panoramic system—A device intended to produce a radiographic image of both dental arches on one film.

Diagnostic source assembly—The tube housing assembly with a beam-limiting device attached.

Diagnostic X-ray system—An X-ray system designed for irradiation of a part of the human body for the purpose of diagnosis or visualization.

Direct scattered radiation—The scattered radiation coming directly from material irradiated by the useful beam and not scattered by other material.

Entrance exposure rate—The exposure in air per unit time at the point where the center of the useful beam enters the patient.

Field emission equipment—Equipment using an X-ray tube in which electrons are emitted from the cathode solely by the force between an electric field and the electrons.

Filter—aterial placed in the useful beam to absorb the less penetrating radiation.

Filtration—Material placed in the useful beam to absorb the less penetrating radiation.

Fluoroscopic imaging assembly—A subsystem in which X-ray photons produce a fluoroscopic image. The term includes the image receptors such as the image intensifier and spot-film device, electrical interlocks, if any, and structural material providing linkage between the image receptor and diagnostic source assembly.

Fluoroscopic system—See fluoroscopic imaging assembly.

Focal spot—The area projected on the anode of the X-ray tube by the electrons accelerated from the cathode and from which the useful beam originates.

Half-value layer (HVL)—The thickness of specified material which attenuates the exposure rate by 1/2 when introduced into the path of a given beam of radiation.

Healing arts screening—The testing of human beings using X-ray machines for the detection or evaluation of health indications when the tests are not specifically and individually ordered by a licensed practitioner of the healing arts legally authorized to prescribe the X-ray tests for the purpose of diagnosis or treatment.

Image intensifier—A device, installed in its housing, which instantaneously converts an X-ray pattern into a corresponding light image of higher energy density.

Image receptor—A device, such as a fluorescent screen or radiographic film, which transforms incident X-ray photons either into a visible image or into another form which can be made into a visible image by further transformations.

Intensifying screen—A fluorescent screen which transforms incident X-ray photons into a visible image.

Intraoral dental radiography—A modality of dental radiography in which the image receptor is placed inside a patient's oral cavity.

kV—Kilovolts

kVp—Peak tube potential (see kilovolts peak).

Kilovolts peak (kVp)—The maximum value of the potential difference across the X-ray tube during an exposure.

Lead equivalent—The thickness of lead affording the same attenuation, under specified conditions, as the material in question.

Leakage radiation—Radiation emanating from the diagnostic or therapeutic source assembly except for the following:

- (i) The useful beam.
- (ii) Radiation produced when the exposure switch or timer is not activated.

Leakage technique factors—The technique factors associated with the tube housing assembly which are used in measuring leakage radiation defined as follows:

- (i) For capacitor energy storage equipment, the maximum-rated peak tube potential and the maximum-rated number of exposures in an hour for operation at the maximum-rated peak tube potential with the charge per exposure being 10 millicoulombs—10 milliamperere seconds—or the minimum charge obtainable from the unit, whichever is larger.
- (ii) For field emission equipment rated for pulsed operation, the maximum-rated peak tube potential and the maximum-rated number of X-ray pulses in an hour for operation at the maximum-rated peak tube potential.
- (iii) For other equipment, the maximum-rated peak tube potential and the maximum-rated continuous tube current for the maximum-rated peak tube potential.

Licensed practitioner of the healing arts—An individual licensed by the Commonwealth to practice the healing arts, which for the purposes of this article shall be limited to medicine, surgery, dentistry, osteopathy, podiatry and chiropractic.

Light field—The area defined by the intersection of the light beam with a plane parallel with the plane of the image receptor. The edge of the field is defined by the points at which the light intensity is 25% of the maximum light intensity in the plane.

Line-voltage regulation—The difference between the no-load and the load line potentials expressed as a percent of the load line potential calculated using the following equation:

$$\text{Percent line-voltage regulation} = 100 (V_n - V_l) / V_l$$

where

V_n = No-load line potential and

V_l = Load line potential.

mA—Milliamperere.

mAs—Milliamperere second.

mR—Milliroentgen.

Maximum line current—The root-mean-square current in the supply line of an X-ray machine operating at its maximum rating.

Mobile X-ray system—see X-ray equipment.

Patient—An individual subjected to healing arts examination, diagnosis or treatment.

Peak tube potential—The maximum value of the potential difference across the X-ray tube during an exposure.

Phototimer—A method for controlling the radiation exposures to an image receptor by measuring the radiation which reaches a radiation monitoring device. The radiation monitoring device is part of an electronic circuit which controls the duration of time the tube is activated.

Portable radiation system—See X-ray equipment.

Position indicating device (PID)—A device on dental X-ray equipment used to indicate the beam position and to establish a definite source-surface (skin) distance.

Positive beam limitation—The automatic or semiautomatic adjustment of an X-ray beam to the size of the selected image receptor, whereby an X-ray exposure cannot be made without an adjustment.

Protective apron—An apron incorporating radiation absorbing materials.

Protective barrier—A barrier of radiation absorbing material used to reduce radiation exposure. The term includes the following types:

- (i) *Primary protective barrier*—Material used to reduce radiation exposure from the useful beam.
- (ii) *Secondary protective barrier*—Material used to reduce exposure from stray, leakage or scattered radiation.

Protective glove—A glove incorporating radiation absorbing materials.

Qualified expert—An individual having the knowledge and training to measure ionizing radiation, to evaluate safety techniques, and to advise regarding radiation protection needs. For example, individuals certified in the appropriate field by the American Board of Radiology, or the American Board of Health Physics, or the American Board of Medical Physics, or those having equivalent qualifications. With reference to the calibration of radiation therapy equipment, an individual having, in addition to the preceding qualifications, training and experience in the clinical applications of radiation physics to radiation therapy. For example, individuals certified in therapeutic radiological physics or X-ray and radium physics by the American Board of Radiology, or radiation oncology physics by the American Board of Medical Physics; or those having equivalent qualifications.

Radiation detector—A device which provides a signal or other indication suitable for measuring one or more quantities of incident radiation.

Radiation therapy simulation system—A radiographic or fluoroscopic X-ray system intended for localizing the volume to be exposed during radiation therapy and confirming the position and size of the therapeutic irradiation field.

Radiograph—An image receptor on which an image is created directly or indirectly by an X-ray pattern and results in a permanent record.

Radiographic imaging system—A system whereby an image is produced on an image receptor by the action of ionizing radiation.

Rating—The operating limits specified by the component manufacturer.

Registrant—A person who is legally obligated to register with the Department under this article and the act.

Research—One of the following:

- (i) Theoretical analysis, exploration or experimentation.
- (ii) The extension of investigative findings and theories of a scientific or technical nature into practical application for experimental and demonstration purposes, including the experimental testing of models, devices, equipment, materials and processes. The term includes the external administration of X-ray radiation to human beings for diagnostic or therapeutic purposes or in an equivalent manner as a diagnostic or therapeutic procedure.

Response time—The time required for an instrument system to reach 90% of its final reading when the instrument system is exposed to a step change from zero radiation flux to a flux sufficient to provide a steady state midscale reading.

SSD—The distance between the source and the skin of the patient.

SID—Source-image receptor distance—The distance from the source to the center of the input surface of the image receptor.

Scattered radiation—Radiation that, during passage through matter, has been deviated in direction.

Screening—See the definition of “healing arts screening.”

Serial radiography—Radiographic images produced in regular sequence.

Shutter—A device attached to the tube housing assembly which can totally intercept the useful beam and which has a lead equivalency not less than that of the tube housing assembly.

Source—The focal spot of the X-ray tube.

Specific prescription—A written or oral directive authorizing a radiographic or fluoroscopic examination of a specified individual.

Spot check—A procedure to assure that a previous calibration continues to be valid.

Spot film—A radiograph which is made during a fluoroscopic examination to permanently record conditions which exist during that fluoroscopic procedure.

Spot-film device—A device intended to transport or position a radiographic image receptor between the X-ray source and fluoroscopic image receptor. The term includes a device intended to hold a cassette in front of the input end of an image intensifier for the purpose of making a radiograph.

Stray radiation—The sum of leakage and scattered radiation.

Technique factors—The following conditions of operation:

- (i) For capacitor energy storage equipment, peak tube potential in kV and quantity of charge in mAs.
- (ii) For field emission equipment rated for pulsed operation, peak tube potential in kV, number of X-ray pulses and either tube current or product of tube current and time.

(iii) For other equipment, peak tube potential in kV and either tube current in mA and exposure time in seconds or the product of tube current and exposure time in mAs.

Therapeutic X-ray system—A system design for irradiation of a part of the human body for the purpose of treatment or alleviation of symptoms of disease.

Timer—An electronic device which is capable of measuring an X-ray exposure.

Tube—An X-ray tube, unless otherwise specified.

Tube housing assembly—The tube housing with the X-ray tube installed. The term includes high-voltage or filament transformers, or both, and other appropriate elements when contained within the tube housing.

Useful beam—The radiation which passes through the tube housing port and the aperture of the beam-limiting device when the exposure switch or timer is activated.

Variable-aperture beam-limiting device—A beam-limiting device which has capacity for stepless adjustment of the X-ray field size.

Visible area—The portion of the input surface of the image receptor over which incident X-ray photons are producing a visible image.

Wedge filter—An added filter effecting continuous progressive attenuation on all or part of the useful beam.

X-ray control—A device which controls input power to the X-ray high-voltage generator or the X-ray tube, or both. The term includes equipment such as timers, phototimers, automatic brightness stabilizers and similar devices, which control the technique factors of an X-ray exposure.

X-ray equipment—An X-ray system, subsystem or component thereof. Types of X-ray equipment are as follows:

(i) *Mobile X-ray equipment*—X-ray equipment mounted on a permanent base with wheels or casters for moving while completely assembled.

(ii) *Portable X-ray equipment*—X-ray equipment designed to be hand-carried.

(iii) *Stationary X-ray equipment*—X-ray equipment which is installed in a fixed location or vehicle.

X-ray field—The area defined by the intersection of the useful beam with a plane parallel with the plane of the image receptor. The edge of the field is defined by the points at which the exposure rate is 25% of the maximum exposure rate in the plane.

X-ray high-voltage generator—A device which transforms electrical energy from the potential supplied by the X-ray control to the tube operating potential.

X-ray subsystem—A combination of two or more components of an X-ray system.

X-ray system—An assembly of components for the controlled production of X-rays. The term includes minimally an X-ray high-voltage generator, an X-ray control, a tube housing assembly, a beam-limiting device and the necessary

supporting structures. Additional components which function with the system are considered integral parts of the system.

X-ray tube—An electron tube which is designed to be used primarily for the production of X-rays.

Authority

The provisions of this § 221.2 amended under sections 301 and 302 of the Radiation Protection Act (35 P. S. §§ 7110.301 and 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20).

Source

The provisions of this § 221.2 adopted February 1, 1972, effective February 2, 1972, 2 Pa.B. 212; amended December 18, 1987, effective December 19, 1987, 17 Pa.B. 5235; amended October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894. Immediately preceding text appears at serial pages (123672) to (123678).

Cross References

This section cited in 25 Pa. Code § 221.35a (relating to fluoroscopic X-ray systems); and 25 Pa. Code § 221.201 (relating to definitions).

§ 221.3. Sale and installation.

No person may sell or install a radiation-producing machine that does not meet the provisions of this article.

Authority

The provisions of this § 221.3 amended under section 302 of the Radiation Protection Act (35 P. S. § 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20).

Source

The provisions of this § 221.3 adopted February 1, 1972, effective February 2, 1972, 2 Pa.B. 212; amended December 18, 1987, effective December 19, 1987, 17 Pa.B. 5235. Immediately preceding text appears at serial page (4881).

Cross References

This section cited in 25 Pa. Code § 221.35a (relating to fluoroscopic X-ray systems).

ADMINISTRATIVE CONTROLS

§ 221.11. Registrant responsibilities.

(a) The registrant is responsible for directing the operation of X-ray systems under his administrative control and shall do the following:

(1) Assure that the requirements of this article are met in the operation of the X-ray systems.

(2) Permit only auxiliary personnel who have met the applicable radiologic requirements of 49 Pa. Code, Part I, Subpart A (relating to professional and occupational affairs) to operate X-ray systems for diagnostic or therapeutic

purposes when employed by a licensed practitioner of the healing arts at the location at which the licensed practitioner practices.

(3) Permit only auxiliary personnel employed by a health care facility regulated by the Department of Health, the Department of Public Welfare or the Federal government to operate X-ray systems for diagnostic or therapeutic purposes in accordance with written job descriptions and employee qualifications.

(b) An individual who operates an X-ray system shall be instructed adequately in the safe operating procedures and be competent in the safe use of the equipment. The instructions shall include, but not be limited to, items included in Appendix A (relating to determination of competence).

(c) A chart, which specifies the techniques for examinations performed with the system, shall be provided in the vicinity of each diagnostic X-ray system's control panel. This chart shall include information pertinent to the particular examination, such as:

(1) The patient's body part and anatomical size, or body part thickness, or age (for pediatrics), versus technique factors to be utilized.

(2) The type and size of the film or film-screen combination.

(3) The type of grid, if any.

(4) The type and location of placement of patient shielding-for example, gonad, and the like.

(5) For mammography, indication of kVp/target/filter combination.

(6) Source to image receptor distance to be used, except for dental intra-oral radiography.

(d) Written safety procedures and rules shall be available at a facility including restrictions of the operating technique required for the safe operation of the particular X-ray system. The operator shall be able to demonstrate familiarity with the rules.

(e) Except for patients who cannot be moved out of the room, only the staff and ancillary personnel or other persons required for the medical procedure or training shall be in the room during the radiographic exposure. The following apply for individuals other than the patient being examined:

(1) Individuals shall be positioned so that no part of the body will be struck by the useful beam unless protected by at least 0.5 millimeter lead equivalent material. The lead equivalent of the material is to be determined at 60 kV.

(2) All persons required for the medical procedure shall be protected from the scatter radiation by protective aprons or whole protective barriers of at least 0.25 millimeter lead equivalent or shall be so positioned that the persons are not in the direct line of the useful beam and the nearest portion of the body is at least 2 meters from both the tube head and the nearest edge of the image receptor.

(3) A patient who cannot be removed from the room shall be protected from the scatter radiation by protective barriers of at least 0.25 millimeter lead equivalent material unless the shield would compromise the health of the indi-

vidual or shall be so positioned that the patient is not in the direct line of the useful beam and the nearest portion of the body is at least 2 meters from both the tube head and the nearest edge of the image receptor.

(4) No individual, other than the patient being examined, may be in the useful beam, unless required to conduct the procedure.

(f) During diagnostic procedures in which the gonads are in the useful beam, gonad shielding of at least 0.5 millimeter lead equivalent shall be used for patients except for cases in which this would interfere with the diagnostic procedure.

(g) An individual may not be exposed to the useful beam except for healing arts purposes or under § 221.15 (relating to use of X-rays in research on humans). An exposure shall be authorized by a licensed practitioner of the healing arts. This provision specifically prohibits deliberate exposure for the following purposes:

(1) Exposure of an individual for training, demonstration or other nonhealing arts purposes.

(2) Exposure of an individual for the purpose of healing arts screening except as authorized by the Department. When requesting authorization, the registrant shall submit the information as outlined in § 221.13 (relating to information to be submitted by persons proposing to conduct healing arts screening).

(h) If a patient or image receptor requires auxiliary support during a radiation exposure the following apply:

(1) Mechanical holding devices shall be used when the technique permits.

(2) The human holder shall be protected as required by subsection (e).

(3) An individual may not be used routinely to hold image receptors or patients.

(i) Procedures and auxiliary equipment designed to minimize patient and personnel exposure commensurate with the needed diagnostic information shall be utilized.

(j) The screen and film system used shall be spectrally compatible. Defective screens may not be used for diagnostic radiological imaging.

(k) With the exception of intraoral dental radiography, film may not be used without intensifying screens for routine diagnostic radiological imaging.

(l) The registrant shall have a quality assurance program. This quality assurance program shall be in accordance with guidelines established by the Department or by another appropriate organization recognized by the Department. The Department's guidelines and a list of recognized organizations will be maintained and made available on the Department's website and on request.

(m) Neither the X-ray tube housing nor the collimating device may be hand-held during the exposure.

Authority

The provisions of this § 221.11 amended under sections 301 and 302 of the Radiation Protection Act (35 P. S. §§ 7110.301 and 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20).

Source

The provisions of this § 221.11 adopted February 1, 1972, effective February 2, 1972, 2 Pa.B. 212; amended December 18, 1987, effective December 19, 1987, 17 Pa.B. 5235; amended October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894; amended November 16, 2001, effective November 17, 2001, 31 Pa.B. 6282. Immediately preceding text appears at serial pages (249282) to (249285).

Cross References

This section cited in 25 Pa. Code § 221.35a (relating to fluoroscopic X-ray systems); and 25 Pa. Code § 221.42a (relating to control of scattered radiation).

§ 221.12. Records, maintenance and associated information.

The registrant shall maintain records of surveys, calibrations, maintenance and modifications performed on the X-ray systems including the names of persons who performed the services. The registrant shall keep these records for inspection by the Department for 5 years.

Authority

The provisions of this § 221.12 amended under sections 301 and 302 of the Radiation Protection Act (35 P. S. §§ 7110.301 and 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20).

Source

The provisions of this § 221.12 adopted February 1, 1972, effective February 2, 1972, 2 Pa.B. 212; amended December 18, 1987, effective December 19, 1987, 17 Pa.B. 5235; amended October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894. Immediately preceding text appears at serial page (123681).

Cross References

This section cited in 25 Pa. Code § 221.35a (relating to fluoroscopic X-ray systems).

§ 221.13. Information to be submitted by persons proposing to conduct healing arts screening.

A person requesting that the Department approve a healing arts screening program shall submit in writing the following information and evaluation. If information submitted to the Department becomes invalid or outdated, the registrant shall immediately notify the Department.

- (1) The name and address of the applicant and, if applicable, the names and addresses of agents within this Commonwealth.
- (2) The diseases or conditions for which the X-ray examinations are to be used.

(3) The description in detail of the X-ray examinations proposed in the screening program.

(4) A description of the population to be examined in the screening program—age, sex, physical condition and other appropriate information.

(5) An evaluation of known alternate methods not involving ionizing radiation which could achieve the goals of the screening program and why these methods are not used in preference to the X-ray examinations.

(6) An evaluation by a qualified expert of the X-ray systems to be used in the screening program. The evaluation shall show that the systems satisfy the requirements of this article. The evaluation shall include a measurement of patient exposures from the X-ray examinations to be performed.

(7) A description of the diagnostic film quality control program.

(8) A copy of the technique chart for the X-ray examination procedures to be used.

(9) The qualifications of an individual who will be operating the X-ray systems.

(10) The qualifications of the physician who will be supervising the operators of the X-ray systems. The extent of supervision and the method of work performance evaluation shall be specified.

(11) The name and address of the individual who will interpret the radiographs.

(12) A description of the procedures to be used in advising the individuals screened and their private practitioners of the healing arts of the results of the screening procedure and further medical needs indicated.

(13) A description of the procedures for the retention or disposition of the radiographs and other records pertaining to the X-ray examination.

(14) Mammography facilities shall comply with 21 CFR Part 900 (relating to mammography).

Authority

The provisions of this § 221.13 amended under sections 301 and 302 of the Radiation Protection Act (35 P. S. §§ 7110.301 and 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20).

Source

The provisions of this § 221.13 adopted February 1, 1972, effective February 2, 1972, 2 Pa.B. 212; amended December 18, 1987, effective December 19, 1987, 17 Pa.B. 5235; amended October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894; amended November 16, 2001, effective November 17, 2001, 31 Pa.B. 6282. Immediately preceding text appears at serial pages (249285) to (249286).

Cross References

This section cited in 25 Pa. Code § 221.11 (relating to registrant responsibilities); and 25 Pa. Code § 221.35a (relating to fluoroscopic X-ray systems).

§ 221.14. [Reserved].**Source**

The provisions of this § 221.14 adopted February 1, 1972, effective February 2, 1972, 2 Pa.B. 212; reserved December 18, 1987, effective December 19, 1987, 17 Pa.B. 5235. Immediately preceding text appears at serial page (4882).

§ 221.15. Use of X-rays in research on humans.

(a) Registrants conducting research using X-rays involving human subjects are exempted from the requirements of this section if the research is conducted, funded, regulated or supported by a Federal agency which has implemented the Federal policy for the protection of human subjects or if the research is carried out in an institution which conducts other Federally funded or supported human research and follows all Federal requirements for protocol review and research subject protection.

(b) If not exempted under subsection (a), a person shall submit, in writing, the following information and evaluation to the Department and receive approval by the Department before conducting the research. If the information submitted to the Department becomes invalid or outdated, the person shall immediately, in writing, notify the Department.

(1) The name and address of the applicant and, if applicable, the names and addresses of agents within this Commonwealth.

(2) A description of the population to be examined in the research program, age, sex, physical condition and other appropriate information.

(3) An evaluation of known alternate methods not involving ionizing radiation which could achieve the goals of the research program and why these methods are not used in preference to the X-ray examinations.

(4) An evaluation by a qualified expert of the X-ray system to be used in the research program. This evaluation shall show that the system satisfies the requirements of this article. The evaluation shall include a projected measurement of individual and cumulative patient exposures from the X-ray examinations to be performed.

(5) A description of the diagnostic X-ray quality control program.

(6) A copy of the chart which specifies the information for the X-ray examination procedures to be used.

(7) The qualifications of all individuals who will be operating the X-ray system.

(8) The qualifications of the physician who will be supervising the operators of the X-ray systems. The extent of supervision and the method of work performance evaluation shall be specified.

(9) The name and address of the individual who will interpret the radiographs.

(10) A copy of the research protocol authorized by a committee consisting of at least three persons. One of the committee members shall be knowledgeable in radiation effects on humans.

(c) Proposed subjects or their legal representative shall sign a statement acknowledging that they have been informed of their anticipated radiation exposure and possible consequences arising from this exposure.

Authority

The provisions of this § 221.15 issued under sections 301 and 302 of the Radiation Protection Act (35 P. S. §§ 7110.301 and 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20).

Source

The provisions of this § 221.15 adopted October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894.

Cross References

This section cited in 25 Pa. Code § 221.11 (relating to registrant responsibilities); and 25 Pa. Code § 221.35a (relating to fluoroscopic X-ray systems).

DIAGNOSTIC INSTALLATIONS GENERAL REQUIREMENTS

§ 221.21. Diagnostic equipment requirements.

Diagnostic systems incorporating one or more certified components shall comply with 21 CFR 1020.30—1020.33.

Authority

The provisions of this § 221.21 amended under sections 301 and 302 of the Radiation Protection Act (35 P. S. §§ 7110.301 and 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20).

Source

The provisions of this § 221.21 adopted February 1, 1972, effective February 2, 1972, 2 Pa.B. 212; amended December 18, 1987, effective December 19, 1987, 17 Pa.B. 5235; amended October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894. Immediately preceding text appears at serial page (123683).

Cross References

This section cited in 25 Pa. Code § 221.35a (relating to fluoroscopic X-ray systems).

§ 221.22. Battery charge indicator.

On battery-powered X-ray generators, the control panel shall have means to indicate visually whether the battery is adequately charged for proper operation.

Authority

The provisions of this § 221.22 amended under section 302 of the Radiation Protection Act (35 P. S. § 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20).

Source

The provisions of this § 221.22 adopted February 1, 1972, effective February 2, 1972, 2 Pa.B. 212; amended December 18, 1987, effective December 19, 1987, 17 Pa.B. 5235. Immediately preceding text appears at serial page (4882).

Cross References

This section cited in 25 Pa. Code § 221.35a (relating to fluoroscopic X-ray systems).

§ 221.23. Leakage radiation from the diagnostic source assembly.

The leakage radiation from the diagnostic source assembly measured at a distance of 1 meter in any direction from the source may not exceed 100 milliroentgens (25.8 $\mu\text{C/kg}$) in 1 hour when the X-ray tube is operated at its leakage technique factors. Compliance shall be determined by measurements averaged over an area of 100 square centimeters with no linear dimension greater than 20 centimeters.

Authority

The provisions of this § 221.23 amended under section 302 of the Radiation Protection Act (35 P. S. § 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20).

Source

The provisions of this § 221.23 adopted February 1, 1972, effective February 2, 1972, 2 Pa.B. 212; amended December 18, 1987, effective December 19, 1987, 17 Pa.B. 5235. Immediately preceding text appears at serial page (4882).

Cross References

This section cited in 25 Pa. Code § 221.35a (relating to fluoroscopic X-ray systems).

§ 221.24. Radiation from components other than the diagnostic source assembly.

The radiation emitted by a component other than the diagnostic source assembly may not exceed 2 milliroentgens (.516 $\mu\text{C/kg}$) in 1 hour at 5 centimeters from an accessible surface of the component when it is operated in an assembled X-ray system under conditions for which it was designed. Compliance shall be determined by measurements averaged over an area of 100 square centimeters with no linear dimension greater than 20 centimeters.

Authority

The provisions of this § 221.24 amended under section 302 of the Radiation Protection Act (35 P. S. § 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20).

Source

The provisions of this § 221.24 adopted February 1, 1972, effective February 2, 1972, 2 Pa.B. 212; amended December 18, 1987, effective December 19, 1987, 17 Pa.B. 5235. Immediately preceding text appears at serial pages (4882) to (4883).

Cross References

This section cited in 25 Pa. Code § 221.35a (relating to fluoroscopic X-ray systems).

§ 221.25. Beam quality.

(a) Diagnostic X-ray systems shall have filtration that satisfies the requirements of Table I. The requirements of this section shall be considered to have been met if it can be demonstrated that the half value layer of the primary beam is not less than that shown in Table II.

TABLE I

Filtration Required vs. Operating Voltage

<i>Operating Voltage (kVp)</i>	<i>Total Filtration (inherent plus added) (millimeters aluminum equivalent)</i>
Below 50.....	.5 millimeters
50—70.....	1.5 millimeters
Above 70.....	2.5 millimeters

TABLE II

<i>Design operating range (Kilovolts peak)</i>	<i>Measured potential (Kilovolts peak)</i>	<i>Half-value layer (millimeters of aluminum)</i>
Below 50.....	30	.3
	40	.4
	49	.5
50 to 70.....	50	1.2
	60	1.3
	70	1.5
Above 70.....	71	2.1
	80	2.3
	90	2.5
	100	2.7
	110	3
	120	3.2
	130	3.5
	140	3.8
	150	4.1

(b) Beryllium window tubes shall have a minimum of .5 millimeter aluminum equivalent filtration permanently installed in the useful beam.

(c) For capacitor energy storage equipment, compliance with this section shall be determined with the maximum quantity of charge per exposure.

(d) The required minimal aluminum equivalent filtration shall include the filtration contributed by materials which are always present between the source and the patient.

(e) For X-ray systems having variable filtration in the useful beam, a means shall be provided to prohibit exposure unless the filtration requirements of subsection (a) are met for the kVp selected.

Authority

The provisions of this § 221.25 issued under section 302 of the Radiation Protection Act (35 P. S. § 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20).

Source

The provisions of this § 221.25 adopted December 18, 1987, effective December 19, 1987, 17 Pa.B. 5235.

Cross References

This section cited in 25 Pa. Code § 221.35a (relating to fluoroscopic X-ray systems).

§ 221.26. Multiple tubes.

If two or more radiographic tubes are controlled by one exposure switch, the tube or tubes which have been selected shall be clearly indicated prior to initiation of the exposure. This indication shall be both on the X-ray control panel and at or near the tube housing assembly which has been selected.

Authority

The provisions of this § 221.26 issued under section 302 of the Radiation Protection Act (35 P. S. § 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20).

Source

The provisions of this § 221.26 adopted December 18, 1987, effective December 19, 1987, 17 Pa.B. 5235.

Cross References

This section cited in 25 Pa. Code § 221.35a (relating to fluoroscopic X-ray systems).

§ 221.27. Mechanical support of tube head.

The tube housing assembly supports shall be adjusted so that the tube housing assembly will remain stable during an exposure unless tube housing movement is a designed function of the X-ray system.

Authority

The provisions of this § 221.27 issued under section 302 of the Radiation Protection Act (35 P. S. § 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20).

Source

The provisions of this § 221.27 adopted December 18, 1987, effective December 19, 1987, 17 Pa.B. 5235.

Cross References

This section cited in 25 Pa. Code § 221.35a (relating to fluoroscopic X-ray systems).

§ 221.28. Technique indicators.

(a) The technique factors for radiographic systems shall be indicated before exposure except for units utilizing automatic exposure controls, in which case the maximum mAs shall be indicated.

(b) The requirement of subsection (a) may be met by permanent markings on equipment having fixed technique factors. Indication of technique factors shall be visible from the operator's position except in the case of spot films made by a fluoroscopist.

Authority

The provisions of this § 221.28 issued under sections 301 and 302 of the Radiation Protection Act (35 P. S. §§ 7110.301 and 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20).

Source

The provisions of this § 221.27 adopted December 18, 1987, effective December 19, 1987, 17 Pa.B. 5235; amended October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894. Immediately preceding text appears at serial page (123686).

Cross References

This section cited in 25 Pa. Code § 221.35a (relating to fluoroscopic X-ray systems).

§ 221.29. Kilovoltage (kV) accuracy.

The kV output may not vary from the set-indicated value by more than 10% over the range of technique factors normally used. Discrepancies of more than 10% between set-indicated and measured kV values shall be investigated by a qualified expert or service engineer and appropriate action taken.

Authority

The provisions of this § 221.29 issued under sections 301 and 302 of the Radiation Protection Act (35 P. S. §§ 7110.301 and 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20).

221-20

Source

The provisions of this § 221.29 adopted October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894; amended November 16, 2001, effective November 17, 2001, 31 Pa.B. 6282. Immediately preceding text appears at serial pages (249292) and (254507).

Cross References

This section cited in 25 Pa. Code § 221.35a (relating to fluoroscopic X-ray systems).

§ 221.30. Exposure reproducibility.

The coefficient of variation of exposure reproducibility may not exceed 0.10 when technique factors are held constant. This requirement shall be deemed to have been met when four exposures are made. This requirement applies when either manual techniques or automatic exposure control is used.

Authority

The provisions of this § 221.30 issued under sections 301 and 302 of the Radiation Protection Act (35 P. S. §§ 7110.301 and 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20).

Source

The provisions of this § 221.30 adopted October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894.

Cross References

This section cited in 25 Pa. Code § 221.35a (relating to fluoroscopic X-ray systems).

§ 221.31. [Reserved].**Source**

The provisions of this § 221.31 adopted February 1, 1972, effective February 2, 1972, 2 Pa.B. 212; amended December 18, 1987, effective December 19, 1987, 17 Pa.B. 5235; reserved October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894. Immediately preceding text appears at serial pages (123686) to (123688).

§ 221.31a. Locks.

Position locking, holding and centering devices on X-ray systems shall function as intended.

Authority

The provisions of this § 221.31a issued under sections 301 and 302 of the Radiation Protection Act (35 P. S. §§ 7110.301 and 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20).

Source

The provisions of this § 221.31a adopted October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894.

Cross References

This section cited in 25 Pa. Code § 221.35a (relating to fluoroscopic X-ray systems).

§ 221.32. [Reserved].**Source**

The provisions of this § 221.32 adopted February 1, 1972, effective February 2, 1972, 2 Pa.B. 212; amended December 18, 1987, effective December 19, 1987, 17 Pa.B. 5235; reserved October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894. Immediately preceding text appears at serial page (123688).

§ 221.32a. Radiographic beam limitation.

- (a) The useful beam shall be limited to the area of clinical interest.
- (b) The beam limiting device shall do one of the following:
 - (1) Indicate numerically the field size in the plane of the image receptor to which it is adjusted to within 2% of the SID.
 - (2) Provide for visually defining the perimeter of the X-ray field except for systems designed for one image receptor size. The total misalignment of the edges of the visually defined field with the respective edges of the X-ray field may not exceed 2% of the distance from the source to the center of the visually defined field when the surface upon which it appears is perpendicular to the axis of the X-ray beam.
- (c) A means shall be provided for stepless (continuous) adjustment of the size of the X-ray field except for systems which use removable fixed operation beam limiting devices.
- (d) A means shall be provided to:
 - (1) Indicate when the axis of the X-ray beam is perpendicular to the plane of the image receptor if the angle between the axis of the X-ray beam and the plane of the image receptor is variable. This paragraph does not apply to portable, mobile or intraoral dental units.
 - (2) Align the center of the X-ray field with respect to the center of the image receptor to within 2% of the SID.
 - (3) Indicate the SID to within 2%.
- (e) Intraoral dental X-ray systems designed for use with an intraoral image receptor shall be provided with means to limit SSD to not less than either of the following:
 - (1) Eighteen centimeters if operable above 50 kVp.
 - (2) Ten centimeters if not operable above 50 kVp.
- (f) Indication of field size dimensions and SIDs shall be specified so that aperture adjustments result in X-ray field dimensions in the plane of the image receptor which correspond to those indicated by the beam-limiting device to within 2% of the SID when the beam axis is indicated to be perpendicular to the plane of the image receptor.

(g) Intraoral dental systems designed for use with an intraoral image receptor shall be provided with a means to limit the X-ray beam so that:

(1) Eighteen centimeters or more, the X-ray field at the minimum SSD shall be containable in a circle having a diameter of no more than 7 centimeters.

(2) Less than 18 centimeters, the X-ray field at the minimum SSD shall be containable in a circle having a diameter of no more than 6 centimeters.

(h) When positive beam limitation is used, the following conditions shall be met:

(1) The radiation beam may not be larger than the linear dimensions of the image receptor being used.

(2) The positive beam limitation device shall allow the operator to further reduce the size of the radiation field.

(i) Mobile or portable radiographic systems, other than intraoral dental X-ray systems, shall be provided with a means to limit the source-to-skin distance to at least 30 centimeters.

(j) Radiographic equipment designed for one or more image receptor sizes at a fixed SID shall be provided with a means to accomplish one of the following:

(1) Limit the field at the plane of the image receptor to dimensions no greater than those of the image receptor and align the center of the X-ray field with the center of the image receptor to within 2% of the SID.

(2) The X-ray field shall be sized and aligned so that at the plane of the image receptor, it does not extend beyond the edge of the image receptor by more than 2% of the SID.

Authority

The provisions of this § 221.32a issued under sections 301 and 302 of the Radiation Protection Act (35 P. S. §§ 7110.301 and 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20).

Source

The provisions of this § 221.32a adopted October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894; amended November 16, 2001, effective November 17, 2001, 31 Pa.B. 6282. Immediately preceding text appears at serial pages (254508) and (249295).

Cross References

This section cited in 25 Pa. Code § 221.35a (relating to fluoroscopic X-ray systems).

§ 221.33. [Reserved].

Source

The provisions of this § 221.33 adopted February 1, 1972, effective February 2, 1972, 2 Pa.B. 212; amended December 18, 1987, effective December 19, 1987, 17 Pa.B. 5235; reserved October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894. Immediately preceding text appears at serial pages (123688) to (123689).

§ 221.33a. Radiation from capacitor energy storage equipment in standby status.

Radiation emitted from an X-ray tube when the exposure switch or timer is not activated may not exceed a rate of 2 milliroentgens (0.516 $\mu\text{C/kg}$) per hour at 5 centimeters from an accessible surface of a fully charged diagnostic source assembly, with the beam-limiting device fully open.

Authority

The provisions of this § 221.33a issued under sections 301 and 302 of the Radiation Protection Act (35 P. S. §§ 7110.301 and 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20).

Source

The provisions of this § 221.33a adopted October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894; amended November 16, 2001, effective November 17, 2001, 31 Pa.B. 6282. Immediately preceding text appears at serial pages (249295) to (249296).

Cross References

This section cited in 25 Pa. Code § 221.35a (relating to fluoroscopic X-ray systems).

§ 221.34. [Reserved].

Source

The provisions of this § 221.34 adopted February 1, 1972, effective February 2, 1972, 2 Pa.B. 212; amended December 18, 1987, effective December 19, 1987, 17 Pa.B. 5235; reserved October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894. Immediately preceding text appears at serial page (123690).

§ 221.34a. Radiation exposure control.

(a) *Radiation exposure control.* A means shall be provided to initiate the radiation exposure by a deliberate action on the part of the operator, such as the depression of a switch. Radiation exposure may not be initiated without such an action.

(b) *Visual indication and audible signal.* A means shall be provided for visual indication observable from the operator's protected position whenever X-rays are produced. In addition, a signal audible to the operator shall indicate that the exposure has terminated.

(c) *Termination of exposure.* A means shall be provided to terminate the exposure at a preset time interval, preset product of current and time, a preset number of pulses or a preset radiation exposure to the image receptor. Except for dental panoramic systems, termination of an exposure shall cause automatic resetting of the time to its initial setting or to "zero."

(d) *Manual exposure control.* An X-ray control shall be incorporated into each X-ray system which allows the operator to terminate an exposure at any time except for one or more of the following:

- (1) Exposure of 1/2 second or less.

(2) During serial radiography in which case a means shall be provided to permit completion of any single exposure of the series in process.

(e) *Automatic exposure control.*

(1) Indication shall be made on the control panel when this mode of operation is selected.

(i) A means shall be provided to terminate irradiation at an appropriate exposure for the projection if the automatic exposure control fails to terminate irradiation.

(ii) A visible signal shall indicate when an exposure has been terminated at the limits required by subparagraph (i), and manual resetting shall be required before further automatically timed exposures can be made.

(2) For X-ray systems operating in automatic exposure control mode, and which lack engineered safeguards that prevent exposure in the event of either a malfunction or a mispositioned X-ray beam with respect to film cassette sensors, the back-up or default mAs shall be set by the operator to an appropriate maximum value for the projection.

(3) X-ray systems utilizing automatic exposure control, in which the back-up mAs values are preset and cannot be selected by the operator, shall prominently indicate the preset mAs value on the console, along with an appropriate warning notice to the operator.

(f) *Exposure control location.*

(1) Stationary X-ray systems shall have X-ray controls permanently mounted in a protected area and situated so that the operator is required to remain in that protected area during the entire exposure.

(2) For mobile and portable X-ray systems the exposure switch shall be arranged so that the operator can stand at least 2 meters from the patient and from the tube head and away from the direction of the useful X-ray beam.

Authority

The provisions of this § 221.34a issued under sections 301 and 302 of the Radiation Protection Act (35 P. S. §§ 7110.301 and 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20).

Source

The provisions of this § 221.34a adopted October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894.

Cross References

This section cited in 25 Pa. Code § 221.35a (relating to fluoroscopic X-ray systems).

§ 221.35. [Reserved].

Source

The provisions of this § 221.35 adopted February 1, 1972, effective February 2, 1972, 2 Pa.B. 212; amended December 18, 1987, effective December 19, 1987, 17 Pa.B. 5235; reserved October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894. Immediately preceding text appears at serial pages (123690) to (123691).

§ 221.35a. Fluoroscopic X-ray systems.

Fluoroscopic X-ray systems shall use an image intensifier and in addition to the requirements of §§ 221.1—221.34a, shall meet the requirements of §§ 221.36a—221.38a (relating to limitation of useful beam of fluoroscopic equipment; activation of fluoroscopic tube; and entrance exposure rate).

Authority

The provisions of this § 221.35a issued under sections 301 and 302 of the Radiation Protection Act (31 P. S. §§ 7110.301 and 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20).

Source

The provisions of this § 221.35a adopted October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894.

Cross References

This section cited in 25 Pa. Code § 221.43a (relating to mobile fluoroscopes); and 25 Pa. Code § 221.61 (relating to radiation therapy simulation systems).

§ 221.36. [Reserved].**Source**

The provisions of this § 221.36 adopted December 18, 1987, effective December 19, 1987, 17 Pa.B. 5235; reserved October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894. Immediately preceding text appears at serial page (123691).

§ 221.36a. Limitation of useful beam of fluoroscopic equipment.

(a) The fluoroscopic imaging assembly shall be provided with a primary protective barrier which intercepts the entire cross section of the useful beam at any source-to-image receptor distance.

(b) The X-ray tube used for fluoroscopy may not produce X-rays unless a barrier is in position to intercept the useful beam and the imaging device is in place and operable.

(c) A means shall be provided for stepless (continuous) adjustment of the field size.

(d) The minimum field size at the greatest source to image receptor distance shall be containable in a square of 5 centimeters by 5 centimeters unless otherwise provided in 21 CFR 1020.32(b) (relating to fluoroscopic equipment).

(e) Equipment may not be operated at a source to skin distance less than 30 centimeters or as required under 21 CFR 1020.32(g) (relating to source-skin distance fluoroscopic equipment).

(f) The width of the X-ray field in the plane of the image receptor may not exceed that of the visible area of the image receptor by more than 3% of the source to image receptor distance. The sum of the excess length and the excess width may not be greater than 4% of the source to image receptor distance.

(g) For rectangular X-ray fields used with a circular image receptor, the error in alignment shall be determined along the length and width dimensions of the X-ray field which passes through the center of the visible area of the image receptor.

(h) Compliance with subsections (a)—(g) shall be determined with the beam axis perpendicular to the plane of the image receptor.

(i) Spot-film devices shall meet the following additional requirements:

(1) A means shall be provided between the source and the patient for adjustment of the X-ray field size to the size of the portion of film which has been selected on the spot-film selector.

(2) The adjustments shall be automatically accomplished except when the X-ray field size in the plane of the film is smaller than that of the film.

(3) The total misalignment of the edges of the X-ray field with the respective edges of the selected portion of the image receptor along the length or width dimensions of the X-ray field in the plane of the image receptor may not exceed 3% of the source-to image receptor when adjusted for full coverage of the selected portion of the image receptor.

(4) The sum, without regard to sign, of the misalignment along any two orthogonal dimensions, may not exceed 4% of the source to image receptor distance.

(5) The center of the X-ray field in the plane of the film shall be aligned with the center of the film within 2% of the source to image receptor distance.

Authority

The provisions of this § 221.36a issued under sections 301 and 302 of the Radiation Protection Act (35 P. S. §§ 7110.301 and 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20).

Source

The provisions of this § 221.36a adopted October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894; amended November 16, 2001, effective November 17, 2001, 31 Pa.B. 6282. Immediately preceding text appears at serial pages (249298) and (254509).

Cross References

This section cited in 25 Pa. Code § 221.35a (relating to fluoroscopic X-ray systems); 25 Pa. Code § 221.43a (relating to mobile fluoroscopes); and 25 Pa. Code § 221.61 (relating to radiation therapy simulation systems).

§ 221.37. [Reserved].

Source

The provisions of this § 221.37 adopted December 18, 1987, effective December 19, 1987, 17 Pa.B. 5235; reserved October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894. Immediately preceding text appears at serial pages (123691) to (123692).

§ 221.37a. Activation of fluoroscopic tube.

X-ray production in the fluoroscopic mode shall be controlled by a device which requires continuous pressure by the fluoroscopist for the entire time of the exposure (dead-man switch). When recording serial fluoroscopic images, the fluoroscopist shall be able to terminate X-ray exposures at any time, but means may be provided to permit completion of any single exposure of the series in process.

Authority

The provisions of this § 221.37a issued under sections 301 and 302 of the Radiation Protection Act (35 P. S. §§ 7110.301 and 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20).

Source

The provisions of this § 221.37a adopted October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894.

Cross References

This section cited in 25 Pa. Code § 221.35a (relating to fluoroscopic X-ray systems); 25 Pa. Code § 221.43a (relating to mobile fluoroscopes); and 25 Pa. Code § 221.61 (relating to radiation therapy simulation systems).

§ 221.38. [Reserved].**Source**

The provisions of this § 221.38 adopted December 18, 1987, effective December 19, 1987, 17 Pa.B. 5235; reserved October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894. Immediately preceding text appears at serial page (123692).

§ 221.38a. Entrance exposure rate.

(a) *Fluoroscopic systems without high level control.* The exposure rate may not exceed 10 roentgens (2.58 mC/kg) per minute except during recording of fluoroscopic images.

(b) *Fluoroscopic systems with high level control.*

(1) When the high level control is activated, the maximum exposure rate shall be 20 roentgens (5.16 mC/kg) per minute.

(2) When the high level control is not activated, the maximum exposure rate shall be 10 roentgens (2.58 mC/kg) per minute.

(3) Special means of activation of high level controls are required. The high level control shall only be operable when continuous manual activation is provided by the operator.

(4) There shall be an indication to the fluoroscopist that the high level control is being used.

(c) *Frequency of output measurements.* Output measurements required by this section shall be made annually and after maintenance that could affect the output of the machine.

(d) *Compliance requirements.* Compliance with subsections (a)—(c) shall be determined as follows:

(1) If the source is below the table, the exposure rate shall be expressed for the center of the useful beam 1 centimeter above the tabletop or cradle with the image intensifier 30 centimeters above the tabletop or cradle.

(2) If the source is above the table, the exposure rate shall be measured at 30 centimeters above the tabletop with the end of the beam-limiting device or spacer positioned as closely as possible to the point of measurement.

(3) In a c-arm type of fluoroscope, the exposure rate shall be measured at 30 centimeters from the input surface of the fluoroscopic imaging assembly with the source at its closest possible position of operation.

(4) The tube potential and current shall be set to give the maximum exposure possible from the X-ray system. For systems with automatic exposure control, at least 3 millimeters of lead shall be placed between the measuring device and image receptor.

(5) The measurement shall be made at the center of the useful beam.

Authority

The provisions of this § 221.38a issued under sections 301 and 302 of the Radiation Protection Act (35 P. S. §§ 7110.301 and 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20).

Source

The provisions of this § 221.38a adopted October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894.

Cross References

This section cited in 25 Pa. Code § 221.35a (relating to fluoroscopic X-ray systems); 25 Pa. Code § 221.43a (relating to mobile fluoroscopes); and 25 Pa. Code § 221.61 (relating to radiation therapy simulation systems).

§ 221.39. [Reserved].

Source

The provisions of this § 221.39 adopted December 18, 1987, effective December 19, 1987, 17 Pa.B. 5235; reserved October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894. Immediately preceding text appears at serial pages (123692) to (123693).

§ 221.39a. Barrier transmitted radiation rate limits.

The protective barrier may not transmit more than 2 milliroentgens (.516 $\mu\text{mC/kg}$) per hour at 10 centimeters from an accessible surface of the fluoroscopic imaging assembly for each roentgen per minute of entrance exposure rate.

Authority

The provisions of this § 221.39a issued under sections 301 and 302 of the Radiation Protection Act (35 P. S. §§ 7110.301 and 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20).

Source

The provisions of this § 221.39a adopted October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894.

Cross References

This section cited in 25 Pa. Code § 221.43a (relating to mobile fluoroscopes); and 25 Pa. Code § 221.61 (relating to radiation therapy simulation systems).

§ 221.40. [Reserved].**Source**

The provisions of this § 221.40 adopted December 18, 1987, effective December 19, 1987, 17 Pa.B. 5235; reserved October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894. Immediately preceding text appears at serial page (123693).

§ 221.40a. Indication of tube voltage and current.

During fluoroscopy and cinefluorography, the voltage and the current shall be continuously indicated.

Authority

The provisions of this § 221.40a issued under sections 301 and 302 of the Radiation Protection Act (35 P. S. §§ 7110.301 and 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20).

Source

The provisions of this § 221.40a adopted October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894.

Cross References

This section cited in 25 Pa. Code § 221.43a (relating to mobile fluoroscopes); and 25 Pa. Code § 221.61 (relating to radiation therapy simulation systems).

§ 221.41. [Reserved].**Source**

The provisions of this § 221.41 adopted February 1, 1972, effective February 2, 1972, 2 Pa.B. 212; amended December 18, 1987, effective December 19, 1987, 17 Pa.B. 5235; reserved October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894. Immediately preceding text appears at serial page (123693).

§ 221.41a. Fluoroscopic timer.

A cumulative timing device activated by the fluoroscope switch shall be provided. It shall indicate the passage of a predetermined period of irradiation either by an audible signal or by temporary or permanent interruption of the irradiation when the increment of exposure time exceeds a predetermined limit not exceeding 5 minutes.

Authority

The provisions of this § 221.41a issued under sections 301 and 302 of the Radiation Protection Act (35 P. S. §§ 7110.301 and 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20).

Source

The provisions of this § 221.41a adopted October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894.

Cross References

This section cited in 25 Pa. Code § 221.35a (relating to fluoroscopic X-ray systems); 25 Pa. Code § 221.43a (relating to mobile fluoroscopes); and 25 Pa. Code § 221.61 (relating to radiation therapy simulation systems).

§ 221.42. [Reserved].**Source**

The provisions of this § 221.42 adopted December 18, 1987, effective December 19, 1987, 17 Pa.B. 5235; reserved October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894. Immediately preceding text appears at serial page (123694).

§ 221.42a. Control of scattered radiation.

(a) Fluoroscopic table designs when combined with normal operating procedures shall be of a type so no unprotected part of the staff or an ancillary individual's whole body is exposed to unattenuated scattered radiation which originates from under the table. The attenuation required may be not less than .25 millimeter lead equivalent.

(b) Equipment configuration when combined with normal operating procedures shall be of a type so that no portion of the staff or an ancillary individual's whole body, except the extremities, is exposed to the unattenuated scattered radiation emanating from above the table top unless one of the following criteria is met:

(1) The individual is at least 120 centimeters from the center of the useful beam.

(2) The radiation has passed through at least .25 millimeter of lead equivalent material—for example, drapes, bucky-slot cover (film-tray cover panel), sliding or folding panel or self supporting curtains—in addition to lead equivalency provided by the protective apron referred to in § 221.11(e) (relating to registrant responsibilities).

Authority

The provisions of this § 221.42a issued under sections 301 and 302 of the Radiation Protection Act (35 P. S. §§ 7110.301 and 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20).

Source

The provisions of this § 221.42a adopted October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894.

Cross References

This section cited in 25 Pa. Code § 221.43a (relating to mobile fluoroscopes); and 25 Pa. Code § 221.61 (relating to radiation therapy simulation systems).

§ 221.43. [Reserved].**Source**

The provisions of this § 221.43 adopted December 18, 1987, effective December 19, 1987, 17 Pa.B. 5235; reserved October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894. Immediately preceding text appears at serial pages (123694) to (123695).

§ 221.43a. Mobile fluoroscopes.

In addition to the other requirements of §§ 221.35a—221.42a, mobile fluoroscopes shall provide image intensification.

Authority

The provisions of this § 221.43a issued under sections 301 and 302 of the Radiation Protection Act (35 P. S. §§ 7110.301 and 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20).

Source

The provisions of this § 221.43a adopted October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894.

Cross References

This section cited in 25 Pa. Code § 221.61 (relating to radiation therapy simulation systems).

§§ 221.44—221.49. [Reserved].**Source**

The provisions of these §§ 221.44—221.49 adopted December 18, 1987, effective December 19, 1987, 17 Pa.B. 5235; reserved October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894. Immediately preceding text appears at serial pages (123694) to (123698).

§§ 221.51—221.55. [Reserved].**Source**

The provisions of these §§ 221.51—221.55 adopted December 18, 1987, effective December 19, 1987, 17 Pa.B. 5235; reserved October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894. Immediately preceding text appears at serial pages (123698) to (123701).

§ 221.56. [Reserved].**Source**

The provisions of this § 221.56 adopted December 18, 1987, effective December 19, 1987, 17 Pa.B. 5235; reserved November 16, 2001, effective November 17, 2001, 31 Pa.B. 6282. Immediately preceding text appears at serial pages (249304) and (254511).

OTHER SYSTEMS**§ 221.61. Radiation therapy simulation systems.**

Radiation therapy simulation systems shall comply with §§ 221.35a—221.43a. Radiation therapy simulation systems are exempt from §§ 221.36a, 221.38a, 221.39a and 221.41a if the systems that do not meet the requirements in § 221.41a (relating to fluoroscopic timer) are provided with a means of indicating the cumulative time that an individual patient has been exposed to X-rays. In this case, procedures shall require that the timer be reset between examinations.

Authority

The provisions of this § 221.61 issued under sections 301 and 302 of the Radiation Protection Act (35 P. S. §§ 7110.301 and 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20).

Source

The provisions of this § 221.61 adopted December 18, 1987, effective December 19, 1987, 17 Pa.B. 5235; amended October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894; amended November 16, 2001, effective November 17, 2001, 31 Pa.B. 6282. Immediately preceding text appears at serial page (254511).

§ 221.62. [Reserved].**Source**

The provisions of this § 221.62 adopted December 18, 1987, effective December 19, 1987, 17 Pa.B. 5235; reserved October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894. Immediately preceding text appears at serial pages (123701) to (123702).

**THERAPEUTIC X-RAY SYSTEMS WITH ENERGIES
LESS THAN 1 MEV****§ 221.71. Equipment requirements.**

(a) When the tube is operated at its leakage technique factors, the leakage radiation may not exceed:

- (1) One hundred milliroentgens (25.8 μ C/kg) per hour at 5 centimeters from the surface of the tube housing assembly for contact therapy systems.
- (2) One roentgen (.258 mC/kg) per hour at 1 meter from the source for 0-150 kVp systems manufactured or installed prior to December 19, 1987.
- (3) One hundred milliroentgens (25.8 μ C/kg) per hour at 1 meter from the source for 0-150 kVp systems manufactured on or after December 19, 1987.
- (4) One roentgen (.258 mC/kg) per hour at 1 meter from the source for 151 to 500 kVp systems.
- (5) One-tenth percent of the exposure rate of the useful beam 1 meter from the source for 501 to 999 kVp systems at 1 meter from the source.

(b) Fixed diaphragms or cones used for limiting the useful beam shall provide at least the same protection as required by the tube housing assembly.

(c) Beam limiting devices may, for the portion of the useful beam blocked by these devices, transmit not more than 5% of the original X-ray beam intensity at the maximum voltage and maximum treatment filter. This requirement does not apply to auxiliary blocks or materials placed in the useful beam to shape the useful beam to the individual patient.

(d) The filter system shall be designed so that:

(1) Filters cannot be accidentally displaced from the useful beam at any possible tube orientation.

(2) The radiation at 5 centimeters from the filter insertion slot opening does not exceed 30 roentgens (7.74 mC/kg) per hour under operating conditions.

(3) A filter is marked as to its material of construction and its thickness. For wedge filters, the wedge factor and wedge angle shall appear on the wedge or wedge tray.

(4) On equipment purchased after January 1, 1971, a filter indication system shall be used on therapy machines using changeable filters. The system shall indicate from the control panel the presence or absence of a filter and shall be designed to permit easy recognition of an added filter in place.

(5) An X-ray system equipped with a beryllium or other low-filtration window shall be clearly labeled as such upon the tube housing assembly and at the control panel.

(e) The tube housing assembly shall be immobilized during stationary treatments.

(f) The tube housing assembly shall be so marked that it is possible to determine the location of the focal spot to within 5 millimeters, and the marking shall be readily accessible for use during calibration procedures.

(g) Contact therapy tube housing assemblies shall have a removable shield of at least .5 millimeter lead equivalency at 100 kVp that can be positioned over the entire useful beam exit port during periods when the beam is not in use.

(h) Systems of greater than 150 kVp manufactured after December 19, 1987, shall have a beam monitor system which shall meet the following requirements:

(1) Not allow irradiation until a preselected value of exposure has been made at the treatment control panel.

(2) Independently terminate irradiation when the preselected exposure has been reached.

(3) Be designed so that, in the event of a system malfunction or electrical power failure or other interruption, the dose administered to a patient prior to the interruption can be accurately determined.

(4) Have a control panel display which maintains the reading until intentionally reset to zero.

(5) Have a control panel display which does not have scale multiplying factors and utilizes a design so that increasing dose is displayed by increasing numbers.

(i) The following apply to times on the equipment:

(1) A timer shall be provided which has a display at the control panel. The timer shall be graduated in minutes and fractions of minutes. The timer shall have a preset time selector and an elapsed time indicator.

(2) The timer shall be a cumulative timer which activates with the radiation and retains its reading after irradiation is interrupted or terminated. After irradiation is terminated and before irradiation can be reinitiated, it shall be necessary to reset the timer to zero.

(3) The timer shall terminate irradiation when a preselected time has elapsed if a dose monitoring system present has not previously terminated irradiation.

(4) The timer shall permit accurate presetting and determination of exposure time as short as 1 second.

(5) The timer may not permit an exposure if set at zero.

(6) The timer may not activate until the shutter is opened when patient irradiation is controlled by a shutter mechanism.

(j) The control panel, in addition to the displays required in this section, shall have:

(1) An indication of power status.

(2) An indication of X-ray production.

(3) The means of indicating X-ray tube current and voltage.

(4) The means of terminating an exposure.

(k) When a control panel may energize more than one X-ray tube, the following requirements shall be met:

(1) It shall be possible to activate only one X-ray tube at one time.

(2) There shall be an indication at the control panel identifying which X-ray tube is energized.

(3) There shall be an indication at the tube housing assembly when that tube is energized.

(l) There shall be a means of determining the SSD to within 5 millimeters.

(m) Unless it is possible to bring the X-ray output to the prescribed exposure parameters within 5 seconds, the entire useful beam shall be automatically attenuated by a shutter having a lead equivalency not less than that of the tube housing assembly.

- (1) After the unit is at operating parameters, the shutter shall be controlled electrically by the operator from the control panel.
- (2) An indication of shutter position shall appear at the control panel.

Authority

The provisions of this § 221.71 issued under section 302 of the Radiation Protection Act (35 P. S. § 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20).

Source

The provisions of this § 221.71 adopted December 18, 1987, effective December 19, 1987, 17 Pa.B. 5235.

§ 221.72. Facility design requirements for systems capable of operating above 50 kVp.

- (a) Provision shall be made to permit continuous observation of and communication with the patient during irradiation.
- (b) Windows, mirror systems or closed-circuit television viewing screens used for observing the patient shall be so located that the operator can maintain direct surveillance over both the control panel and the patient.
- (c) Treatment rooms which contain an X-ray system capable of operating above 150 kVp shall meet the following additional requirements:
 - (1) Necessary shielding, except for a beam interceptor, shall be provided by fixed barriers.
 - (2) The control panel shall be outside the treatment room or in a shielded booth.
 - (3) Doors of the treatment room shall be electrically interlocked to the control panel so that X-ray production cannot occur unless entrance doors are closed.
 - (4) Interlocks shall be provided so that, when a door of the treatment room is opened, either the machine will shut off automatically or the radiation level within the room will be reduced to an average of not more than two milliroentgens (.52 $\mu\text{C/kg}$) per hour and a maximum of ten milliroentgens (2.58 $\mu\text{C/kg}$) per hour at a distance of 1 meter in any direction from the target; or interlocks shall energize a conspicuous visible or audible alarm signal so that the individual entering and the operator are made aware of the entry. After a shut-off or reduction in output, it shall be possible to restore the machine to full operation only from the control panel.
 - (5) Treatment room entrances shall be provided with warning lights, which will indicate when the useful beam is on, in a readily observable position near the outside of access doors.

Authority

The provisions of this § 221.72 issued under section 302 of the Radiation Protection Act (35 P. S. § 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20).

Source

The provisions of this § 221.72 adopted December 18, 1987, effective December 19, 1987, 17 Pa.B. 5235.

§ 221.73. Surveys.

(a) A facility shall have a survey made by, or under the direction of, a qualified expert or a radiological physicist. The survey shall also be done after a change in the facility or equipment which might cause a change in radiation levels.

(b) The qualified expert or radiological physicist shall report the survey results in writing to the individual in charge of the facility and a copy of the report shall be maintained by the registrant for inspection by the Department. The facility shall be operated in compliance with limitations indicated by the survey.

Authority

The provisions of this § 221.73 issued under section 302 of the Radiation Protection Act (35 P. S. § 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20).

Source

The provisions of this § 221.73 adopted December 18, 1987, effective December 19, 1987, 17 Pa.B. 5235.

§ 221.74. Calibration.

(a) The calibration of an X-ray system shall be performed at intervals not to exceed 1 year and after a change of replacement of components which could cause a change in the radiation output.

(b) The calibration of the radiation output of the X-ray system shall be performed by or under the direction of a radiological physicist who is physically present at the facility during the calibration.

(c) The calibration of the radiation output of an X-ray system shall be performed with a calibrated instrument. The calibration of the instrument shall be traceable to a national standard. The instrument shall have been calibrated within the preceding 2 years.

(d) Calibrations made under this section shall be made so that the dose at a reference point in soft tissue may be calculated as accurately as possible but with an uncertainty of no greater than 5%.

(e) The calibration of the X-ray system shall include, but is not limited to, the following determinations:

- (1) The exposure rates for each combination of field size, technique factors, filter and treatment distance used.
- (2) The degree of congruence between the radiation field and the field indicated by the localizing device if a device is present.
- (3) An evaluation of the uniformity of the largest radiation field used.

(f) Records of calibration performed under this section shall be maintained by the registrant for at least 5 years after completion of the calibration.

(g) A copy of the most recent X-ray system calibration shall be available at the control panel.

Authority

The provisions of this § 221.74 issued under section 302 of the Radiation Protection Act (35 P. S. § 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20).

Source

The provisions of this § 221.74 adopted December 18, 1987, effective December 19, 1987, 17 Pa.B. 5235.

Cross References

This section cited in 25 Pa. Code § 221.75 (relating to spot checks).

§ 221.75. Spot checks.

Spot checks shall be performed on X-ray systems capable of operation at greater than 150 kVp. The spot checks shall meet the following requirements:

(1) The procedures shall be in writing and shall have been developed by a radiological physicist.

(2) If a radiological physicist does not perform the spot check measurements, the results of the spot check measurements shall be reviewed by a radiological physicist within 15 days.

(3) The measurements taken during the spot checks shall demonstrate the degree of consistency of the operating characteristics which can affect the radiation output of the X-ray system.

(4) The spot-check procedure shall specify the frequency at which tests or measurements are to be performed and the acceptable tolerance for each parameter measured in the spot check when compared to the value for that parameter determined in the calibration specified in § 221.74 (relating to calibration).

(5) The procedure shall also note conditions which require that the system be recalibrated under § 221.74.

(6) Records of spot-check measurements performed under this section shall be maintained by the registrant for 5 years following the measurement.

(7) Spot check measurements shall be performed using a dosimetry system that has been calibrated under § 221.74(c). Alternatively, a dosimetry system used solely for spot check measurements may be calibrated by direct intercomparison with a system that has been calibrated under § 221.74(c). The alternative calibration method shall have been performed within the previous year and after each servicing that may have affected the system calibration.

Authority

The provisions of this § 221.75 issued under section 302 of the Radiation Protection Act (35 P. S. § 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20).

Source

The provisions of this § 221.75 adopted December 18, 1987, effective December 19, 1987, 17 Pa.B. 5235.

§ 221.76. Operating procedures.

(a) Therapeutic X-ray systems shall be secured to prevent unauthorized use whenever the system is unattended.

(b) If a patient is held in position for radiation therapy, mechanical supporting or restraining devices shall be used.

(c) The tube housing assembly may not be held by an individual during exposures.

(d) No individual other than the patient may be in the treatment room during irradiation unless protected by a shielded booth.

(e) Interlocks, on-off beam control mechanisms and safety and warning devices shall be checked and appropriately serviced at least once in a calendar year.

Authority

The provisions of this § 221.76 issued under section 302 of the Radiation Protection Act (35 P. S. § 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20).

Source

The provisions of this § 221.76 adopted December 18, 1987, effective December 19, 1987, 17 Pa.B. 5235.

§§ 221.81—221.102. [Reserved].**Source**

The provisions of these §§ 221.81—221.102 adopted December 18, 1987, effective December 19, 1987, 17 Pa.B. 5235; reserved October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894. Immediately preceding text appears at serial pages (123709) to (123726).

COMPUTED TOMOGRAPHY X-RAY SYSTEMS**§ 221.201. Definitions.**

In addition to the definitions in §§ 215.2 and 221.2 (relating to definitions), the following words and terms when used in this section and §§ 221.202—221.205, have the following meanings, unless the context clearly indicates otherwise:

CS—Contrast scale—The change in the linear attenuation coefficient per CT number relative to water; that is:

$$CS = (U_x - U_w) / ((CT)_x - (CT)_w)$$

Where:

U_x = Linear attenuation coefficient of the material of interest

U_w = Linear attenuation coefficient of water

$(CT)_x$ = CT number of the material of interest

$(CT)_w$ = CT number of water

CT number—The number used to represent the X-ray attenuation associated with each elemental area of the CT image.

CT—Computed tomography—The production of a tomogram by the acquisition and computer processing of X-ray transmission data.

CTDI—Computed tomography dose index—The integral of the dose profile along a line perpendicular to the tomographic plane divided by the product of the nominal tomographic section thickness and the number of tomograms produced in a single scan.

CT conditions of operation—The selectable parameters governing the operation of a CT X-ray system including, but not limited to, nominal tomographic section thickness, filtration and the technique factors as defined in this chapter.

Elemental area—The smallest area within a tomogram for which the X-ray attenuation properties of a body are depicted.

Gantry—The tube housing assemblies, beam-limiting devices, detectors, transformers, if applicable, and the supporting structures and frames which hold these components.

Lux—A unit illumination equivalent to 1 lumen per square centimeter or .0929 foot-candles.

MSAD—Multiple scan average dose—The calculated average dose to the tissue within each slice in a series utilizing an ion chamber. The MSAD is calculated using the following equation:

$$MSAD = (F \times K \times L \times E) / (T \times N)$$

Where

F = Factor to convert exposure in air to absorbed dose in lucite in RADS/mR

K = Calibration factor to account for the ion chamber's response and volume.

L = Effective length of ion chamber in millimeters (mm)

E = Exposure reading in milliroentgen (mR)

T = Nominal slice thickness in millimeters (mm) and

N = Number of slices per scan

Multiple tomogram system—A computed tomography X-ray system which obtains X-ray transmission data simultaneously during a single scan to produce more than one tomogram.

Noise—The standard deviation of the fluctuations in the CT number expressed as a percentage of the attenuation coefficient of water. Its estimate (S_n) is calculated using the following expression:

$$S_n = 100 \times CS \times S/U_w$$

Where:

CS = Contrast scale

U_w = Linear attenuation coefficient of water.

S = Estimated standard deviation of the CT number of picture elements in a specified area of the CT image.

Nominal tomographic section thickness—The full-width at half-maximum of the sensitivity profile taken at the center of the cross-sectional volume over which X-ray transmission data are collected.

Performance phantom—A phantom which has a capability of providing an indication of contrast scale, noise, nominal tomographic section thickness, the resolution capability of the CT system for low and high contrast objects, and measuring the mean CT number for water or other reference materials.

Picture element—See elemental area.

Pixel—See elemental area.

Reference plane—A plane which is at a known fixed distance—which could be zero—to the tomographic plane and parallel to it.

Scan—The complete process of collecting X-ray transmission data for the production of a tomogram. Data may be collected simultaneously during a single scan for the production of one or more tomograms.

Scan increment—The amount of relative displacement of the patient with respect to the CT X-ray system between successive scans measured along the direction of the displacement.

Scan sequence—A preselected set of two or more scans performed consecutively under preselected CT conditions of operation.

Scan time—The period of time between the beginning and end of X-ray transmission data accumulation for a single scan.

Sensitivity profile—The relative response of the CT X-ray system as a function of position along a line perpendicular to the tomographic plane.

Single tomogram system—A CT X-ray system which obtains X-ray transmission data during a scan to produce a single tomogram.

Technique factors—The conditions of operation, specified as follows:

(i) For CT equipment designed for pulsed operations, peak tube potential, scan time in seconds, X-ray pulse width in seconds and the number of X-ray pulses per second or per mAs.

(ii) For CT equipment not designed for pulsed operation, peak tube potential, and either tube current and scan time in seconds or the product of tube current and exposure time in mAs.

Tomogram—The depiction of the X-ray attenuation properties of a section through a body.

Tomographic plane—The geometric plane which is identified as corresponding to the output tomogram.

Tomographic section—The volume of an object whose X-ray attenuation properties are imaged in a tomogram.

Authority

The provisions of this § 221.201 issued under sections 301 and 302 of the Radiation Protection Act (35 P. S. §§ 7110.301 and 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20).

Source

The provisions of this § 221.201 adopted October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894.

Cross References

This section cited in 25 Pa. Code § 221.204 (relating to radiation measurements and performance evaluations).

§ 221.202. Equipment requirements.

(a) *Termination of exposure.* The operator shall be able to terminate the X-ray exposure at any time during a scan, or series of scans under X-ray system control, of greater than 0.5 second duration. Termination of the X-ray exposure shall necessitate resetting of the conditions of operation prior to initiation of another scan.

(b) *Tomographic plane indication and alignment.*

(1) For any single tomogram system, a means shall be provided to permit visual determination of the tomographic plane or a reference plane offset from the tomographic plane.

(2) For any multiple tomogram system, a means shall be provided to permit visual determination of the location of a reference plane. This reference plane may be offset from the location of the tomographic plane.

(c) *Status indicators and control switches.*

(1) The CT X-ray control and gantry shall provide visual indication whenever X-rays are produced and, if applicable, whether the shutter is open or closed.

(2) The emergency buttons or switches shall be clearly labeled as to their function.

(3) Each individual scan or series of scans shall require initiation by the operator.

(d) *Indication of CT conditions of operation.* The CT X-ray system shall be designed so that the CT conditions of operation to be used during a scan or a scan sequence are indicated prior to the initiation of a scan or a scan sequence. On equipment having all or some of these conditions of operation at fixed values, this requirement may be met by permanent markings. Indication of CT conditions of operation shall be visible from any position from which scan initiation is possible.

(e) *Leakage radiation.* The leakage radiation from the diagnostic source assembly measured at a distance of 1 meter in any direction from the source may not exceed 100 milliroentgens (25.8 $\mu\text{C/kg}$) in 1 hour when the X-ray tube is

operated at its leakage technique factors. Compliance shall be determined by measurements averaged over an area of 100 square centimeters with no linear dimension greater than 20 centimeters.

(f) *Beam quality.* The HVL shall be at least 3.2 millimeters aluminum at 120 kVp.

(g) *Additional requirements applicable to CT X-ray systems containing a gantry manufactured after September 3, 1985.*

(1) The total error in the indicated location of the tomographic plane or reference plane by the light field or laser indicator may not exceed 5 millimeters.

(2) If the X-ray production period is less than 0.5 second, the indication of X-ray production shall be actuated for at least 0.5 second. Beam-on and shutter status indicators at or near the gantry shall be discernible from any point external to the patient opening where insertion of any part of the human body into the primary beam is possible.

(3) The CT X-ray system shall be normalized to water.

(4) The CT number for water for a region of interest, not exceeding 100 square millimeters, shall be 0 ± 10.0 CT number units. The facility's performance phantom shall be utilized, with the technique factors specified by the qualified expert, to confirm compliance. In instances when a CTN of 0 for water is inappropriate, as in 3D treatment planning, the qualified expert may establish and maintain an equivalent value.

(5) With the performance phantom, the mean CT number of water of one group of pixels may not differ from the mean CT number of water of a second group of pixels equal size within the same image by more than the manufacturer's published specifications.

(6) The noise, utilizing the facility's performance phantom, may not exceed the manufacturer's published specifications.

(7) The total error between the indicated and actual slice thickness may not exceed 2.0 millimeters.

(8) A distance of at least 100 millimeters measured in a CT image shall agree with the actual distance to within $\pm 5\%$.

(9) Premature termination of the X-ray exposure by the operator shall necessitate resetting the CT conditions of operation prior to the initiation of another scan.

(h) *Exemption of CT units used solely for therapy simulations.* CT units used solely for therapy simulations are exempt from this section and §§ 221.203—221.205.

Authority

The provisions of this § 221.202 issued under sections 301 and 302 of the Radiation Protection Act (35 P. S. §§ 7110.301 and 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20).

Source

The provisions of this § 221.202 adopted October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894; amended November 16, 2001, effective November 17, 2001, 31 Pa.B. 6282. Immediately preceding text appears at serial pages (249314) to (249315).

Cross References

This section cited in 25 Pa. Code § 221.201 (relating to definitions); and 25 Pa. Code § 221.204 (relating to radiation measurements and performance evaluations).

§ 221.203. Facility design requirements.

(a) *Oral communication.* Provision shall be made for oral communication between the patient and the operator at the control panel.

(b) *Viewing systems.*

(1) A means shall be provided to permit continuous observation of the patient during irradiation and shall be located so that the operator can observe the patient from the control panel.

(2) If the primary viewing system is by electronic means, an alternate viewing system, which may be electronic, shall be available for use in the event of failure of the primary viewing system.

Authority

The provisions of this § 221.203 issued under sections 301 and 302 of the Radiation Protection Act (35 P. S. §§ 7110.301 and 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20).

Source

The provisions of this § 221.203 adopted October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894.

Cross References

This section cited in 25 Pa. Code § 221.201 (relating to definitions); and 25 Pa. Code § 221.202 (relating to equipment requirements).

§ 221.204. Radiation measurements and performance evaluations.

(a) *Radiation measurements.*

(1) The CTDI or MSDAD along the two axes specified in paragraph (2)(ii) shall be measured. The CT dosimetry phantom shall be oriented so that the measurement point 1.0 centimeter from the outer surface and within the phantom is in the same angular position within the gantry at the point of maximum surface exposure identified. The CT conditions of operation shall be reproducible and correspond to typical values used by the registrant. If the point of maximum surface exposure constantly changes due to system design, then

measurements shall be taken at four different locations—top left, top right, bottom left, bottom right—1 centimeter from the outer surface of the phantom.

(2) CT dosimetry phantoms shall be used in determining the radiation output of a CT X-ray system. The phantoms shall meet the definition for a CT dosimetry phantom under 21 CFR 1020.33(b)(6) (relating to computed tomography (CT) equipment).

(i) The phantoms shall be specifically designed for CT dosimetry and deemed appropriate by the facility's qualified expert and the Department.

(ii) CT dosimetry phantoms shall provide a means for the placement of dosimeters along the axis of rotation and along a line parallel to the axis of rotation 1.0 centimeter from the outer surface and within the phantom. The means for the placement of dosimeters or alignment devices at other locations may be provided.

(iii) Any effects on the doses measured due to the removal of phantom material to accommodate dosimeters shall be accounted for through appropriate corrections to the reported data or included in the statement of maximum deviation for the values obtained using the phantom.

(iv) Dose measurements shall be performed with the CT dosimetry phantom placed on the patient couch or support device without additional attenuation materials present.

(3) In addition to the items in subsection (b), the following items shall be evaluated annually or after any component repair or change which in the opinion of the qualified expert may effect the performance of the CT unit:

(i) HVL (half value layer) determination at the most commonly used kVp or 120 kVp.

(ii) CTDI or MSAD as specified in § 221.201 (relating to definitions) for commonly used techniques.

(iii) Tomographic plane indication (light/laser alignment).

(iv) Slice thickness as specified in § 221.202(g)(7) (relating to equipment requirements).

(v) Distance readout calibration.

(4) The measurement of the radiation output of a CT X-ray system shall be performed with a dosimetry system that has calibration traceable to National Institute of Standards and Technology. The calibration of the system shall be in accordance with an established calibration protocol. The calibration protocol published by the AAPM is accepted as an established protocol. Other protocols which are equivalent will be accepted, but the user shall submit that protocol to the Department for concurrence that the protocol is equivalent.

(5) An mR/mAs value shall be determined at least annually for the head and body.

(6) Procedures and results shall be maintained for 5 years and be available for review by the Department.

(b) *Performance evaluations.*

(1) Written performance evaluation procedures shall be developed by a qualified expert. These procedures shall be available for review by the Department.

(2) The performance evaluation procedures shall include at least the following using the facility's performance phantom:

- (i) Noise.
- (ii) Contrast scale.
- (iii) Spatial resolution (low and high contrast).
- (iv) Mean CT number for water.
- (v) Acceptable tolerances.

(3) The performance evaluation shall be performed at intervals not to exceed 3 months by the qualified expert or an individual designated by the qualified expert.

(4) The qualified expert need not be present during the performance evaluation, but shall be informed within 48 hours of any problems or unacceptable deviations.

(5) Performance evaluations shall include acquisition of images obtained with the performance phantom using the same processing mode and CT conditions of operation as are used to perform the measurements required by subsection (a).

(6) Records of the performance evaluations shall be maintained for inspection by the Department for at least 4 years.

Authority

The provisions of this § 221.204 issued under sections 301 and 302 of the Radiation Protection Act (35 P. S. §§ 7110.301 and 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20).

Source

The provisions of this § 221.204 adopted October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894.

Cross References

This section cited in 25 Pa. Code § 221.201 (relating to definitions); and 25 Pa. Code § 221.202 (relating to equipment requirements).

§ 221.205. Operating procedures.

(a) Information shall be available at the control panel regarding the operation and performance evaluations of the system. The information shall include the following:

- (1) The dates of the latest radiation measurements and performance evaluation and the location within the facility where the results of those tests may be obtained.

(2) Instructions on the use of the CT phantoms including a schedule of performance evaluations appropriate for the system, allowable variations for the indicated parameters and the results of at least the most recent performance evaluation conducted on the system.

(3) The distance in millimeters between the tomographic plane and the reference plane if the reference plane is utilized.

(4) A current technique chart available at the control panel which specifies for each routine examination the CT conditions of operation and the number of scans per examination.

(b) If the radiation measurements and performance evaluation of the CT X-ray system indicates that a system operating parameter has exceeded a tolerance established by the qualified expert, the use of the CT X-ray system on patients shall be limited to those uses permitted by established written instructions of the qualified expert.

Authority

The provisions of this § 221.205 issued under sections 301 and 302 of the Radiation Protection Act (35 P. S. §§ 7110.301 and 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20).

Source

The provisions of this § 221.205 adopted October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894.

Cross References

This section cited in 25 Pa. Code § 221.201 (relating to definitions); and 25 Pa. Code § 221.202 (relating to equipment requirements).

**APPENDIX A
DETERMINATION OF COMPETENCE**

The following are areas in which an individual shall have expertise for the competent operation of diagnostic X-ray equipment:

- (1) *Familiarization with equipment.*
 - (i) Identification of controls.
 - (ii) Function of each control.
 - (iii) How to use a technique chart.
- (2) *Radiation protection.*
 - (i) Collimation.
 - (ii) Filtration.
 - (iii) Gonad shielding and other patient protection devices if used.
 - (iv) Restriction of X-ray tube radiation to image receptor.
 - (v) Personnel protection.
 - (vi) Grids.
 - (vii) Proper use of personnel dosimetry, if required.
 - (viii) Understanding units of radiation.
- (3) *Film processing.*
 - (i) Film speed as related to patient exposure.
 - (ii) Film processing parameters.
 - (iii) Quality assurance program.
 - (iv) Identification of film artifacts and corrective actions, if necessary.
 - (v) Identification of adequate film exposure on the resultant radiograph, and corrective actions, if necessary.
- (4) *Procedures.*
 - (i) Knowledge of anatomy and physiology.
 - (ii) Knowledge of positioning and radiographic demonstration of the requested anatomy with corrective actions, if necessary.
- (5) *Emergency procedures.* Termination of exposure in event of automatic timing device failure.
- (6) *Continuing education.* Continuing education annually to include radiation protection.

Authority

The provisions of this Appendix A issued under sections 301 and 302 of the Radiation Protection Act (35 P. S. §§ 7110.301 and 7110.302); and section 1920-A of The Administrative Code of 1929 (71 P. S. § 510-20).

Source

The provisions of this Appendix A adopted October 2, 1998, effective October 3, 1998, 28 Pa.B. 4894.

[Next page is 222-1.]

221-48