



INDEV GAUGING SYSTEMS

P-8

99990001/2002001

TO: Judith Joustra  
COMPANY: USNRC - Region I  
FAX: 610-337-5269  
Phone: 610-337-5355  
DATE: 10/29/03

FROM: David Smalley  
FAX: 815.282.4797  
Phone: 815.282.4463  
PAGES: 3

Judy:

Here is a copy of the letter going to Pepper and a copy of our radioactive shipment receipt report form.

If you need any other information, let me know.

Regards,

A handwritten signature in black ink, appearing to read "David Smalley". The signature is stylized with a large, looped "D" and a long, sweeping "S" that extends to the right.

David Smalley  
Operations Manager/RSO

NMSS/RGNI MATERIALS-004



## INDEV GAUGING SYSTEMS

October 29, 2003

Mr. Steve Pepper  
Pepper Environmental Services  
2701 East Tioga Street  
Philadelphia, PA 19134

Mr. Pepper:

This letter is notification that Indev Gauging Systems has received the radioactive source listed below for reclamation from SD Richman & Sons in Philadelphia, PA, per the request of Pepper Environmental Services, as transferred by David Smalley of Indev Gauging Systems:

Serial Number	Isotope	Original Strength	Assay Date	Manufacturer and Model Number
4193	Kr-85	100mCi	10/84	3M - 3E40

The source was received at Indev Gauging Systems on October 29, 2003, in good condition. The receipt of this source at Indev Gauging Systems releases Pepper Environmental Services and SD Richman & Sons from all responsibility to this source.

If you have any questions regarding this letter or the transfer of the devices, please call me at 815-282-4463. Thank you.

Regards,

A handwritten signature in black ink, appearing to read "David Smalley".

David Smalley  
Radiation Safety Officer

PEPPER ENVIRONMENTAL

## RADIOACTIVE SHIPMENT RECEIPT REPORT

Company SD RICHMAN & SONS PO# 1001-2702 Date 10/29/03  
Address 2435 WHEATSHEAF LANE  
PHILADELPHIA, PA 19137

Condition of Packages	OK	Crushed	Punctures	Other
Package #1	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Package #2	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Package #3	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>

## Transport Index

Package #1 0.1 Package #2 0.1 Package #3 0.1

Measured Radiation Levels	Surface (mR/hr)	At 1 Meter (mR/hr)
Package #1	<u>9</u>	<u>0.1</u>
Package #2	<u>0.1</u>	<u>0.1</u>
Package #3	<u>0.1</u>	<u>0.1</u>

S/N 4193  
190mCi

Note: If any surface reading exceeds 200 mR/hr or any 1 meter reading exceeds 10 mR/hr, STOP! NOTIFY THE RSO IMMEDIATELY!

Do Packing Slip and contents agree? Yes ☒ No ☐ If not, notify the RSO.

Survey results of packing material and cartons only

#1 Bg mR/hr

#2 Background mR/hr

#3 Background mR/hr

WIPE TEST OF Sr-90 PACKAGE REQUIRED BY LAW! ☐

If RCA/Carrier notification was required, give time, date, and person(s) notified:

Time                      Date                      Person(s) notified                     

Indev Engineer Signature David Smalley Date 10/29/03