

From: Nima Ashkeboussi
To: Victor.Hoang@Draeger.com
Date: 11/7/03 10:13AM
Subject: Request for Additional Information

Dear Mr. Hoang,

This e-mail is in reference to your application dated May 20, 2003 and your letter dated October 27, 2003, requesting registration for Draeger Safety, Inc. Models 5000, 5100, 5600, and 5700 gas monitoring devices. In order to complete our review, please complete the additional information below:

1. Please provide the number of devices Draeger projects selling annually in the United States.
2. Please describe the quality control procedures that Draeger Safety, Inc. will perform on devices received from Draeger Germany. Specifically, regarding:
 - That the device received meets statements submitted in this application (including radiation levels, materials, dimension, etc.)
 - That all units are leak tested to 0.005 microcurie (185 Bq)
 - All units are tested for proper operation of all safety features

Please respond to this request as soon as possible. If we do not receive a response from you in 30 days, we will consider your request as having been abandoned and void the active control for your request.

If you have any questions, please contact me at 301-415-7637 or John Jankovich at 301-415-7904.

Sincerely,

Nima Ashkeboussi
Engineering Aide
Material Safety and Inspection Branch
U.S. Nuclear Regulatory Commission
Mail-Stop T8-F5
Washington, DC 20555
(301) 415-7637
naa@nrc.gov

Mail Envelope Properties (3FABB682.7F9 : 15 : 21386)

Subject: Request for Additional Information
Creation Date: 11/7/03 10:13AM
From: Nima Ashkeboussi

Created By: NAA@nrc.gov

Recipients	Action	Date & Time
Draeger.com	Transferred	11/07/03 10:13AM
Victor.Hoang (Victor.Hoang@Draeger.com)		

Post Office	Delivered	Route
		Draeger.com

Files	Size	Date & Time
MESSAGE	2819	11/07/03 10:13AM

Options

Auto Delete: No
Expiration Date: None
Notify Recipients: Yes
Priority: Standard
Reply Requested: No
Return Notification: None

Concealed Subject: No
Security: Standard

To Be Delivered: Immediate
Status Tracking: Delivered & Opened