

U.S. NUCLEAR REGULATORY COMMISSION		Conversation Date: 9-16-04	
TELEPHONE CONVERSATION RECORD		Time: 3 pm	
Mail Control No.:	License No.:	Docket No.:	
135632	37-01374-03	030-02983	
Licensee/Applicant Participant(s):	Organization:	Telephone No.:	
Dr. Chen, Chief Medical Physicist	Mercy Hospital	973-568-9495	
Person(s) Calling: Penny Lanzisera			
Subject: amendment to license			
<p>Summary:</p> <p>Requested that Dr. Chen clarify the following:</p> <ol style="list-style-type: none"> 1) Location of Clinac 4 and Clinac 2100 vault in relation to previous location. Dr. Chen confirmed that vaults are on basement floor of the General Services Building. 2) Confirm whether Clinac 2100 vault is the same as the Clinac 20 vault indicated on the shielding diagrams. Dr. Chen confirmed this. 3) Provide letter from management of Mercy Hospital confirming that they concur with statements made in the September 2 and 15 letters. 4) Indicate if the new Nucletron HDR sources listed for use in the unit are required. 5) Dr. Chen stated that "source calibration" is calibration of source for medical use. 6) Confirm that the descriptions of the: i) warning systems & restricted area controls (locks, signs, warning lights, alarms, and door interlocks); (ii) area radiation monitoring equipment and location; and (iii) audio-visual equipment and location will meet the commitments made in the 5-29-01 application for each new room. Dr. Chen described all of the controls and equipment, which met the regulatory requirements. 7) Describe the steps taken to ensure that the Clinacs are not operated during HDR use. Dr. Chen will submit information on the toggle switch installed. <p>Dr. Chen will submit the information to answer items 3, 4, 6, and 7 as soon as possible, and understands that a Corrected Copy of Amendment 65 will be issued.</p>			
Action Required/Taken: await response and issue revised license amendment			
Prepared By: Penny Lanzisera		Date: 9-16-04	