

U.S. NUCLEAR REGULATORY COMMISSION		Conversation Date: 10-26-04	
<b>TELEPHONE CONVERSATION RECORD</b>		Time:	
Mail Control No.: 135216	License No.: 37-12110-02	Docket No.: 030-031557	
Licensee/Applicant Participant(s): Janet Bryant, Consultant for Montgomery Hospital	Organization: Montgomery Hospital	Telephone No.: 610-270-2000-X4612	
Person(s) Calling: T. Weidner			
Subject: License Renewal			
<p>Summary:</p> <p>Ms. Bryant will provide the following additional information:</p> <ol style="list-style-type: none"> <li>1. For types of use permitted by 10 CFR 35.300 and 35.400, describe the room(s) where patients will be housed if they cannot be released under 10 CFR 35.75. The discussion should include room numbers, adjacent areas and rooms, including those above and below the patient's room and a description of shielding, if applicable,</li> <li>2. For use permitted by 10 CFR 35.400; <ol style="list-style-type: none"> <li>a. indicate the location where sealed sources will be stored.</li> <li>b. describe the emergency response equipment available for manual brachytherapy facilities.</li> </ol> </li> <li>3. In your application dated June 24, 2004, you listed yttrium-90 SIR-Spheres, however you did not include information on this program. Please confirm that you intend to continue the program described in your letters dated November 5, 2002, February 13, 2003, March 7, 2003 and March 27, 2003.</li> </ol>			
Action Required/Taken: Wait for fax			
Prepared By: T. Weidner		Date: 10-26-04	