



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION I
475 ALLENDALE ROAD
KING OF PRUSSIA, PENNSYLVANIA 19406-1415

November 22, 2004

Docket No. 03005249
Control No. 135220

License No. 29-01035-01

Shaaban F. ElNaggar, Ph.D.
Radiation Safety Officer
FMC Corporation
Chemical Research & Development Center
P.O. Box 8
Princeton, NJ 08543

SUBJECT: FMC CORPORATION, ISSUANCE OF CORRECTED COPY OF LICENSE,
CONTROL NO. 135220

Dear Mr. ElNaggar:

Enclosed is the Corrected Copy of Amendment No. 32 for License No. 29-01035-01. In accordance with the telephone call on November 22, 2004, Condition No. 8. I. has been changed to 0.001 millicuries rather than 0.001 microcuries.

We apologize for any inconvenience this error may have caused.

Sincerely,

Original signed by John D. Kinneman

John D. Kinneman, Chief
Security and Industrial Branch
Division of Nuclear Materials Safety

Enclosure:
Corrected Copy of Amendment No. 32

S. ElNaggar
FMC Corporation

2

DOCUMENT NAME: E:\Filenet\ML043340129.wpd

To receive a copy of this document, indicate in the box: "C" = Copy w/o attach/encl "E" = Copy w/ attach/encl "N" = No copy

OFFICE	DNMS/RI	N	DNMS/RI	N	DNMS/RI			
NAME	JJohansen /JMJ/		JKinneman /JDK/					
DATE	11/22/04		11/22/2004					

OFFICIAL RECORD COPY

MATERIALS LICENSE

Corrected Copy

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 36, 39, 40, and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations, and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

<p>Licensee</p> <p>1. FMC Corporation Chemical Research and Development Center</p> <p>2. P. O. Box 8 Princeton, New Jersey 08543</p>	<p>In accordance with the application dated June 21, 2004,</p> <p>3. License number 29-01035-01 is amended in its entirety to read as follows:</p> <hr/> <p>4. Expiration date November 30, 2014</p> <hr/> <p>5. Docket No. 030-05249 Reference No.</p>
--	---

<p>6. Byproduct, source, and/or special nuclear material</p> <p>A. Hydrogen 3</p> <p>B. Carbon 14</p> <p>C. Phosphorus 32</p> <p>D. Sulfur 35</p> <p>E. Chlorine 36</p> <p>F. Calcium 45</p> <p>G. Iodine 125</p> <p>H. Nickel 63</p> <p>I. Cesium 137</p>	<p>7. Chemical and/or physical form</p> <p>A. Any</p> <p>B. Any</p> <p>C. Any</p> <p>D. Any</p> <p>E. Any</p> <p>F. Any</p> <p>G. Non-volatile Compounds</p> <p>H. Plated Foils</p> <p>I. Sealed Source</p>	<p>8. Maximum amount that licensee may possess at any one time under this license</p> <p>A. 1 curie</p> <p>B. 2 curies</p> <p>C. 20 millicuries</p> <p>D. 20 millicuries</p> <p>E. 20 millicuries</p> <p>F. 10 millicuries</p> <p>G. 10 millicuries</p> <p>H. 800 millicuries</p> <p>I. 0.001 millicuries</p>
--	---	---

9. Authorized use:

- A. Through G. Research and development as defined in 10 CFR 30.4; animal studies.
- H. To be used for sample analysis in compatible gas chromatography devices that have been registered either with the U.S. Nuclear Regulatory Commission under 10 CFR 32.210 or with an Agreement State and have been distributed in accordance with a Commission or Agreement State specific license authorizing distribution to persons specifically authorized by a Commission or Agreement State license to receive, possess, and use the devices.
- I. Calibration and checking of the licensee's instruments.

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**License Number
29-01035-01Docket or Reference Number
030-05249Amendment No. 32
Corrected Copy**CONDITIONS**

10. Licensed material may be used or stored only at the licensee's facilities located at U. S. Route 1 and Plainsboro Road, Princeton, New Jersey.
11. Licensed material shall be used by, or under the supervision of, John M. Becker, Bruce Black, M. K. Chaguturu, Audrey W. Chen, Peter Dierks, James R. Eldridge, Sha'aban El Naggar, Robert Henrie, David Kiefer, Lyle P. Kinne, G. Russell Peters, or Sandra Simpson.
12. The Radiation Safety Officer for this license is Sha'aban El Naggar.
13. The licensee shall not use licensed material in or on human beings.
14. The licensee shall not use licensed material in field applications where it is released except as provided otherwise by specific condition of this license.
15. Experimental animals, or the products from experimental animals, that have been administered licensed materials shall not be used for human consumption.
16.
 - A. Sealed sources shall be tested for leakage and/or contamination at intervals not to exceed the intervals specified in the certificate of registration issued by the U.S. Nuclear Regulatory Commission under 10 CFR 32.210 or under equivalent regulations of an Agreement State.
 - B. In the absence of a certificate from a transferor indicating that a leak test has been made within the intervals specified in the certificate of registration issued by the U.S. Nuclear Regulatory Commission under 10 CFR 32.210 or under equivalent regulations of an Agreement State, prior to the transfer, a sealed source received from another person shall not be put into use until tested and the test results received.
 - C. Sealed sources need not be tested if they contain only hydrogen-3; or they contain only a radioactive gas; or the half-life of the isotope is 30 days or less; or they contain not more than 100 microcuries of beta- and/or gamma-emitting material or not more than 10 microcuries of alpha-emitting material.
 - D. Sealed sources need not be tested if they are in storage and are not being used; however, when they are removed from storage for use or transferred to another person and have not been tested within the required leak test interval, they shall be tested before use or transfer. No sealed source shall be stored for a period of more than 10 years without being tested for leakage and/or contamination.
 - E. The leak test shall be capable of detecting the presence of 0.005 microcurie (185 becquerels) of radioactive material on the test sample. If the test reveals the presence of 0.005 microcurie (185 becquerels) or more of removable contamination, a report shall be filed with the U.S. Nuclear Regulatory Commission in accordance with 10 CFR 30.50(c)(2), and the source shall be removed

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**License Number
29-01035-01Docket or Reference Number
030-05249Amendment No. 32
Corrected Copy

immediately from service and decontaminated, repaired, or disposed of in accordance with Commission regulations.

- F. Tests for leakage and/or contamination, including leak test sample collection and analysis, shall be performed by the licensee or by other persons specifically licensed by the U.S. Nuclear Regulatory Commission or an Agreement State to perform such services.
- G. Records of leak test results shall be kept in units of microcuries and shall be maintained for 5 years.
17. Sealed sources or detector cells containing licensed material shall not be opened or sources removed from source holders by the licensee.
18. The licensee shall conduct a physical inventory every six months, or at other interval approved by the U.S. Nuclear Regulatory Commission, to account for all sealed sources and/or devices received and possessed under the license.
19. Maintenance, repair, cleaning, replacement, and disposal of foils contained in detector cells shall be performed only by the device manufacturer or other persons specifically authorized by the U.S. Nuclear Regulatory Commission or an Agreement State to perform such services.
20. The licensee is authorized to hold radioactive material with a physical half-life of less than or equal to 120 days for decay-in-storage before disposal in ordinary trash, provided:
- A. Waste to be disposed of in this manner shall be held for decay a minimum of 10 half-lives.
- B. Before disposal as ordinary trash, the waste shall be surveyed at the container surface with the appropriate survey instrument set on its most sensitive scale and with no interposed shielding to determine that its radioactivity cannot be distinguished from background. All radiation labels shall be removed or obliterated.
- C. A record of each such disposal permitted under this license condition shall be retained for 3 years. The record must include the date of disposal, the date on which the byproduct material was placed in storage, the radionuclides disposed, the survey instrument used, the background dose rate, the dose rate measured at the surface of each waste container, and the name of the individual who performed the disposal.
21. The licensee is authorized to transport licensed material in accordance with the provisions of 10 CFR Part 71, "Packaging and Transportation of Radioactive Material."
22. Except as specifically provided otherwise in this license, the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the documents, including any enclosures, listed below. The U.S. Nuclear Regulatory Commission's regulations shall govern unless the statements, representations, and procedures in the licensee's application and correspondence are more restrictive than the regulations.

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**License Number
29-01035-01Docket or Reference Number
030-05249Amendment No. 32
Corrected Copy

A. Application dated June 21, 2004



For the U.S. Nuclear Regulatory Commission

Date November 22, 2004

By

Original signed by Jenny JohansenJenny Johansen
Security and Industrial Branch
Division of Nuclear Materials Safety
Region I
King of Prussia, Pennsylvania 19406