



UNITED STATES  
**NUCLEAR REGULATORY COMMISSION**  
REGION I  
475 ALLENDALE ROAD  
KING OF PRUSSIA, PENNSYLVANIA 19406-1415

December 14, 2004

Docket No. 03033893  
Control No. 135880

License No. 45-25339-01MD

Kory Kodimer, Ph.D.  
Manager, Health Physics  
Cardinal Health  
1807-H Murry Road  
Roanoke, VA 24018

SUBJECT: CARDINAL HEALTH, REQUEST FOR ADDITIONAL INFORMATION  
CONCERNING APPLICATION FOR AMENDMENT TO LICENSE, CONTROL  
NO. 135880

Dear Mr. Kodimer:

This is in reference to your letter dated October 13, 2004 requesting to amend Nuclear Regulatory Commission License No. 45-25339-01MD. In order to continue our review, we need the following additional information:

1. The information you submitted concerning the remodeling of your facility at 713-B Harris Street, Charlottesville, VA where a restricted area will become an office did not include the residual radiation levels and the levels of residual removable contamination. Please submit the following additional information:
  - a. results of surveys made for radiation levels in milliroentgen per hour (mR/hr) and of wipe test surveys made for removable contamination in disintegrations per minute per 100 square centimeters (dpm/100 cm<sup>2</sup>), and
  - b. the manufacturer, model number, and sensitivity of the survey instruments used to perform these measurements.

Please note that on October 25, 2004, the NRC suspended public access to ADAMS, and initiated an additional security review of publicly available documents to ensure that potentially sensitive information is removed from the ADAMS database accessible through the NRC's web site. Interested members of the public may obtain copies of the referenced documents for review and/or copying by contacting the NRC Public Document Room pending resumption of public access to ADAMS. The NRC Public Document Room is located at NRC Headquarters in Rockville, MD, and can be contacted at 800-397-4209 or 301-415-4737 or [pdr@nrc.gov](mailto:pdr@nrc.gov).

We will continue our review upon receipt of this information. Please reply to Farrah Gaskins at the Region I Office and refer to Mail Control No. 135880. If you have any technical questions regarding this deficiency letter, please call Farrah Gaskins at (610) 337-5143.

K. Kodimer  
Cardinal Health

2

If we do not receive a reply from you within 30 calendar days from the date of this letter, we shall assume that you do not wish to pursue your application.

Sincerely,

***Original signed by Elizabeth Ullrich***

James P. Dwyer, Chief  
Commercial and R&D Branch  
Division of Nuclear Materials Safety

Enclosure:  
10 CFR Parts 19, 20, and 30

cc:  
Robert W. Beightol, Radiation Safety Officer

DOCUMENT NAME: E:\Filenet\ML043510196.wpd

To receive a copy of this document, indicate in the box: "C" = Copy w/o attach/encl "E" = Copy w/ attach/encl "N" = No copy

OFFICE	DNMS/RI	N	DNMS/RI	N	DNMS/RI			
NAME	Jdwyer/EXU for		FGaskins/FCG					
DATE	12/14/04		12/14/04					

OFFICIAL RECORD COPY