

Sr. Mercado

Central Industrial Services, Inc.

52-25470-01
030-35030
6/8/1999

Please see the X on the following pages.

I need your commitment about a radiation meter - either to buy one or to be able to get one from the local emergency authority for use in case of an accident or gauge accident. The instrument should be calibrated each 12 months and must be capable of detecting cesium 137.

I need your commitment about physical inventory every six months for all sealed sources and devices possessed and received during the period.

Please call me if you have any questions.

Thank You David Collins

258350

NMSS/RGNI MATERIALS-001