

U.S. NUCLEAR REGULATORY COMMISSION		Conversation Date: 12/29/04	
TELEPHONE CONVERSATION RECORD		Time:	
Mail Control No.: 136185	License No.: 45-00385-06	Docket No.: 03006510	
Licensee/Applicant Participant(s): Joe Tenhet, RSO	Organization: Phillip Morris USA	Telephone No.: 804-274-2946	
Person(s) Calling: J. Joustra			
Subject: letter dated 12/20/04			
<p>Summary:</p> <ol style="list-style-type: none"> 1. How many of the Molins devices do you posses? 20 but only 8 are in use. These devices will be replaced/disposed in about 3 to 4 years and the licensee will obtain devices from another manufacture. 2. How often have you exchanged sources in the past. Never but they have removed sources in order to service the devices. 3. How will the sources arrive? They will arrive in a shielded container. The source screws into the device. The source is shielded except for the window. The licensee has procedures and equipment to remove and install the sources. 4. Will any changes be made to the device. The licensee indicated that they will change the name plate on the device to reflect the correct mode/serial number of the sealed source. 5. The licensee will send a copy of attachment VZ-2530 as mentioned in the letter from AEA attached to the licensee letter. 			
Action Required/Taken: wait for attachment and complete TAR form for panel 1/6/05.			
Prepared By: J. Joustra		Date: 1/3/05	