

TELEPHONE CONVERSATION RECORD		
<b>Licensee Name:</b> Georgia-Pacific Corporation	<b>Date:</b> 1/7/2005	<b>Time:</b> 1 PM
<b>License No.:</b> 45-25050-01	<b>Docket No.:</b> 03031124	<b>Control No.:</b> 136198
<b>Licensee Contact:</b> Denise J. Fluharty, RSO	<b>Telephone No:</b> 434-634-6133	
<b>NRC Contact:</b> Kathy Dolce Modes	<b>Telephone No:</b> 610.337.5251	
<b>Subject: Request for Additional Information</b>		
<p><b>Summary:</b></p> <ol style="list-style-type: none"> <li>1. Please provide the correct model numbers of the gauges. DMC Model OSP-200-108 and OSP-200-120 are not listed in the Sealed Source and Device Registrations.</li> <li>2. Please pick one person to be designated as the RSO on the license. Item 7 in your application dated December 20, 2004 lists 3 individuals to be RSO.</li> <li>3. Items 10.1 (survey instruments) and 10.3 (dosimetry) in your application are not in conformance with NUREG-1556, Volume 4 (see Appendix B page B-4).</li> <li>4. Please provide the RSO's phone, fax, and business email address for our records.</li> </ol>		
<b>Action Required/Taken: Wait for fax.</b>		
<b>Signature:</b> Kathy Dolce Modes	<b>Date:</b> January 11, 2005	