

**From:** Michelle Beardsley  
**To:** Shashadhar.M.Mohapatra@Medstar.net  
**Date:** Fri, Jan 7, 2005 2:38 PM  
**Subject:** Amendment request dated January 3, 2005  
**Place:** LAT@nrc.gov

Control No. 136242  
License No. 08-03604-03  
Docket No. 03001325

S. Mohapatra  
Radiation Safety Officer  
Washington Hospital Center  
Washington, D.C. 20010

Dr. Mohapatra:

Regarding your request in the above letter to add a self-shielded irradiator to the license, the following additional information is needed (Please be aware that NUREG-1556, Vol. 5 is used to review applications regarding self-shielded irradiators):

1. Please confirm that you, the RSO, and future RSO's will have successfully completed the training described in Appendix G in the above listed NUREG.

2. Please confirm that physical inventories will be conducted at intervals not to exceed 6 months to account for these sealed sources.

3. Please address Item 10.6 of the above NUREG regarding Operating and Emergency procedures.

You may find a copy of this NUREG on our website at nrc.gov, under Nuclear Materials-->Quicklinks-->Guidance

If you need additional information, you may contact me either by email or phone at 610-337-6942. You may fax your response to my attention at 610-337-5393. Please reference the control number listed above on any correspondence regarding this matter.

Sincerely,  
Michelle Beardsley  
Health Physicist  
USNRC RI  
475 Allendale Road  
King of Prussia, PA 19406

**CC:** LAT@nrc.gov  
**Mail Envelope Properties**  
(41DEE536.794 : 12 : 35316)

**Subject:** Amendment request dated January 3, 2005  
**Creation Date:** Fri, Jan 7, 2005 2:38 PM  
**From:** Michelle Beardsley

**Created By:** MRB@nrc.gov

**Recipients**

Medstar.net

Shashadhar.M.Mohapatra (Shashadhar.M.Moha

nrc.gov

kp1\_po.KP\_DO

LAT CC (LAT@nrc.gov)

**Post Office**

kp1\_po.KP\_DO

**Route**

Medstar.net

nrc.gov

**Files**

MESSAGE

**Size**

2658

**Date & Time**

Friday, January 7, 2005 2:38 PM

**Options**

**Expiration Date:**

None

**Priority:**

Standard

**Reply Requested:**

No

**Return Notification:**

None

**Concealed Subject:**

No

**Security:**

Standard