



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION I
475 ALLENDALE ROAD
KING OF PRUSSIA, PENNSYLVANIA 19406-1415

January 14, 2005

Docket No. 03036776
Control No. 136113

License No. 45-30980-01

Carol P. Cardinale, M.D.
Radiation Safety Officer
Kaiser Permanente
12011 Lee-Jackson Highway
Fairfax, VA 22033

SUBJECT: KAISER PERMANENTE, REQUEST FOR ADDITIONAL INFORMATION
CONCERNING APPLICATION FOR NEW LICENSE, CONTROL NO. 136113

Dear Dr. Cardinale:

This is in reference to your application dated November 23, 2004 applying for a Nuclear Regulatory Commission license. In order to continue our review, we need the following additional information:

1. It is unclear that your application was signed by a management representative. Please confirm that you are a management representative or submit a letter signed by a management representative indicating that management has reviewed the application and concurs in the statements and representations contained therein. Note also that a management representative should sign all future correspondence that requests a change in your license.
2. In an effort to update the NRC's ability to communicate with licensees, we request that you provide your main phone number, and the phone number, facsimile number and e-mail address by which we could contact you or the Radiation Safety Officer in case of emergency.
3. Please confirm that your proposed location of use, listed in Item 3 of your application, is not a location of use for any other active NRC license.
4. You have requested that Carol P. Cardinale, M.D. be named Radiation Safety Officer (RSO) on your license. Sufficient documentation has been provided concerning the training and experience specified in 10 CFR 35.900(b) and 35.50(b), however, written certification, signed by a preceptor RSO, was not provided as stated in Table C.3, Item 7, "Radiation Safety Officer." NRC Form 313A, Item 7 is marked as N/A and Item 9 is signed by Steven Bergman, M.D. as an Authorized User. Please submit another copy of NRC Form 313A completing Items 1., 5a., 7., 9., 11a., 11b., and 12. signed by the RSO for New York License No. 75-2878-01 or a signed letter by the RSO with significantly the same information.

C. Cardinale
Kaiser Permanente

2

5. Your application does not identify adjacent areas across the walls from use and storage locations. Please revise your diagram to indicate what those adjacent areas include.

Please note that on October 25, 2004, the NRC suspended public access to ADAMS, and initiated an additional security review of publicly available documents to ensure that potentially sensitive information is removed from the ADAMS database accessible through the NRC's web site. Interested members of the public may obtain copies of the referenced documents for review and/or copying by contacting the NRC Public Document Room pending resumption of public access to ADAMS. The NRC Public Document Room is located at NRC Headquarters in Rockville, MD, and can be contacted at 800-397-4209 or 301-415-4737 or pdr@nrc.gov.

We will continue our review upon receipt of this information. Please reply to my attention at the Region I Office and refer to Mail Control No. 136113. If you have any technical questions regarding this deficiency letter, please call me at (610) 337-5075.

If we do not receive a reply from you within 30 calendar days from the date of this letter, we shall assume that you do not wish to pursue your application.

Sincerely,

Original signed by Steven Courtemanche

Steven Courtemanche
Health Physicist
Commercial and R&D Branch
Division of Nuclear Materials Safety

Enclosures:

1. 10 CFR Parts 19, 20, and 30
2. NRC Form 313A

cc:

Jamey Andrews, Clinical Operations Manager

C. Cardinale
Kaiser Permanente

3

DOCUMENT NAME: E:\Filenet\ML050240068.wpd

To receive a copy of this document, indicate in the box: "C" = Copy w/o attach/encl "E" = Copy w/ attach/encl "N" = No copy

OFFICE	DNMS/RI	N	DNMS/RI		DNMS/RI			
NAME	SCourtemanche/src							
DATE	01/14/2005							

OFFICIAL RECORD COPY