

U.S. NUCLEAR REGULATORY COMMISSION		Conversation Date: 01/05/05	
TELEPHONE CONVERSATION RECORD		Time: 10:30 a.m.	
Mail Control No.:	License No.:	Docket No.:	
135354	08-07398-03	03014754	
Licensee/Applicant Participant(s):	Organization:	Telephone No.:	
Susan F. Krueger-Schmidt	Krueger-gilbert Health Physics, Inc.	410-665-5447	
Person(s) Calling: Randolph C. Ragland, Jr., 610-337-5083 office, 610-310-7799 cell			
Subject: Sibley Memorial Hospital, 7/14/04 License Renewal Application			
<p>Summary:</p> <p>Ms. Krueger-Schmidt confirmed the following:</p> <ol style="list-style-type: none"> 1) Sibley Memorial Hospital would like to retain all Authorized Users and Medical Physicists as listed on NRC License 08-07398-03 Amendment No. 23. 2) Each Am-241 sealed source appearing on the NRC License 08-07398-03, Amendment 23 has an activity less than 30 millicuries. We explained to Ms Krueger-Schmidt that possession of this isotope quantity is covered under 10 CFR 35.65 "Calibration, reference, and transmission sources" and no longer needs to be listed as a line item on the license, and will be removed when the license is renewed. 3) Each Barium 133 source included in NRC License 08-07398 has an activity less than 30 millicuries. Ms. Krueger-Schmidt added that these sources are currently not in use and are in being stored in preparation for a future radioactive waste shipment. We explained to Ms Krueger-Schmidt that possession of this isotope quantity is covered under 10 CFR 35.65 "Calibration, reference, and transmission sources" and no longer needs to be listed as a line item on the license. Accordingly, authorization for "Barium 133 permitted by 10 CFR 35.500 will be removed from the license when the license is renewed. 			
Action Required/Taken:		This telecon will be placed into ADAMS and referenced during license renewal.	
Prepared By: Randolph C. Ragland, Jr. Michelle R. Beardsley		Date: 01/05/05	

RM
1/14/05