

U.S. NUCLEAR REGULATORY COMMISSION		Conversation Date: 1/31/05
TELEPHONE CONVERSATION RECORD		Time: 08:40
Mail Control No.: N/A	License No.: 52-25154-01	Docket No.: 030-32370
Licensee/Applicant Participant(s): Receptionist for Jorge Hernández, M.D.	Organization: Jorge Hernandez	Telephone No.: 787-841-7030 787-844-1125 (fax)
Person(s) Calling: Randolph C. Ragland, Jr.		
Subject: NRC License and Disposition of Sr-90 Source		
<p>Summary:</p> <p>Dr. Hernández's Receptionist provided a fax number so NRC could send this message to Dr. Hernández:</p> <p>As discussed with Mr. Hector Bermúdez on January 10, 2005, NRC requires that you have your Sr-90 eye applicator calibrated in accordance with 10 CFR 35.432. You indicated that because of this requirement, you plan to discontinue use of your Sr-90 eye applicator and will take action to transfer or dispose of the source at a licensed facility. We need for you to take the following action:</p> <p>Please fax NRC a license amendment request @ 610-337-5269 requesting NRC to amend your license to place your license into a "Possession Only" condition, and state that you plan to take action to transfer your Sr-90 eye applicator to a licensed facility or licensed disposal site.</p> <p>Regards,</p> <p>Randolph C. Ragland, Jr. NRC Region I 475 Allendale Road King of Prussia, PA 19406</p>		
Action Required/Taken: Mr. Ragland will call Dr. Hernández this afternoon to ask Dr. Hernández to fax NRC a license amendment as discussed above.		
Prepared By: Randolph C. Ragland, Jr.		Date: 1/31/05