

U.S. NUCLEAR REGULATORY COMMISSION		Conversation Date: 2-1-05	
<b>TELEPHONE CONVERSATION RECORD</b>		Time:	
Mail Control No.:	License No.:	Docket No.:	
136365	29-03047-01	03002465	
Licensee/Applicant Participant(s):	Organization:	Telephone No.:	
Marc Belanich	PBI Reg. Med. Ctr.		
Person(s) Calling: Michelle Beardsley , RI			
Subject: Letter dated January 28, 2005			
<p>Summary:</p> <p>Regarding the above, I requested the following additional information:</p> <ol style="list-style-type: none"> <li>1. Confirm that all operators, users will receive vendor training;</li> <li>2. Submit detailed spot-check and emergency procedures;</li> <li>3. Submit diagram and exposure calculations for all adjacent areas;</li> <li>4. Specify who will perform ,and where maintenance/repair and source exchanges will occur, i.e. licensee, AROS, Varian.</li> <li>5. Dr. Arbor is not authorized for 35.300 on license submitted, please submit evidence of continuing ed./experience for these materials;</li> <li>6.Dr. Brown is only auth. User for 31.11 materials; please specify another au, delete these from license, or use under general license;</li> <li>7. Specify range of survey meter used for HDR;</li> <li>8. Submit copy of receipt confirmation of IVB sources in order to remove them from license.</li> </ol>			
Action Required/Taken: Submit additional info.			
Prepared By: M. Beardsley		Date: 2-2-05	