

TELEPHONE CONVERSATION RECORD	Date: January 25, 2005	Time: 14:00
Mail Control No.: N/A Inspection No.: N/A	License No.: 47-18437-01	Docket No.: 030-15079
Person Called: Jack M. Pryor, Corporate Radiation Safety Officer	Licensee: U. S. Silica Company	Telephone No.: 304-258-8250
Person Calling: Steven Courtemanche/(610) 337-5075		
Subject: Request to retire NRC license.		
<p><u>Summary:</u> I spoke with Mr. Pryor about the licensee's letter dated January 7, 2005, and the telephone evaluation of March 26, 2003. I informed Mr. Pryor that the evaluation indicated that the licensee possessed only a Hewlett Packard Model 5890 gas chromatograph while the letter indicated that the licensee disposed of a Perkin Elmer gas chromatograph. Mr. Pryor stated that upon review the Hewlett Packard unit was possessed under a general license and the Perkin Elmer unit was specifically-licensed. The Perkin Elmer unit had been in storage for a number of years because it was non-functional. I asked Mr. Pryor about the attachment to the letter from Chase Environmental Group, Inc. The attachment describes a TN Technologies Model 5084 from Mapleton, PA and an ASOMA XRF unit Model 8620 being picked up for disposal. Mr. Pryor indicated that the items were generally-licensed devices which the licensee no longer needed. I stated that if the units were being disposed of without replacements then a notice needed to be sent to the Director of NMSS in NRC Headquarters. Mr. Pryor stated that he would look into the matter and confirm whether the required notice had been sent in to NRC Headquarters.</p>		
Action Required/Taken: Place into ADAMS.		
Prepared by Steven Courtemanche	Date: January 27, 2005	