

NOTE TO FILE

Licensee: St. Francis Medical Center  
License No.: 29-00232-08  
Docket No.: 03002423  
Mail Control: 136433

- 1) Additional request letter dated March 18, 2005 shows closeout survey wipe tests in net cpm rather than net dpm. I am accepting these based on very low net cpm in comparison with background, plus use of only short half-life material. In cover letter, will direct licensee to submit future closeout survey results in units of net dpm per 100 square cm..
- 2) Deficiency letter dated April 8, 2005 states that the Sr-90 eye applicator has been shipped for disposal, however no documentation was provided. In cover letter, will inform licensee that they must submit documentation of receipt of source by authorized recipient in order to remove this source from license.

S. Gabriel  
Senior Health Physicist  
4/11/2005