

ROBERT WOOD JOHNSON  
UNIVERSITY HOSPITAL  
HAMILTON

March 9, 2005

United States Nuclear Regulatory Commission  
Licensing Assistance Section  
475 Allendale Road  
King of Prussia, PA 19406-1415

29-16145-01  
03010491


05 MAR 18 P 1:21

RECEIVED  
REGION 1

To Whom It May Concern,

Robert Wood Johnson University Hospital at Hamilton, license number 29-16145-01 wishes to submit the final survey and wipe test of our cardiac stress testing room, prior to release as an unrestricted area. Should you have any questions, please contact our Radiation Safety Officer, Ms. Janet Bryant, MS through her office at 908-788-9440.

Sincerely,

  
Christy Stephenson  
President and Chief Executive Officer

cc. Mr. Jim Kilmer

136743

UNRESTRICTED MATERIALS-002

Today's Wipes

		Net Counting Rate		Net Activity	
stress loc 6 <i>Counter</i>	Background Full Spectrum	50.00 cpm	03/02/2005 10:54 141.0 cpm	60 sec	Trigger: 2000 dpm
				57.47 dpm	
stress loc 5 <i>Counter</i>	Background Full Spectrum	66.00 cpm	03/02/2005 10:53 141.0 cpm	60 sec	Trigger: 2000 dpm
				75.86 dpm	
stress loc 4 <i>Floor</i>	Background Full Spectrum	81.00 cpm	03/02/2005 10:52 141.0 cpm	60 sec	Trigger: 2000 dpm
				93.10 dpm	
stress loc 3 <i>Treadmill</i>	Background Full Spectrum	48.00 cpm	03/02/2005 10:50 141.0 cpm	60 sec	Trigger: 2000 dpm
				55.17 dpm	
stress loc 2 <i>Floor</i>	Background Full Spectrum 74 keV	94.00 cpm 39.00 cpm	03/02/2005 10:48 141.0 cpm	60 sec	Trigger: 2000 dpm
				108.0 dpm	
stress loc 1 <i>Floor</i>	Background Full Spectrum 79 keV	170.0 cpm 78.00 cpm	03/02/2005 10:47 141.0 cpm	60 sec	Trigger: 2000 dpm
				195.4 dpm	
<del>stress loc 1</del>	Background Full Spectrum 93 keV	343.0 cpm 194.0 cpm	03/02/2005 10:45 141.0 cpm	60 sec	Trigger: 2000 dpm
				394.3 dpm	
<del>Package 3</del>	Background Full Spectrum	-10.00 cpm	03/02/2005 07:21 161.0 cpm	60 sec	Trigger: 2200 dpm
				0.000 dpm	

all areas surveyed < 2000 dpm and < 0.02 mR/hr  
Background = 0.01 mR/hr  
meter used = Ludlum 3, SN: 48147, calibrated 11/24/04  
performed by: Janet Bryant, M.S. RAB  
Medical Physicist  
Radiation Safety Officer

This is to acknowledge the receipt of your letter/application dated

3/9/2005, and to inform you that the initial processing which includes an administrative review has been performed.

☒ AMEND. 29-1645-01  
There were no administrative omissions. Your application was assigned to a technical reviewer. Please note that the technical review may identify additional omissions or require additional information.

☐ Please provide to this office within 30 days of your receipt of this card

A copy of your action has been forwarded to our License Fee & Accounts Receivable Branch, who will contact you separately if there is a fee issue involved.

Your action has been assigned Mail Control Number 136743.  
When calling to inquire about this action, please refer to this control number.  
You may call us on (610) 337-5398, or 337-5260.

BETWEEN: : (FOR LFMS USE)  
 : INFORMATION FROM LTS  
 : -----  
 :  
 License Fee Management Branch, ARM : Program Code: 02120  
 and : Status Code: 0  
 Regional Licensing Sections : Fee Category: 7C  
 : Exp. Date: 20130228  
 : Fee Comments: CODE 23  
 : Decom Fin Assur Reqd: N  
 : ::::::::::::::::::::::::::::::::::::::

LICENSE FEE TRANSMITTAL

A. REGION *I*

1. APPLICATION ATTACHED

Applicant/Licensee: ROBERT WOOD JOHNSON UNIV. HOSP. AT  
Received Date: 20050318  
Docket No: 3010491  
Control No.: 136743  
License No.: 29-16145-01  
Action Type: Amendment

2. FEE ATTACHED

Amount: \_\_\_\_\_  
Check No.: \_\_\_\_\_

3. COMMENTS

Signed *M. G. Parkin*  
Date *4/10/2005*

B. LICENSE FEE MANAGEMENT BRANCH (Check when milestone 03 is entered /\_\_/)

1. Fee Category and Amount: \_\_\_\_\_

2. Correct Fee Paid. Application may be processed for:

Amendment \_\_\_\_\_  
Renewal \_\_\_\_\_  
License \_\_\_\_\_

3. OTHER \_\_\_\_\_  
\_\_\_\_\_

Signed \_\_\_\_\_  
Date \_\_\_\_\_