



Susquehanna Health System

Divine Providence Hospital • Muncy Valley Hospital
The Williamsport Hospital & Medical Center

Richard W. DeWald
Chairman of the Board

Sister Joanne Bednar
Vice Chairman

Steven P. Johnson
President & CEO

April 25, 2005

NRC License No: 37-04185-01

03003037

J-1

U.S. Nuclear Regulatory Commission
Region I
Nuclear Materials Section B
475 Allendale Road
King Of Prussia, PA 19406

Dear Sirs,

Attached please find the closeout surveys performed by Sam Payne MS, Health Physicist Consultant. Surveys were done in the stress labs that we requested to be removed from our license and release for general use. Please reference letter dated April 19, 2005.

We are also asking for this amendment to be expedited since reconstruction is to begin on 4/25/2005. If possible please fax the amendment releasing the rooms for general use to (570) 321-2496.

Thank you for your attention to this matter.

Sincerely

A handwritten signature in black ink, appearing to read "Jeff Kemmler".

Jeff Kemmler, Director of Nuclear Medicine

NRCcloseoutstresslab TWH.word

136860

NMCG/RONI MATERIALS-CJC

**Williamsport Hospital
Williamsport, PA**

**Facility Decommissioning Survey
Department of Nuclear Medicine**

Date: April 25, 2005

This report is submitted in keeping with Nuclear Regulatory Commission Regulations:
***Guidelines for Decontaminations of Facilities and Equipment Prior to Release for
Unrestricted Use.***

The Closeout Survey was conducted to close out Stress Rooms A and B for the Williamsport Hospital, Williamsport, Pennsylvania, NRC License No.: 37-04185-01 and DEP License No: PA-0092. These areas were used to stress patients undergoing cardiology exams using primarily Tc-99m or Tl-201.

This survey was conducted utilizing multiple wipe test samples utilizing soft absorbent paper over 100 cm. sq. areas made at various locations throughout the department (see area map attached). Room surveys for ambient radiation exposure levels were also made utilizing a GM Meter.

Survey Results:

All areas wiped and surveyed for ambient exposure levels reveal that no residual contamination was detected in excess of 1000 dpm Beta or Gamma per 100 sq. cm. Refer to attached survey. All areas were indistinguishable from normal background measurements.

It is therefore recommended that this area be released for use as an unrestricted area.

Survey by:



Douglas E. Heim
Consultant Physicist

Facility Decommissioning Survey

Williamsport Hospital

April 25, 2005

Facility Decommissioning Survey Results

April 25, 2005

Location:		Removable Contamination dpm/100 cm ²	Area Survey (mR/hr)
1	Floor	33	0.03
2	Floor	75	0.03
3	Floor	25	0.02
4	Floor	84	0.03
5	Floor	53	0.03
6	Floor	50	0.02
7	Floor	84	0.03
8	Floor	20	0.04
9	Floor	54	0.03
10	Floor	82	0.03
11			
12			
13			
14			

Monitoring Equipment

Survey Instrument: Ludlum Model 14C S/N: 103879

Calibration Date: 10/11/04 Background: .03 mR/hr

Well Counter:

Captus 500 S/N:

