

TELEPHONE CONVERSATION RECORD	Date: April 22, 2005	Time: 15:00
Mail Control No.: 136639 Inspection No.: N/A	License No.: 06-28798-01	Docket No.: 030-32975
Person Called: Alan McGinnes, Chief Executive Officer	Licensee: Grove Hill Medical Center, P.C.	Telephone No.: 203-223-0220
Person Calling: Steven Courtemanche/(610) 337-5075		
Subject: Clarification of amended license.		
<p><u>Summary:</u> I spoke with Mr. McGinnis about the licensee's request to amend the license to add a patient attenuation device. I explained that all of the sources requested could not be put on the license because the Sealed Source and Device Registration for the device mentioned is only authorized for IPL Model No. NES-8497. I also mentioned that the Authorized Users and the personnel operating the device should be trained in the use of the device before its operation. Mr. McGinnis stated he would inform his medical physicist of the conversation.</p>		
Action Required/Taken: Document for ADAMS		
Prepared by Steven Courtemanche	Date: April 22, 2005	