

**From:** Richard McKinley  
**To:** eseyler@harthosp.org  
**Date:** 5/5/05 2:42PM  
**Subject:** License renewal of Lic. No. 06-14854-01,docket 03008163, control no. 136719

Dear Dr. Seyler:

Pursuant to your request to renew the above license, I need to ask for the following information:

- 1) Please send the statement regarding dose calibrator calibration contained in Section 8.17, Item 9, page 8-32.
- 2) What is above and below the I-131 delivery area?
- 3) Did you mean to say that your dose calibrator measured up to 2000 microcuries?
- 4) Please confirm that your survey instrument will measure to the highest dose rate you will encounter.
- 5) You have committed to Appendix K procedures for calibrating survey instruments. Do you intend to calibrate your own instrument? If not, please specify which vendor you currently use.
- 6) Do you have access to a backup survey instrument?
- 7) When you do contamination surveys, what instrument will you use to count wipes?

Please respond to these items by hardcopy, either through the mail or to fax number 610-337-5269. Thank you very much for your cooperation.

Sincerely,  
Richard McKinley  
Health Physicist  
USNRC, Region I

**CC:** rwm1

**Mail Envelope Properties**

(427A6ACD.F34 : 14 : 10472)

**Subject:** CTMultispecialty Group-Endocrine Division, 06-14854-01, 03008163, control 136719

**Creation Date:** Thu, May 5, 2005 2:49 PM

**From:** Richard McKinley

**Created By:** RWM1@nrc.gov

**Recipients**

nrc.gov

kp1\_po.KP\_DO

LAT (LAT)

**Post Office**

kp1\_po.KP\_DO

**Route**

nrc.gov

**Files**

MESSAGE

License renewal of Lic. No. 06-14854-01,docket 03008163, control no. 136719.mlm

**Size**

411

1972

**Date & Time**

Thursday, May 5, 2005 2:49 PM

Thursday, May 5, 2005 2:46 PM

**Options**

**Expiration Date:**

None

**Priority:**

Standard

**Reply Requested:**

No

**Return Notification:**

None

**Concealed Subject:**

No

**Security:**

Standard