

U.S. NUCLEAR REGULATORY COMMISSION		Conversation Date: 05/09/05
<b>TELEPHONE CONVERSATION RECORD</b>		Time: 1300 hours
Mail Control No.: 136679	License No.: 45-08482-01	Docket No.: 03003327
Licensee/Applicant Participant(s): Peter Paras, RSO,	Organization: Northern Virginia Radiology	Telephone No.: 301.670.1095
Person(s) Calling: S Xu, RI		
Subject: license renewal		
<p>Summary:</p> <p>Following issues have been discussed with Dr. Romero:</p> <ol style="list-style-type: none"> <li>1. Confirm their license name. The name on the application and on the current license are different.</li> <li>2. Asked him to provide us the list of their survey instruments.</li> <li>3. Asked him to provide us with their security measures of their seeds storage room.</li> <li>4. Confirm that their radiation instruments will be calibrated by a person qualified to perform survey meter calibrations.</li> <li>5. Informed him that we don't review their treatment protocol.</li> </ol>		
Action Required/Taken: await for response		
Prepared By: S Xu	Date: 05/09/2005	