


Government of the District of Columbia



RECEIVED  
REGION 1

'05 MAY -2 P1:33



**District Department of Transportation**  
Infrastructure Project Management Administration  
Quality Assurance/Quality Control Division

April 28, 2005

**Mr. Samuel Collins**  
Region 1  
U.S. Nuclear Regulator  
Commission  
475 Allendale Road  
King of Prussia, PA 19406

Dear Mr. Collins:

Transferring Nuclear Gauges  
License No. 08-23557-01

03028905

This letter is to confirm that seven of our nuclear gauges have been transferred to Introk Tek Inc. at 5908 Triangle Drive, Raleigh, NC 27617 permanently . ( see attachment) Leak tests for these seven gauges are also attached.

Please contact me at (202) 671-2422 if you have any questions.

Sincerely,

A handwritten signature in cursive script that reads "Lawrence Chung".  
Lawrence Chung  
Radiation Safety Officer

Attachments

Cc: Kathy Modes

INOC/REGI MATERIALS-002

# Memo

**To:** Mr. Lawrence Chung, Washington DC DOT, QC division

**From:** InstroTek, Inc.

**Date:** 4/27/2005

**Re:** Transfer of gauges into InstroTek Inventory

We have received and inspected the following gauges and have now permanently transferred these gauges into our inventory at 5908 Triangle Drive, Raleigh, NC 27617, USA.

Gauge Manufacturer	Model	S/N
Troxler	3241 B	497
Troxler	3241 C	1167
Troxler	3241 C	1895
Troxler	4640	459
Troxler	3440	19294
Troxler	3411 B	16053
Troxler	3411 B	9103

If we can be of any assistance, please do not hesitate to contact us.

Sincerely,



Ali Regimand  
President

InstroTek, Inc.  
5908 Triangle Dr  
Raleigh, NC 27617  
(919) 875-8371 Fax (919) 875-8328

Date: 4/21/2005  
Test Number: 1

LARRY CHUNG  
DISTRICT OF COLUMBIA, DOT  
2000 14TH ST N.W.  
WASHINGTON D.C., DC 20009

Phone: (202 ) 671 -2422  
FAX: (202 ) 671 -0646

## LEAK TEST CERTIFICATE

NC Materials License # 092-1073-1

This certifies that leak test analysis was conducted on the sample with the following information.  
The results shown below accurately represent the level of removable contamination.

Gauge model: 3241-B

Gauge S/N: 497

Leak Test Date: 3/28/2005

Source	Reading in microCuries
60-7758	0.000000
N/A	0.000183

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by: Harry Yessitt

Date: 4/21/05

InstroTek, Inc.  
5908 Triangle Dr  
Raleigh, NC 27617  
(919) 875-8371 Fax (919) 875-8328

Date: 4/21/2005  
Test Number: 1

LARRY CHUNG  
DISTRICT OF COLUMBIA, DOT  
2000 14TH ST N.W.  
WASHINGTON D.C., DC 20009

Phone: (202 ) 671 -2422  
FAX: (202 ) 671 -0646

## LEAK TEST CERTIFICATE

NC Materials License # 092-1073-1

This certifies that leak test analysis was conducted on the sample with the following information.  
The results shown below accurately represent the level of removable contamination.

Gauge model: 3411-B

Gauge S/N: 9103

Leak Test Date: 3/28/2005

Source	Reading in microCuries
40-6415	0.000000
N/A	0.000201

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by: Harry Yacobi

Date: 4/21/05

InstroTek, Inc.  
5908 Triangle Dr  
Raleigh, NC 27617  
(919) 875-8371 Fax (919) 875-8328

Date: 4/21/2005  
Test Number: 1

LARRY CHUNG  
DISTRICT OF COLUMBIA, DOT  
2000 14TH ST N.W.  
WASHINGTON D.C., DC 20009

Phone: (202) 671-2422  
FAX: (202) 671-0646

## LEAK TEST CERTIFICATE

NC Materials License # 092-1073-1

This certifies that leak test analysis was conducted on the sample with the following information.  
The results shown below accurately represent the level of removable contamination.

Gauge model: 3411-B

Gauge S/N: 16053

Leak Test Date: 3/28/2005

Source	Reading in microCuries
50-5037	0.000000
N/A	0.000188

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by: Harry Yount

Date: 4/21/05

InstroTek, Inc.  
5908 Triangle Dr  
Raleigh, NC 27617  
(919) 875-8371 Fax (919) 875-8328

Date: 4/21/2005  
Test Number: 1

LARRY CHUNG  
DISTRICT OF COLUMBIA, DOT  
2000 14TH ST N.W.  
WASHINGTON D.C., DC 20009

Phone: (202 )671 -2422  
FAX: (202 )671 -0646

## LEAK TEST CERTIFICATE

NC Materials License # 092-1073-1

This certifies that leak test analysis was conducted on the sample with the following information.  
The results shown below accurately represent the level of removable contamination.

Gauge model: 3440

Gauge S/N: 19294

Leak Test Date: 3/28/2005

Source	Reading in microCuries
50-8952	0.000002
N/A	0.000176

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by: Harry Yeast

Date: 4/21/05

InstroTek, Inc.  
5908 Triangle Dr  
Raleigh, NC 27617  
(919) 875-8371 Fax (919) 875-8328

Date: 4/21/2005  
Test Number: 1

LARRY CHUNG  
DISTRICT OF COLUMBIA, DOT  
2000 14TH ST N.W.  
WASHINGTON D.C., DC 20009

Phone: (202 ) 671 -2422  
FAX: (202 ) 671 -0646

## LEAK TEST CERTIFICATE

NC Materials License # 092-1073-1

This certifies that leak test analysis was conducted on the sample with the following information.  
The results shown below accurately represent the level of removable contamination.

Gauge model: 4640

Gauge S/N: 459

Leak Test Date: 3/28/2005

Source	Reading in microCuries
4808	0.000000
N/A	0.000178

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by: Harry Yacov

Date: 4/21/05

4-26-05; 8:35PM  
InstroTek, Inc.  
5908 Triangle Dr  
Raleigh, NC 27617  
(919) 875-8371 Fax (919) 875-8328

9198758320  
P.1  
Date: 4/26/2005  
Test Number: 1

LARRY CHUNG  
DISTRICT OF COLUMBIA, DOT  
2000 14TH ST N.W.  
WASHINGTON D.C., DC 20009

Phone: (202 ) 671 -2422  
FAX: (202 ) 671 -0646

## LEAK TEST CERTIFICATE

NC Materials License # 092-1073-1

This certifies that leak test analysis was conducted on the sample with the following information.  
The results shown below accurately represent the level of removable contamination.


Gauge model: Troxler 3241-C Gauge S/N: 1895

Leak Test Date: 4/25/2005

Source	Reading in microCuries
559180	0.000000
559180	0.000248

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by:



Date:

4/26/05



InstroTek, Inc.  
5908 Triangle Dr  
Raleigh, NC 27617  
(919) 875-8371 Fax (919) 875-8328

Date: 4/26/2005  
Test Number: 1

LARRY CHUNG  
DISTRICT OF COLUMBIA, DOT  
2000 14TH ST N.W.  
WASHINGTON D.C., DC 20009

Phone: (202 ) 671 -2422  
FAX: (202 ) 671 -0646

## LEAK TEST CERTIFICATE

NC Materials License # 092-1073-1

This certifies that leak test analysis was conducted on the sample with the following information.  
The results shown below accurately represent the level of removable contamination.


Gauge model: Troxler 3241-C Gauge S/N: 1167

Leak Test Date: 4/25/2005

<u>Source</u>	<u>Reading in microCuries</u>
558395	0.000002
558395	0.000218

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by:



Date:

4/26/05