

U.S. NUCLEAR REGULATORY COMMISSION		Conversation Date: 05/09/05
TELEPHONE CONVERSATION RECORD		Time: 1300 hours
Mail Control No.: 136679	License No.: 45-08482-01	Docket No.: 03003327
Licensee/Applicant Participant(s): Peter Paras, RSO,	Organization: Northern Virginia Radiology	Telephone No.: 301.670.1095
Person(s) Calling: S Xu, RI		
Subject: license renewal		
<p>Summary:</p> <p>Following issues have been discussed with Dr. Paras:</p> <ol style="list-style-type: none"> 1. Confirm their license name. The name on the application and on the current license are different. 2. Asked him to provide us the list of their survey instruments. 3. Asked him to provide us with their security measures of their seeds storage room. 4. Confirm that their radiation instruments will be calibrated by a person qualified to perform survey meter calibrations. 5. Informed him that we don't review their treatment protocol. <p>5/16/2005</p> <ol style="list-style-type: none"> 1. Informed him that Pd-103 is state regulated material. 2. Confirmed that they want 2 Ci for I-125. 3. On their renewal application listed physicists. Informed them that Part 35 has no provision to appoint an authorized medical physicist for 35.400 manual brachytherapy. 4. Suggested they obtain a thin sodium iodide crystal detector for better I-125 detection efficiency, but he said they are not getting one at this point. 		
Action Required/Taken: place in ADAMS		
Prepared By: S Xu	Date: 05/09/2005 and 5/16/2005	