



UNITED STATES  
**NUCLEAR REGULATORY COMMISSION**  
REGION I  
475 ALLENDALE ROAD  
KING OF PRUSSIA, PENNSYLVANIA 19406-1415

June 7, 2005

Docket No. 03036890  
Control No. 136601

License No. 45-31026-01

Michael D. Schorr  
Environmental, Health and Safety Manager  
U.S. Filter Recovery Services  
5800 Farrington Avenue  
Alexandria, VA 22304

SUBJECT: U.S. FILTER RECOVERY SERVICES, VOIDANCE OF APPLICATION FOR  
NEW LICENSE, CONTROL NO. 136601

Dear Mr. Schorr:

This concerns the subject application for an NRC license. The gas chromatograph that you desired contains a generally licensed source which does not require a specific license. We have, therefore, voided your application. This action is taken without prejudice to the resubmission of your request.

Sincerely,

***Original signed by Todd J. Jackson, CHP***

Todd J. Jackson, CHP  
Senior Health Physicist  
Commercial and R&D Branch  
Division of Nuclear Materials Safety

M. Schorr  
U.S. Filter Recovery Services

2

DOCUMENT NAME: E:\Filenet\ML051600014.wpd

To receive a copy of this document, indicate in the box: "C" = Copy w/o attach/encl "E" = Copy w/ attach/encl "N" = No copy

OFFICE	DNMS/RI	N	DNMS/RI	N	DNMS/RI			
NAME	DLawyer/DRL1		TJackson/TJJ					
DATE	06/07/2005		6/7/2005					

OFFICIAL RECORD COPY