

U.S. NUCLEAR REGULATORY COMMISSION		Conversation Date: 6-6-05	
TELEPHONE CONVERSATION RECORD		Time:	
Mail Control No.:	License No.:	Docket No.:	
136876	45-09001-01	030-03330	
Licensee/Applicant Participant(s):	Organization:	Telephone No.:	
Harold Prussia	Riverside Regional Medical Center	757-594-2757	
Person(s) Calling: T. Weidner			
Subject: License amendment			
Summary: Riverside Regional requested that they be able to calibrate the survey instruments of the hospitals that are part of the Riverside Health System. According to P. Lanzisera as long as Riverside Regional is doing this service free of charge they will not need a service license. If the facilities would be charged for the calibrations, then Riverside Regional would need to have a service license.			
Action Required/Taken: Complete the amendment			
Prepared By: T. Weidner		Date: 6-7-05	