

U.S. NUCLEAR REGULATORY COMMISSION		Conversation Date: 06/8+06/10/05	
<b>TELEPHONE CONVERSATION RECORD</b>		Time: various	
Mail Control No.:	License No.:	Docket No.:	
137096	37-00485-04	03002960	
Licensee/Applicant Participant(s):	Organization:	Telephone No.:	
Rene Smith, Ph.D., RSO	The Reading Hospital & Medical Center	610-988-8144	
Person(s) Calling: SGabriel, RI			
Subject: Expedited amendment request			
<p>Summary:</p> <p>6/8: I called Rene to ask how quickly the amendment is needed. He said plan is to move Nuclear Cardiology and HDR next week. I told him we need additional information for the HDR move (shielding calcs and other facility safety items). He said HDR move might be able to be delayed. He asked if it is permissible to move an HDR unit within contiguous hospital buildings with the source contained in the unit. After consulting with the branch chief, I told him this is permissible if a post-move survey is performed to confirm no change in shielding and if a full calibration is then performed. Rene will let me know if licensee wishes to withdraw the request to move the HDR and resubmit this later.</p> <p>6/10: Rene said he received the e-mailed HDR questions and is working on them. We discussed whether the move of Nuclear Cardiology could be considered a notification, not requiring a license amendment prior to the move. Rene confirmed that no material other than 35.100 and 35.200 is used in Nuclear Cardiology. I told him we will treat the request to move Nuclear Cardiology as a notification so the move can proceed. Licensee will submit closeout surveys after the move. The remainder of this amendment request will be addressed after the HDR questions are answered.</p>			
Action Required/Taken:		E-mail HDR questions, await response to questions, treat move of Nuclear Cardiology as a notification	
Prepared By: SGabriel		Date: 06/10/05	