

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED: InstroTek - N.E. 78 South Main Street Newtown, CT 06470 REPORT Nos 2005-001		2. NRC/REGIONAL OFFICE U.S. Nuclear Regulatory Commission Region I, 475 Allendale Road King of Prussia, Pennsylvania 19406-1415
3. DOCKET NUMBER(S) 030-35611	4. LICENSE NUMBER(S) 06-30610-01	5. DATE(S) OF INSPECTION 8. 30+31, 2005

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- ☐ 1. Based on the inspection findings, no violations were identified.
- ☐ 2. Previous violation(s) closed.
- ☐ 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.
- ☐ Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):

- ☒ 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.

10 CFR 20.1101(c) requires the licensee to at least annually review their entire radiation protection program. The licensee has not renewed their program in the last 3 years. The licensee will complete a review of their program in 30 days. Documentation will be maintained. License Condition 17 requires a lock on the gauge or its container when in storage. Five gauges were not locked, nor in their container when in their storage room. The licensee will purchase locks & lock all gauges when in storage.

Licensee's Statement of Corrective Actions for Item 4, above.

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE	Philip C Palilla	Philip C Palilla	8/31/05
NRC INSPECTOR	Kathy Dolce Modes	K.D. Modes	8.31.05

NRC-002