

| | | |
|--|----------------------------|------------------------|
| U.S. NUCLEAR REGULATORY COMMISSION | | Date: 9/2/2005 |
| TELEPHONE CONVERSATION RECORD | | Time: 1:30 pm |
| Mail Control 137530 or Report No(s). | License No(s). 37-00554-03 | Docket No(s). 03002961 |
| <p>Name of Licensee: Sacred Heart Hospital</p> <p>Name of Participant(s): Frank Altomare, RSO</p> <p>Telephone No. 610-776-4684</p> <p>Subject: Additional information re: proposed RSO <small>(NOTE: This will be used as the Documents Title in ADAMS)</small></p> | | |
| <p>Summary: I spoke with Dr. Altomare regarding the following information needed to continue review of Ralph Natale's, the proposed RSO, qualifications:</p> <ol style="list-style-type: none"> 1. Need confirmation (written statement) that Mr. Natale has one year of full-time experience, within the last 7 years, performing the duties of a radiation safety technologist at a medical institution under an RSO's supervision. 2. Need to commit to updating the hospital's emergency procedures to include appropriate information for new RSO. 3. Because Mr. Natale's training and experience relate to diagnostic and therapeutic nuclear medicine, describe the training received or to be provided for Mr. Natale in the radiation safety aspects of manual brachytherapy and HDR. <p>Dr. Altomare stated he would provide the information by facsimile as soon as possible.</p> | | |
| <p>Action Required: Wait for licensee response.</p> | | |
| <p>Document Availability: <input checked="" type="checkbox"/> Publicly Available <input type="checkbox"/> Non-Publicly Available</p> <p><input checked="" type="checkbox"/> Non-Sensitive <input type="checkbox"/> Non-Sensitive Copyright <input type="checkbox"/> Sensitive <input type="checkbox"/> Sensitive Copyright</p> <p><input type="checkbox"/> Immediate Release <input checked="" type="checkbox"/> Normal Release <input type="checkbox"/> Delay Release Date</p> <p>Prepared & SISF Review Completed By: / RA / Donna M. Janda Date: 9/2/2005</p> | | |