

U.S. NUCLEAR REGULATORY COMMISSION		Date: 9/6/2005
TELEPHONE CONVERSATION RECORD		Time: 13:30
Mail Control 137506 or Report No(s).	License No(s). 07-31076-01	Docket No(s). 03036998
<p>Name of Licensee: David S. Grubbs, M.D.</p> <p>Name of Participant(s): Mark Liddington, Medical Consultant Steven Courtemanche, Health Physicist/NRC</p> <p>Telephone No. 717-291-9813 610-337-5075</p> <p>Subject: Telephone Deficiency (NOTE: This will be used as the Documents Title in ADAMS)</p>		
<p>Summary: Mr. Liddington called to determine the status of the licensing action for his client. I informed him that I had reviewed the action last week and that I had three questions that needed to be addressed in a letter signed by the licensee. (1) I needed the phone number, facsimile number and, if possible, the e-mail address for the licensee. (2) Confirmation that the facilities in which the licensee will be operating are not occupied by another NRC licensee. (3) The application indicates that the licensee possesses only one survey instrument for the performance of radiation dose rate surveys. The licensee must commit to either obtaining a loaner instrument when their survey instrument is off-site or obtaining a second instrument. Mr. Liddington stated that he would see to it that a signed letter addressing all of the above would be sent in to the NRC.</p>		
<p>Action Required: Place in ADAMS</p>		
<p>Document Availability: <input checked="" type="checkbox"/> Publicly Available <input type="checkbox"/> Non-Publicly Available</p> <p><input checked="" type="checkbox"/> Non-Sensitive <input type="checkbox"/> Non-Sensitive Copyright <input type="checkbox"/> Sensitive <input type="checkbox"/> Sensitive Copyright</p> <p><input type="checkbox"/> Immediate Release <input checked="" type="checkbox"/> Normal Release <input type="checkbox"/> Delay Release Date</p> <p>Prepared & SISP Review Completed By: / RA / Steven Courtemanche Date: 9/6/2005</p>		