

U.S. NUCLEAR REGULATORY COMMISSION		Date: September 20, 2005
TELEPHONE CONVERSATION RECORD		Time: 9:00
Mail Control	136778	License No(s). 45-18488-01
or Report No(s).		Docket No(s). 030-15183
<p>Name of Licensee: Riverside Tappahannock Hospital</p> <p>Name of Participant(s): Nancy Hillyer, Director of Radiology Randolph C. Ragland, Jr., NRC (610-337-5083)</p> <p>Telephone No. 804-443-6044</p> <p>Subject: License No. 45-18488-01 License Renewal: Request for Additional Information (NOTE: This will be used as the Documents Title in ADAMS)</p>		
<p>Summary: I requested the following information for Riverside Tappahannock Hospital's NRC License Renewal:</p> <ul style="list-style-type: none"> 1) Dr. Randall A. Sikes, M.D. did not appear on the renewal application. Please state whether you would like Dr. Sikes included or deleted in the renewed NRC License. 2) Facility Diagram <ul style="list-style-type: none"> A) Please include a diagram that shows where the Nuclear Medicine Suite and Cardiac Stress Lab are located with reference to the rest of the hospital. B) Please include details of the Nuclear Medicine Hot lab including waste storage area, any shielding used, location of dose calibrator, etc. C) Please confirm that there are no occupied locations above or below the Hot Lab (e.g., Roof above and ground below) D) Please state which areas are designated as restricted areas (i.e., controlled to minimize radiation exposure). 		
<p>Action Required: Ms. Hillyer, stated that she would assemble the information and fax it to the NRC at 610-337-5269</p>		
<p>Document Availability: <input checked="" type="checkbox"/> Publicly Available <input type="checkbox"/> Non-Publicly Available</p> <p><input checked="" type="checkbox"/> Non-Sensitive <input type="checkbox"/> Non-Sensitive Copyright <input type="checkbox"/> Sensitive <input type="checkbox"/> Sensitive Copyright</p> <p><input type="checkbox"/> Immediate Release <input checked="" type="checkbox"/> Normal Release <input type="checkbox"/> Delay Release Date</p> <p>Prepared & SISP Review Completed By: / RA / R. Ragland Date: 9/20/05</p>		