U.S. NUCLEAR REGULATORY COMMISSION

TELEPHONE CONVERSATION RECORD

Date: 09/21/2005

Time: 1:05 PM

Mail Control or Report No(s). 137539 License No(s). 29-30629-01 Docket No(s). 03035672

Name of Licensee:	The Heart Center of the Oranges
Name of Participant(s):	Elaine Rovazzi, M.S., D.A.B.R.
Telephone No.	973-322-5118
Subject: (NOTE: This will be used as the Documents Title in ADAMS)	RAI, Leak Test Results
Summary:	Leak test results are needed for all sealed sources which require leak testing that were transferred to the new facility. The test result must be within 6 months of the transfer.
Action Required:	Ms. Rovazzi said that she send them into Region I.

Document Availability: X Publicly Available ___ Non-Publicly Available

X Non-Sensitive ___ Non-Sensitive Copyright ___ Sensitive ___ Sensitive Copyright

___ Immediate Release X Normal Release ___ Delay Release Date

Prepared & SISP Review Completed By: ***/ RA /*** Dennis Lawyer Date: 09/21/2005