

U.S. NUCLEAR REGULATORY COMMISSION		Date: 9/27/2005
TELEPHONE CONVERSATION RECORD		Time: 10:45
Mail Control 137647 or Report No(s).	License No(s). 06-30693-01	Docket No(s). 030-35868
<p>Name of Licensee: Protometrix - An Invitrogen Company</p> <p>Name of Participant(s): Gregory Waterbury, Facilities Manager and Safety Officer Steven Courtemanche, Health Physicist NRC</p> <p>Telephone No. 203-848-1104 610-337-5075</p> <p>Subject: Telephone Deficiency (NOTE: This will be used as the Documents Title in ADAMS)</p>		
<p>Summary: I spoke with Mr. Waterbury about their recent license amendment request and the additional information required. (1) I asked if they only possessed GM survey instruments. He stated that they do. I informed him that a GM survey instrument would not be the best instrument for measuring Iodine-125 contamination. I would ask them to confirm that they would obtain an instrument appropriate for the measurement of Iodine-125. (2) Personnel need to be trained in the hazards peculiar to Iodine-125 since they are different from what is currently authorized by the licensed program. (3) Recommend that the licensee perform a base-line thyroid count on individuals in case of a future contamination incident. Mr. Waterbury indicated that he would look into the requests and would like a copy of the deficiencies in writing.</p>		
<p>Action Required: Place into ADAMS</p>		
<p>Document Availability: <input checked="" type="checkbox"/> Publicly Available <input type="checkbox"/> Non-Publicly Available</p> <p><input checked="" type="checkbox"/> Non-Sensitive <input type="checkbox"/> Non-Sensitive Copyright <input type="checkbox"/> Sensitive <input type="checkbox"/> Sensitive Copyright</p> <p><input type="checkbox"/> Immediate Release <input checked="" type="checkbox"/> Normal Release <input type="checkbox"/> Delay Release Date</p> <p>Prepared & SISP Review Completed By: / RA / Steven R. Courtemanche Date: 9/27/2005</p>		