

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:
Cardiology Center of North Jersey
1030 Clifton Avenue
Clifton, NJ 07013

2. NRC/REGIONAL OFFICE

**U.S. Nuclear Regulatory Commission
Region I, 475 Allendale Road
King of Prussia, Pennsylvania 19406-1415**

REPORT Nos 2005-001

3. DOCKET NUMBER(S)
030-35176

4. LICENSE NUMBER(S)
29-30528-01

5. DATE(S) OF INSPECTION
September 21, 2005

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- ☐ 1. Based on the inspection findings, no violations were identified.
- ☐ 2. Previous violation(s) closed.
- ☒ 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.

☒ Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):

10 CFR 30.34(c) requires, in part, that each licensee confine his possession and use of byproduct materials to the locations and purposes authorized by the license.

Condition No. 15 of License No. 29-30528-01 requires that the licensee conduct its program in accordance with the statements, representations, and procedures contained in the application dated July 22, 1999, and the letters dated September 11, 1999, and February 1, 2000.

(Continued on Part 2)

- ☐ 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.

Licensee's Statement of Corrective Actions for Item 4, above.

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE	Elliot Brown	/RA/	10/3/05
NRC INSPECTOR	Joseph Nick	/RA/	10/3/05