

U.S. NUCLEAR REGULATORY COMMISSION		Date: 9/26/2005
TELEPHONE CONVERSATION RECORD		Time: 2:50 pm
Mail Control or Report No(s). 137580	License No(s). 37-07161-01	Docket No(s). 03003085
<p>Name of Licensee: York Hospital</p> <p>Name of Participant(s): Jean Gresick-Schugsta, RSO</p> <p>Telephone No. 717-851-5166</p> <p>Subject: Questions on former nuclear cardiology area close out surveys (NOTE: This will be used as the Documents Title in ADAMS)</p>		
<p>Summary:</p> <ol style="list-style-type: none"> 1. The survey forms submitted stated that the calibration date for the Eberline E-520 survey meter was 6/8/99. The surveys were performed on 8/5 & 8/8/05. Ms. Gresick-Schugsta stated that the forms were incorrect. The correct calibration date for the survey instrument used was 8/9/04. She stated that she would submit the calibration certificate for the survey meter in question. 2. Confirm that only radioactive material with half-lives less than 120 days were used/stored in the location to be released. 3. Confirm that all radioactive material has been removed from the location to be released. <p>Ms. Gresick-Schugsta stated that she will fax a written response to these questions along with the calibration certificate.</p>		
<p>Action Required: Wait for licensee information.</p>		
<p>Document Availability: <input checked="" type="checkbox"/> Publicly Available <input type="checkbox"/> Non-Publicly Available</p> <p><input checked="" type="checkbox"/> Non-Sensitive <input type="checkbox"/> Non-Sensitive Copyright <input type="checkbox"/> Sensitive <input type="checkbox"/> Sensitive Copyright</p> <p><input type="checkbox"/> Immediate Release <input checked="" type="checkbox"/> Normal Release <input type="checkbox"/> Delay Release Date</p> <p>Prepared & SISP Review Completed By: / RA / Donna M. Janda Date: 9/26/2005</p>		