



RECEIVED
SECTION 1

J-5

2005 OCT -6 AM 9:47

**Respond to: Paul J. Early, Corporate RSO
P.O. Box 340
Bemus Point, NY 14712**

**PH: 716.386.3860
FX: 716.386.5118
VM: 858.537.2137**

October 4, 2005

VIA FAX (610-337-5393)

U.S. Nuclear Regulatory Commission, Region I
475 Allendale Road
King of Prussia PA 19406-1415

Re: Amendment for License No. 31-30666-01 03035802

To Whom It May Concern:

Please amend our radioactive material license to **DELETE** the following **BASE LOCATION**:

810 Hospital Drive
Blacksburg, VA 24060

(Close-out surveys are attached.)

Thank you for your immediate attention to this matter.

Sincerely,

A handwritten signature in cursive script that reads "Paul J. Early".

Paul J. Early, DABSNM, DABPM

137643
NMSS/RGNI MATERIALS-002



Close out wipe
test of 810
Hospital Dr.
Blacksburg, VA 24060

Weekly Wipe Test Report

Medical Associates of SW VA
810 Hospital Drive
Blacksburg, VA 24060

Date: 9-26-05

Equipment: Ludlum Survey Meter & NaI Probe

Cs-137 Standard Count: 30K CPM

Probe Efficiency: 1657

Probe BKG Count: .2K CPM 131 dpm

<u>Wipe Test Location</u>	<u>Result (CPM)</u>	<u>Result (DPM)</u>	<u>MR/hr</u>
1. Radiopharmacy - Counter Top	<u>.1K</u>	<u>65</u>	<u>.02</u>
2. Radiopharmacy - Floor	<u>.1K</u>	<u>65</u>	<u>.02</u>
3. Radiopharmacy - Decay Storage Bin	<u>.15K</u>	<u>98</u>	<u>.02</u>
4. SPECT Chair Arm Rest Floor	<u>.05K</u>	<u>32</u>	<u>.02</u>
5. Treadmill Surface	<u>.05K</u>	<u>32</u>	<u>.02</u>
6. Injection Area	<u>.1K</u>	<u>65</u>	<u>.02</u>
7. Injection Area	<u>.05K</u>	<u>32</u>	<u>.02</u>

* Action Level = 2000 DPM/100 cm²

Action Taken? ☒ No

Yes: Explain: _____

Technologist: T. L. L. L. L. L.

810 Hospital Dr. Blacksburg, VA
24060

Sep. 29. 2005 11:34AM MEDICAL ASSOC. OF SW VA

No. 1168 P. 3

