

U.S. NUCLEAR REGULATORY COMMISSION		Date: 9/29/2005															
TELEPHONE CONVERSATION RECORD		Time: 10:45															
Mail Control or Report No(s).	137679	License No(s). 29-23401-01															
		Docket No(s). 030-21026															
<table style="width: 100%; border: none;"> <tr> <td style="width: 30%; padding-bottom: 10px;">Name of Licensee:</td> <td style="padding-bottom: 10px;">Rider University</td> </tr> <tr> <td style="padding-bottom: 10px;">Name of Participant(s):</td> <td style="padding-bottom: 10px;">Kelly Bidle, Ph.D., Radiation Safety Officer Steven Courtemanche, Health Physicist NRC</td> </tr> <tr> <td style="padding-bottom: 10px;">Telephone No.</td> <td style="padding-bottom: 10px;">609-895-5418 610-337-5075</td> </tr> <tr> <td style="padding-bottom: 10px;">Subject: (NOTE: This will be used as the Documents Title in ADAMS)</td> <td style="padding-bottom: 10px;">Additional Information for License Amendment</td> </tr> </table>			Name of Licensee:	Rider University	Name of Participant(s):	Kelly Bidle, Ph.D., Radiation Safety Officer Steven Courtemanche, Health Physicist NRC	Telephone No.	609-895-5418 610-337-5075	Subject: (NOTE: This will be used as the Documents Title in ADAMS)	Additional Information for License Amendment							
Name of Licensee:	Rider University																
Name of Participant(s):	Kelly Bidle, Ph.D., Radiation Safety Officer Steven Courtemanche, Health Physicist NRC																
Telephone No.	609-895-5418 610-337-5075																
Subject: (NOTE: This will be used as the Documents Title in ADAMS)	Additional Information for License Amendment																
<table style="width: 100%; border: none;"> <tr> <td style="width: 30%; vertical-align: top; padding-bottom: 10px;">Summary:</td> <td style="padding-bottom: 10px;"> <p>I spoke with Dr. Bidle. I informed her that the gas chromatograph listed on the Certificate of Disposition was for a Varian Gas chromatograph and not a Perkin Elmer. Dr. Bidle went over the steps she has taken so far to locate the missing ECD. (1) Edwin Roth Ph.D. is dead. (2) She has spoken to individuals who have worked with Dr. Roth. They had no memory of the ECD. (3) The first floor of the Science Building has been renovated during the summer of 2004. All construction waste and obsolete equipment was transported to a landfill. (4) A review of records in Bills Receivable did not indicate that NRD was used, per Perkin Elmer's instructions, for the disposal of the ECD. I requested that Dr. Bidle speak to the former RSO and determine whether the licensee had shipped any radioactive waste containing tritium and Carbon-14 by a vender since 1990. She should review the waste manifests and determine whether the ECD had been placed in the shipment. Once the licensee has finished the review, they should draft a letter and document their actions so that a determination could be made on whether to remove the item from the license.</p> </td> </tr> </table>			Summary:	<p>I spoke with Dr. Bidle. I informed her that the gas chromatograph listed on the Certificate of Disposition was for a Varian Gas chromatograph and not a Perkin Elmer. Dr. Bidle went over the steps she has taken so far to locate the missing ECD. (1) Edwin Roth Ph.D. is dead. (2) She has spoken to individuals who have worked with Dr. Roth. They had no memory of the ECD. (3) The first floor of the Science Building has been renovated during the summer of 2004. All construction waste and obsolete equipment was transported to a landfill. (4) A review of records in Bills Receivable did not indicate that NRD was used, per Perkin Elmer's instructions, for the disposal of the ECD. I requested that Dr. Bidle speak to the former RSO and determine whether the licensee had shipped any radioactive waste containing tritium and Carbon-14 by a vender since 1990. She should review the waste manifests and determine whether the ECD had been placed in the shipment. Once the licensee has finished the review, they should draft a letter and document their actions so that a determination could be made on whether to remove the item from the license.</p>													
Summary:	<p>I spoke with Dr. Bidle. I informed her that the gas chromatograph listed on the Certificate of Disposition was for a Varian Gas chromatograph and not a Perkin Elmer. Dr. Bidle went over the steps she has taken so far to locate the missing ECD. (1) Edwin Roth Ph.D. is dead. (2) She has spoken to individuals who have worked with Dr. Roth. They had no memory of the ECD. (3) The first floor of the Science Building has been renovated during the summer of 2004. All construction waste and obsolete equipment was transported to a landfill. (4) A review of records in Bills Receivable did not indicate that NRD was used, per Perkin Elmer's instructions, for the disposal of the ECD. I requested that Dr. Bidle speak to the former RSO and determine whether the licensee had shipped any radioactive waste containing tritium and Carbon-14 by a vender since 1990. She should review the waste manifests and determine whether the ECD had been placed in the shipment. Once the licensee has finished the review, they should draft a letter and document their actions so that a determination could be made on whether to remove the item from the license.</p>																
<table style="width: 100%; border: none;"> <tr> <td style="width: 30%; vertical-align: top;">Action Required:</td> <td>Place in ADAMS</td> </tr> </table>			Action Required:	Place in ADAMS													
Action Required:	Place in ADAMS																
<table style="width: 100%; border: none;"> <tr> <td style="width: 30%;">Document Availability:</td> <td style="text-align: center;"><input checked="" type="checkbox"/> Publicly Available</td> <td style="text-align: center;"><input type="checkbox"/> Non-Publicly Available</td> </tr> <tr> <td><input checked="" type="checkbox"/> Non-Sensitive</td> <td style="text-align: center;"><input type="checkbox"/> Non-Sensitive Copyright</td> <td style="text-align: center;"><input type="checkbox"/> Sensitive</td> </tr> <tr> <td style="text-align: center;"><input type="checkbox"/> Immediate Release</td> <td style="text-align: center;"><input checked="" type="checkbox"/> Normal Release</td> <td style="text-align: center;"><input type="checkbox"/> Delay Release Date</td> </tr> <tr> <td colspan="3">Prepared & SISF Review Completed By: / RA / Steven R. Courtemanche</td> </tr> <tr> <td colspan="3" style="text-align: right;">Date: 9/29/2005</td> </tr> </table>			Document Availability:	<input checked="" type="checkbox"/> Publicly Available	<input type="checkbox"/> Non-Publicly Available	<input checked="" type="checkbox"/> Non-Sensitive	<input type="checkbox"/> Non-Sensitive Copyright	<input type="checkbox"/> Sensitive	<input type="checkbox"/> Immediate Release	<input checked="" type="checkbox"/> Normal Release	<input type="checkbox"/> Delay Release Date	Prepared & SISF Review Completed By: / RA / Steven R. Courtemanche			Date: 9/29/2005		
Document Availability:	<input checked="" type="checkbox"/> Publicly Available	<input type="checkbox"/> Non-Publicly Available															
<input checked="" type="checkbox"/> Non-Sensitive	<input type="checkbox"/> Non-Sensitive Copyright	<input type="checkbox"/> Sensitive															
<input type="checkbox"/> Immediate Release	<input checked="" type="checkbox"/> Normal Release	<input type="checkbox"/> Delay Release Date															
Prepared & SISF Review Completed By: / RA / Steven R. Courtemanche																	
Date: 9/29/2005																	