

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:

*Center for Diagnostic Imaging
Indianapolis, IN*

2. NRC/REGIONAL OFFICE

U.S. Nuclear Regulatory Commission
Region III
2443 Warrenville Road
Suite 210
Lisle, Illinois 60532-4351

REPORT *2005-001*

3. DOCKET NUMBER(S)

030-35142

4. LICENSEE NUMBER(S)

13-32194-01

5. DATE(S) OF INSPECTION

Nov. 2, 2005

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection

☒ 1. Based on the inspection findings, no violations were identified.

☐ 2. Previous violation(s) closed.

☐ 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.

_____ Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):

☐ 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.

(Violations and Corrective Actions)

Licensee's Statement of Corrective Actions for Item 4, above.

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Deborah A. Piskura	<i>D. Piskura</i>	<i>11/2/05</i>

Docket File Information
SAFETY INSPECTION REPORT
AND COMPLIANCE INSPECTION

1. LICENSEE Center for Diagnostic Imaging REPORT NUMBER(S) 2005-001		2. NRC/REGIONAL OFFICE Region III 2443 Warrenville Road, Suite 210 Lisle, IL 60532	
3. DOCKET NUMBER(S) 030-35142	4. LICENSE NUMBER(S) 13-32194-01	5. DATE(S) OF INSPECTION Nov. 2, 2005	
6. INSPECTION PROCEDURES 87130 & 87131	7. INSPECTION FOCUS AREAS 03.01, 03.02, 03.03, 03.04, 03.05, 03.06, 03.07, and 03.08		
SUPPLEMENTAL INSPECTION INFORMATION			
1. PROGRAM CODE(S) 02200	2. PRIORITY G 3	3. LICENSEE CONTACT Andrea Carpenter, CNMT, RSO	4. TELEPHONE NUMBER 317.846.0717
<input checked="" type="checkbox"/> Main Office Inspection		Next Inspection Date: 11/2008	
<input type="checkbox"/> Field			
<input type="checkbox"/> Temporary Job Site			

PROGRAM SCOPE

This licensee was a large multi-specialty diagnostic office and authorized to use licensed material permitted by Sections 35.100, 35.200 and 35.300. The nuclear medicine department was staffed with one full-time technologist and one temporary technologist who performed approximately 100 diagnostic procedures monthly. The licensee operated a very active I-131 therapy program. Typically, in a year, the clinic administered 10+ therapies for thyroid carcinoma, 5-10 dosages for hyperthyroidism and 5-10 whole body follow up scans. The licensee received unit doses and I-131 capsules from a licensed radiopharmacy. The clinic retained the services of a consulting physicist who audited the radiation safety program on a quarterly basis.

This inspection consisted of interviews with licensee personnel, a review of select records, tour of the nuclear medicine department, and independent measurements. The inspector observed licensee personnel demonstrate dose calibrator QA tests, and radiation surveys.