

TELEPHONE CONVERSATION

To: Mr. Scott Walker, RSO
From: Rachel Browder /RA/
Date: November 17, 2005
Subject: Clarifications on new license application for portable gauge.

I contacted Mr. Walker to discuss the following clarifications to Treadwell & Rollo, Inc. submittal for a new portable gauge license.

1. Provided Mr. Walker with the HOO telephone number (301-816-5100) so he could update the list of Emergency Telephone numbers in the Radiation Safety Program. I informed Mr. Walker that the HOO was staffed 24/7 and this is the number to call for any reporting requirements.
2. Mr. Walker clarified that Treadwell & Rollo, Inc. will take wipe tests with the kit provided by the manufacturer (CPN) and that CPN will perform the leak test analysis.
3. Mr. Walker clarified that Treadwell & Rollo, Inc. will only perform routine maintenance as allowed by the manufacturer and they will send the gauge to the manufacturer (CPN) for any non-routine maintenance and calibration.
4. I asked Mr. Walker about survey instruments and he stated that they did not have any survey instruments. They have had one gauge incident, in which a gauge had been run-over and they requested the manufacturer to respond to the incident in California. I requested that they determine how, as an NRC licensee, they would respond to an incident in NRC jurisdiction and provide that determination in writing to my attention. Essentially, they should either have a survey meter or have access to a survey meter. Mr. Walker stated that he would respond to my request in writing.

Docket No. 030-37096
License No. 04-29219-01
Control No. 470762