

## SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:

Salem Memorial District Hospital  
PO Box 774  
Salem, MO 65560

2. NRC/REGIONAL OFFICE

U.S. Nuclear Regulatory Commission  
Region III  
2443 Warrenville Road, Suite 210  
Lisle, IL 60532-4351

REPORT 2005/001

3. DOCKET NUMBER(S)

030-17320

4. LICENSEE NUMBER(S)

24-18891-01

5. DATE(S) OF INSPECTION

11/11/2005

## LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:



1. Based on the inspection findings, no violations were identified.



2. Previous violation(s) closed.



3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.

Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):




4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.

(Violations and Corrective Actions)

## Licensee's Statement of Corrective Actions for Item 4, above.

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Tony S. Go		11/10/05

NRC FORM 591M PART 3 (10-2003) 10 CFR 2.201		U.S. NUCLEAR REGULATORY COMMISSION	
<b><i>Docket File Information</i></b> <b>SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION</b>			
1. LICENSEE <b>Salem Memorial District Hospital</b> REPORT NUMBER(S) <b>03017320/2005001</b>		2. NRC/REGIONAL OFFICE U.S. Nuclear Regulatory Commission Region III	
3. DOCKET NUMBER(S) <b>030-17320</b>	4. LICENSE NUMBER(S) <b>24-18891-01</b>	5. DATE(S) OF INSPECTION <b>11/ 10 /2005</b>	
6. INSPECTION PROCEDURES <b>87130</b>	7. INSPECTION FOCUS AREAS <b>03.01 - 03.07</b>		
<b>SUPPLEMENTAL INSPECTION INFORMATION</b>			
1. PROGRAM <b>2220</b>	2. PRIORITY <b>3</b>	3. LICENSEE CONTACT <b>Qasim Bajwa, MD</b>	4. TELEPHONE NUMBER <b>573-729-6626</b>
<div style="display: flex; justify-content: space-between;"> <div> <input checked="checked" type="checkbox"/> Main Office Inspection  <input type="checkbox"/> Field Office  <input type="checkbox"/> Temporary Job Site         </div> <div>           Next Inspection Date: <b>11/08</b> </div> </div>			
<b>PROGRAM SCOPE</b>			

NRC FORM 591M PART 3 (10-2003)

### Program Scope

This licensee is a 50-bed hospital located in Salem, MO, about 110 miles Southwest of St. Louis. From this location, the licensee performed routine diagnostic nuclear medicine studies and also provided a mobile service to a client located in Houston, MO, for one day a week. However, most nuclear medicine studies were performed at the main hospital in Salem, MO, and Salem Hospital is the licensee current central location for the mobile service. At the main hospital site, the licensee performed approximately 30 diagnostic nuclear medicine procedures per month. The licensee ordered daily unit doses from a pharmacy.

### Performance Observations

During the inspection, the licensee's technologist demonstrated the following activities to the inspector: the daily radiation survey; and the weekly removable contamination surveys. During the performance-based inspection, the inspector observed that the individual used both whole body and extremity dosimeters during administration and during handling of licensed materials. The inspector observed the nuclear medicine technologist use, lab coats and gloves while handling radioactive materials. No violations of NRC requirements were identified during the inspection.

To the next inspector, because of the distance between hospitals, please call ahead before your next inspection.