

From: Donna Janda
To: Healthphysics@hpcva.com
Date: Tue, Nov 22, 2005 10:02 AM
Subject: Additional information for Maryview Medical Center license renewal

Licensee: Maryview Medical Center
License No. 45-10831-02
Docket No. 03003347
Mail Control No. 137316

ATTN: Roy F. Heltzel, Jr.

Dear Mr. Heltzel,

Your name was provided as the contact person for the Maryview Medical Center license renewal application. In order to continue our review of the renewal application, please provide the following information:

1. Your request did not list byproduct material permitted by 10 CFR 31.11 in the form of prepackaged kits, which is currently authorized on Maryview's NRC license. If this material is still required, please state so and provide a maximum possession limit for this material. If this material is no longer required, please confirm that no byproduct material permitted by 10 CFR 31.11 remains at Maryview. In addition, please provide the location(s) where the prepackaged kits were used/stored. Is licensed material still being used/stored in these areas? Did any of the kits contain licensed material with half-lives greater than 120 days? If so, which radioisotopes were used?
2. Your renewal application requests authorization for gadolinium-153 in sealed source form for storage only. This material is not currently authorized on your license. Gadolinium-153 has a half-life greater than 120 days and cannot be held for decay-in-storage. It must be properly disposed if you have no future plans for use of the licensed material. Please describe your future plans for use or disposal of this licensed material.
3. Please confirm that the authorized uses requested in Items 6.c. and 6.d. of the renewal application are for outpatient use only. For patients who cannot be released under 10 CFR 35.75, please provide the location(s) where sources are stored. Describe the rooms where patients will be housed and provide a description of shielding. Describe the adjacent areas and/or rooms, including those above and below the rooms where patients will be housed. For manual brachytherapy facilities, provide a description of your emergency response equipment.
4. Item 7 of your application and your current license lists Dennis Owens, M.S., and Leland Kirkland, Ph.D., as authorized users for "non-human uses associated with the performance of his radiation safety duties". Please describe the specific types of activities performed under this authorization (e.g., use cesium-137 sealed sources for calibration of instruments). Are these activities conducted at the Maryview facility or at another location? If these activities are conducted at another location, is the facility affiliated with Maryview Medical Center?

Please respond to these questions by December 10, 2005 and include Mail Control No. 137316 in your response. You may fax your response to my attention at 610-337-5269. If you have any questions regarding this email, please call me at 610-337-5371.

Thank you very much for your cooperation.

Sincerely,

Donna M. Janda
Health Physicist, Medical Branch

Division of Nuclear Materials Safety
NRC Region I

CC: Tara Weidner

Mail Envelope Properties

(438332EC.D1F : 8 : 55404)

Subject: Additional information for Maryview Medical Center license renewal
Creation Date: Tue, Nov 22, 2005 10:02 AM
From: Donna Janda

Created By: DMJ@nrc.gov

Recipients

hpcva.com
Healthphysics (Healthphysics@hpcva.com)

nrc.gov
kp1_po.KP_DO
TLW CC (Tara Weidner)

Post Office

kp1_po.KP_DO

Route

hpcva.com
nrc.gov

Files

MESSAGE

Size

4284

Date & Time

Tuesday, November 22, 2005 10:02 AM

Options

Expiration Date: None
Priority: Standard
Reply Requested: No
Return Notification: None

Concealed Subject: No
Security: Standard