

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:

United Hospital  
333 Smith Ave.  
St. Paul, MN  
REPORT 2005001

2. NRC/REGIONAL OFFICE

US NRC Reg. III  
2443 Warrenville Rd  
Suite 210  
Lisle, IL 60532

3. DOCKET NUMBER(S)

03002207

4. LICENSEE NUMBER(S)

22-01914-02

5. DATE(S) OF INSPECTION

Nov 16, 2005

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- ☒ 1. Based on the inspection findings, no violations were identified.
- ☐ 2. Previous violation(s) closed.
- ☐ 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.

\_\_\_\_\_ Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):

- ☐ 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11:

(Violations and Corrective Actions)

Licensee's Statement of Corrective Actions for Item 4, above.

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	G. Tarkenton	[Signature]	Nov 16 2005

(10-2003)  
10 CFR 2.201**Docket File Information**  
**SAFETY INSPECTION REPORT**  
**AND COMPLIANCE INSPECTION**

1. LICENSEE <b>United Hospital</b>		2. NRC/REGIONAL OFFICE Region III 2443 Warrenville Road Lisle, IL 60532	
REPORT NUMBER(S) 2005-001			
3. DOCKET NUMBER(S) 030-02207	4. LICENSE NUMBER(S) 22-01914-02	5. DATE(S) OF INSPECTION 11/16/05	
6. INSPECTION PROCEDURES USED 87131	7. INSPECTION FOCUS AREAS 03.01-03.07		
<b>SUPPLEMENTAL INSPECTION INFORMATION</b>			
1. PROGRAM CODE(S) 2240	2. PRIORITY 2	3. LICENSEE CONTACT John Smith	4. TELEPHONE NUMBER 651/241-5525
<input checked="" type="checkbox"/> Main Office Inspection		Next Inspection Date: 11/2007	
<input type="checkbox"/> Field Office			
<input type="checkbox"/> Temporary Job Site			

**PROGRAM SCOPE**

Licensee is a large hospital focusing on pediatrics located in St. Paul, Minnesota. Licensee has a large nuclear medicine program with four techs performing approximately 16 procedures per day. Licensee is currently authorized for IVB but is in the process of shipping its sources back to the manufacturer. Licensee does perform iodine treatments at the rate of approximately six per month.

**Performance Observations**

The inspector toured the facilities and interviewed authorized users/staff members. Each appeared knowledgeable in radiation safety and isotope handling techniques. Package receipt procedures were demonstrated for the inspector as well as rad waste handling practices. Independent surveys by the inspector did not detect any abnormal reading and were within the expected range.

The inspector reviewed the written directives for iodine treatments performed in 2005, and found no discrepancies. All procedures were followed and the written directive was reflective of the treatment given.

No abnormalities were observed.